

EQUIPMENT IDENTITY AND CUSTOMER DATA

Received Date : Calibration Date :

EQUIPMENT

1. Name :
2. Capacity : (....)
3. Resolution : (....)
4. Nilai *e* dan kelas : *e* = (.....) Kelas = (.....)
5. Type/Model :
6. Serial Number :
7. Merk/Manufacturing :

STANDARD

STANDARD

Weight Set E2 100 mf- 200g ☐
Weight Set F1 100 mg-500g ☐ 10 kg ☐
Weight F2 1kg ☐ 2 kg ☐ 5 kg ☐
Weight M2 20kg ☐

Calibration Metode : SIDIK-IK-CAL-0505
Calibration Location :

ENVIRONMENT COND.

First

End

Units

°C

%RH

THERMOHYGRO:

☐ TH-2
☐ TH-6
☐ TH-7

OWNER

1. Name :
2. Address :

CALIBRATION RESULT

1. SCALL OBSERVATION

Position	Standard Weight ()	The Reading ()
1	2	3
		$z_1 =$
Before		$m_1 =$
Adjustment		$m_2 =$
		$z_2 =$
		$z'_1 =$
After		$m'_1 =$
Adjustment		$m'_2 =$
		$z'_2 =$
Standar deviasi yang lalu (SD)		= ()

4. REPEATABILITY

No.	Middle Capacity		Maximum Capacity	
	... (...)		... (...)	
	Zero	Reading	Zero	Reading
	(<i>zi</i>)	(<i>mi</i>)	(<i>zi</i>)	(<i>mi</i>)
	1	2	1	2
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				

2. EFFECT OF TARE

Standard Weight ()	The Reading ()
	$m_1 =$
	$m_2 =$

Bentuk Pan : ...
Diameter Pan : ... ()

M = rata-rata ($m_1:m_2$)
z= Rata-rata ($z_1:z_{20}$)
C= $M_s - (M - z)$
Jika $C > 5 \times \text{resolusi}$, maka timbangan harus dilakukan adjustment
(Reference: Tabel Kriteria Kinerja Timbangan dari nilai LOP)

3. ACCURACY

Nominal ()	The Reading ()
	$z_1 =$ $m_1 =$ $m'_1 =$
	$z_2 =$ $m_2 =$ $m'_2 =$
	$z_3 =$ $m_3 =$ $m'_3 =$
	$z_4 =$ $m_4 =$ $m'_4 =$
	$z_5 =$ $m_5 =$ $m'_5 =$
	$z_6 =$ $m_6 =$ $m'_6 =$
	$z_7 =$ $m_7 =$ $m'_7 =$
	$z_8 =$ $m_8 =$ $m'_8 =$
	$z_9 =$ $m_9 =$ $m'_9 =$
	$z_{10} =$ $m_{10} =$ $m'_{10} =$
	$z_{11} =$

5. LOADING INFLUENCE ON EACH POSITION ()

Untuk keamanan loadcell timbangan, Gunakan beban uji 1/3 ~ 1/2 dari kapasitas maksimal alat, atau dapat disesuaikan dengan spek. manufaktur

Position	Center	Front	Back	Left	Right
Reading					

6. HYSTERISIS

Standar Weight ()	Reading 1 ()	Reading 2 ()
M =	$M_{(p1)} =$	$M_{(p3)} =$
M' =	$M+M' =$	$M+M' =$
	$M_{(q1)} =$	$M_{(q3)} =$
	Zero =	Zero =
	$M+M' =$	$M+M' =$
	$M_{(q2)} =$	$M_{(q4)} =$
	Zero =	Zero =
	$M_{(p2)} =$	$M_{(p4)} =$

Calibrated By	Checked By
Name : Sign :	Name : Sign :

7. DRIFT

Drift Massa Standar (*d*) = (...)

Catatan: