| PT. SIDIK | Calibration Worksheet - Conductivity Meter | LK-285-IDN |
|---|---|---|

**Received Date** :
**Calibration Date** :

## EQUIPMENT IDENTITY AND CUSTOMER DATA

1. **Name** :
2. **Range/Resolution** : ( )
   **Resolusi di tulis di masing-masing titik*
3. **Type/Model** :
4. **Serial Number** :
5. **Merk/Manufacture** :
6. **Thermohygro used** : *Insitu:* ☐TH-2 ☐TH-6 ☐TH-7
   *Inlab:* ☐TH-4

**OWNER**

1. **Name** :
2. **Address** :

**STANDARD**

| Standard Name | Usage Check |
|---|---|
| Std Solution 84 µS | ☐ |
| Std Solution 1413 µS | ☐ |
| Std Solution 5000 µS | ☐ |
| Std Solution 80000 µS | ☐ |
| PRT PT100 | ☐ |
| Victor 14+ | ☐ |

**CALIBRATION DATA**

1. **Location** :
2. **Calibration Methode** : *SIDIK-IK-CAL-0507*

| Calibrated by : | Chacked by : |
|---|---|
| Sign<br>Name | Sign<br>Name |

## CALIBRATION RESULT

Env. Condition :

| | | | |
|---|---|---|---|
| First | (°C) | First | (%RH) |
| End | (°C) | End | (%RH) |

**Before Adjustment Reading ( )**

| Solution Standard | 84 | | ☐ 1413 µS<br>☐ 1.413 mS | | ☐ 5000 µS<br>☐ 5 mS | | ☐ 80000 µS<br>☐ 80 mS | |
|---|---|---|---|---|---|---|---|---|
| | Resolusi: ( ) | | Resolusi: ( ) | | Resolusi: ( ) | | Resolusi: ( ) | |
| Repeat | µS | °C | ceklis salah satu | °C | ceklis salah satu | °C | ceklis salah satu | °C |
| 1 | | | | | | | | |
| 2 | | | | | | | | |
| 3 | | | | | | | | |
| 4 | | | | | | | | |
| 5 | | | | | | | | |

**After Adjustment Reading ( )**

| Solution Standard | 84 | | ☐ 1413 µS<br>☐ 1.413 mS | | ☐ 5000 µS<br>☐ 5 mS | | ☐ 80000 µS<br>☐ 80 mS | |
|---|---|---|---|---|---|---|---|---|
| | Resolusi: ( ) | | Resolusi: ( ) | | Resolusi: ( ) | | Resolusi: ( ) | |
| Repeat | µS | °C | ceklis salah satu | °C | ceklis salah satu | °C | ceklis salah satu | °C |
| 1 | | | | | | | | |
| 2 | | | | | | | | |
| 3 | | | | | | | | |
| 4 | | | | | | | | |
| 5 | | | | | | | | |

**Catatan :**