

PT. SIDIK	Calibration Worksheet - UV/VIS Spectrophotometer	LK-285-IDN
------------------	---	-------------------

Received Date :
Calibration Date :

Environment Condition

Time	Temperature		Humidity	
First		°C		%RH
End		°C		%RH

Thermohygro used:

Insitu: ☐ TH-2
☐ TH-6
☐ TH-7
Inlab: ☐ TH-4

EQUIPMENT IDENTITY AND CUSTOMER DATA			
EQUIPMENT			
1. Name	:		
2. Range	:	%T	nm
3. Sensitivity/Resolusi	:	%T	nm
4. Type/Model	:		
5. Serial Number	:		
6. Manufacture	:		
OWNER			
1. Name	:		
2. Address	:		
STANDAR			
1. Filter Std. Holmium/PG.Instruments/SPG080982.A		☐	
2. Filter Std. Didynium/PG.Instruments/SPG080982.B		☐	
3. Filter Std. Phptpmetric Linearity/PG.Instruments/SPG080982.C		☐	
CALIBRATION DATA			
1. Location	:		
2. Calibration Methode	:	SIDIK-IK-CAL-0508	

2. Absorbans or Transmitan Calibration		(Cuvet Size : 10 mm, at λ : 560 nm)		
λ (nm)	Neutral Filter (%T)			
	Std. Value	Measuremenet Result		
		X1	X2	X3
560	0.0			
	10.0			
	20.2			
	30.5			
	100.0			

Calibrated by :	Chacked by :
Sign	Sign
Name	Name
Catatan:	

CALIBRATION RESULT

1. Wave Length Calibration

a. Holmium Standard

No	Std. Value (λ 1)	Wavelength (nm) - Holmium		
		Measurement Result		
		X1	X2	X3
1	287.4			
2	333.6			
3	360.6			
4	418.4			
5	445.5			
6	453.3			
7	459.1			
8	536.1			
9	637.2			

b. Didiynium Standard

No	Std. Value (λ 1)	Wavelength (nm) - Didiynium		
		Measurement Result		
		X1	X2	X3
1	478			
2	513.4			
3	529.4			
4	572.4			
5	585.3			
6	684.4			
7	738.5			
8	747.7			
9	806.3			

*) Measured at 25°C and with spectral bandwidth 1 nm.