

PT. SIDIK		Calibration Worksheet - Viscometer		LK-285-IDN			
General Information							
Customer		:					
Address		:					
Calibration Date		:					
Location of Calibration		:		Equipment Name		:	
T awal		:		Manufacturer		:	
T akhir		:		Type		:	
RH awal		:		SN		:	
RH akhir		:		Range		:	
Thermohygro used		:		Resolution		:	
Inlab:		:		Methode		:	

Data Result

Before Adjustment

Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				
Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				
Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				

Mohon pilih salah satu tipe visco di bawah ini (checklist):

MODEL (on body)	MODEL (on screen display)	
DV2TLV	LV	☐
2.5DV2TLV	L3	☐
5DV2TLV	L5	☐
1/4DV2TRV	RQ	☐
1/2DV2TRV	RH	☐
DV2TRV	RV	☐
DV2THA	HA	☐
2DV2THA	A2	☐
2.5DV2THA	A3	☐
DV2THB	HB	☐
2DV2THB	B3	☐
2.5DV2THB	B5	☐

After Adjustment

Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				
Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				
Standard ()		UUT Reading ()				
		1	2	3	4	5
cP						
°C						
Resolusi UUT: ()		Rpm used :(rpm) Spindle used : No.				

Mohon pilih salah satu tipe spindle di bawah ini (lingkari):

RV1	HB1	SPIRAL	SC4-14	CPE-51 or CPA-51Z
RV2	HB2	T-A	SC4-15	CPE-52 or CPA-52Z
RV3	HB3	T-B	SC4-16	V-71
RV4	HB4	T-C	sc4-18	V-72
RV5	HB5	T-D	SC4-21	V-73
RV6	HB6	T-E	SC4-25	V-74
RV7	HB7	T-F	SC4-27	V-75
HA1	LV1	ULA	SC4-28	
HA2	LV2	DIN-81	SC4-29	
HA3	LV3	DIN-82	SC4-31	
HA4	LV4 or 4B2	DIN-83	SC4-34	
HA5	LV5	DIN-85	CPE-40 or CPA-40Z	
HA6	LV-2C	DIN-86	CPE-41 or CPA-41Z	
HA7	LV-3C	DIN-87	CPE-42 or CPA-42Z	

Standard Used:		Calibrated by:		Corrected by:	
Larutan Std Visco 100 cP		Name:		Name:	
Larutan Std Visco 100 cP		Sign:		Sign:	
Larutan Std Visco 30000 cP					
Larutan Std Visco 60000 cP					

Catatan:
