

PT. SIDIK		Calibration Worksheet – DO Meter			LK-285-IDN				
Received Date : Calibration Date :									
EQUIPMENT IDENTITY AND CUSTOMER DATA				CALIBRATION RESULT					
1. Name : 2. Range/Resolution : 3. Type/Model : 4. Serial Number/LPI : 5. Merk/Manufacture : OWNER 1. Name : 2. Address : STANDARD Standard Used : Zero Oxygen Std. 0.0 mg/l RTD Sensor/SH1/20 Victor 14+/992613877 CALIBRATION DATA 1. Location : 2. Calibration Methode : SIDIK-IK-CAL-0530				Env. Condition :		First (°C) End (°C)		First (%RH) End (%RH)	
				Solution Standard		Before adjustment Reading		Thermohygro used: Insitu: ☐ TH-2 ☐ TH-6 ☐ TH-7 Inlab: ☐ TH-4	
						0.00			
				Repeat	mg/l	°C			
				1					
				2					
				3					
				4					
				5					
				Solution Standard		After adjustment Reading			
		0.00							
Repeat	mg/l	°C							
1									
2									
3									
4									
5									
Calibrated by :		Checked by :							
Sign Name		Sign Name							
Catatan:									