

PT. SIDIK		CALIBRATION WORKSHEET FLOW METER : FLOWRATE						LK-285-IDN			
IDENTITAS ALAT			KONDISI LINGKUNGAN						PENGGERJAAN		
Nama Alat :			Suhu Ruangan Sebelum		Sebelum		%RH Thermohygro used:		Lokasi Kalibrasi : Inlab PT SIDIK		
Range/Capacity : ()			Sesudah		Sesudah		%RH ☐ TH-4 ☐ RH-5		Metode Kalibrasi : SIDIK-IK-CAL-0528		
Resolusi : ()			Reading of UUT								
Type/Model :			UUT READING ()								
Serial Number :			Set Point 1 (3 gigi valve)			Set Point 2 (4 gigi valve)			Set Point 3 (8 gigi valve)		
Merk/Manufacture :			20" 40" 60"			20" 40" 60"			20" 40" 60"		
Tanggal Terima :			1								
Tanggal Kalibrasi :			2								
			3								
IDENTITAS CUSTOMER			Reading of Weighing Result								
Nama :			STANDARD READING ()								
Alamat :			Set Point 1 (3 gigi valve)			Set Point 2 (4 gigi valve)			Set Point 3 (8 gigi valve)		
			1								
			2								
			3								
			Time ()								
STANDARD			Temperature								
			Temp.		Set Point 1		Set Point 2		Set Point 3		
			Awal Akhir		Awal Akhir		Awal Akhir		Awal Akhir		
			(°C)								
☐ Timbangan Elektronik (Dini Argeo)											
☐ Timbangan Elektronik (Sartorius)											
☐ Timbangan Elektronik (Excellent)											
☐ Timbangan Eektronik (Fujitsu)											
☐ Stopwatch/Timer											
☐ Sensor Suhu : ☐ RTD/PT 100 ☐ TC Type K ☐ TC Type N											
☐ Standar Meter Suhu : ☐ Temp. Calibrator-Constant											
Dikalibrasi Oleh:		Diperiksa Oleh:		Catatan:				Volume Pipa dari Std. ke UUT (V) = Liter			
TTD		TTD									
Nama:		Nama:									