



PT Sistem Dirgantara Inovasi Teknologi (PT Sidik)

Terakreditasi SNI ISO/IEC 17025:2017 (ISO/IEC 17025:2017) · No. LK-285-IDN

Kawasan Niaga MTC/ MIM Blok J No 25, Buahbatu, Kota Bandung, Jawa Barat

CALIBRATION CERTIFICATE

Certificate Number	CAL/2026/07/0017	Page	1 of 1
Owner	Pt THE MAGNUM ICE CREAM INDONESIA	Order Number	—
Address	Jababeka Industrial Estate, Jl. jababeka IX-B No.29 Blok D1, Wangunharja, Kec. Cikarang Utara, Kabupaten Bekasi, Jawa Barat 17531	Received Date	—
Equipment Name	pH Meter	Manufacturer	SI Analytics
Calibration Location	Onsite — Jl. Arteri Primer A-10 RT. 01 RW.12 Nyalindung Kec. Cicalengka, Kab. Bandung, Jawa Barat	Model/Type	Lab. 885
Calibration Date	30 Juli 2026	Serial Number	IMTE-WQ-129
Calibration Method	SIDIK-IK-CAL-0506	Capacity/Graduation	0-14 pH / 0,01 pH
Env. Condition	T: 25,0°C ± 1,7°C — %RH: 55,00% ± 4,8%	Technician ID	DR

CALIBRATION REPORT

Standard Value	Unit Under Test	Correction	U95% (±)
4,00	4,01	-0,01	0,02
7,00	7,00	0,00	0,02
10,01	10,00	0,01	0,03

The Uncertainty is taken at a Confidence Level 95% and Coverage Factor (k) = 2

Calibration results are not to be announced and only apply to related tools

STANDARD USED

Name	Merk/Type	Serial Number	Traceable to SI through
pH Buffer Solution 4	Supelco/Merck	HC32513535	Merck KGaA
pH Buffer Solution 7	Supelco/Merck	HC46341939	Merck KGaA
pH Buffer Solution 10	Supelco/Merck	HC45400338	Merck KGaA
TH-2	—	TH-2	LK-285-IDN

Issuance Date: 30 Juli 2026

Alex Misramto
Technical Manager