

Assignment #2:

Analysis of electrocardiographic (ECG) signals

Žane Bučan
OBSS 2025/26 , FRI, UL
zb26346@student.uni-lj.si

Abstract—An algorithm classifying ECG heartbeats as normal (N) or ventricular ectopic (V) was implemented using linear norm strategies. Performance was evaluated on the MIT-BIH Arrhythmia Database [1], showing high accuracy. The multi distance classifier is compared to the single distance classifier.

I. INTRODUCTION

An electrocardiogram (ECG) is a non-invasive medical test that records the electrical activity of the heart over time. Using electrodes, it measures the tiny electrical signals produced during each heart beat, which reflect the heart's rhythm and function. ECG signals are widely used to detect and monitor cardiovascular conditions, such as arrhythmia, myocardial infarction and other heart disorders. The analysis of these signals involves classifying heart beat types to properly diagnose a person, if his heart beats normal or abnormal. That is why it is important that we develop robust algorithms to automatically classify each heart beat in the signal.

II. METHODOLOGY

The algorithm reads a record, QRS locations, and annotations. Distances between beats and a reference normal beat are computed using:

$$d_1(x, y) = \frac{1}{N} \sum_{i=1}^N |x_i - y_i| \quad (1)$$

$$d_2(x, y) = \sqrt{\frac{1}{N} \sum_{i=1}^N (x_i - y_i)^2} \quad (2)$$

$$d_\infty(x, y) = \max_{i=1, \dots, N} |x_i - y_i| \quad (3)$$

$$d_{\text{corr}}(x, y) = 1 - \frac{\sum_{i=1}^N (x_i - \bar{x})(y_i - \bar{y})}{\sqrt{\sum_{i=1}^N (x_i - \bar{x})^2 \sum_{i=1}^N (y_i - \bar{y})^2}} \quad (4)$$

$$d_r(x, y) = 1 - \frac{\sum_{i=1}^N (x_i - \bar{x})(y_i - \bar{y})}{\sqrt{\sum_{i=1}^N (x_i - \bar{x})^2 \sum_{i=1}^N (y_i - \bar{y})^2}} \quad (5)$$

with $\bar{x} = \frac{1}{N} \sum_i x_i$, $\bar{y} = \frac{1}{N} \sum_i y_i$.

Then the algorithm includes the next steps ('N' - normal beat, 'V' - ventricular beat):

- **QRS Extraction:** 60 ms before and 100 ms after each annotated 'N' and 'V' peak.
- **Normalization:** To adjust QRS complexes by amplitude the algorithm normalizes the signal with the function:

$$\hat{q} = \frac{q - \mu}{\max |q - \mu|} \quad \text{where} \quad \mu = \frac{1}{N} \sum_{i=1}^N q_i$$

Where q is a QRS complex and q_i is a value in the signal.

- **Reference building:** The algorithm needs a good 'N' heart beat to compare it to the other ones. That is why it takes the first 5 minutes of the signal, extracts 'N' heart beats from the 5 minute time frame and builds a reference beat of the average 'N' heart beats. The standard d_2 norm classifier declares the first 'N' heart beat as a reference.
- **Threshold estimation:** The goal of the threshold estimation is to determine, for each distance strategy, a threshold that separates 'N' from 'V' QRS complexes, and to assign a weight to each strategy for voting. For each QRS beat q_i and a chosen reference QRS q_{ref} , it computes its distance according to a selected strategy s :

$d_s(q_i, q_{\text{ref}})$ = distance between q_i and q_{ref} using strategy s

From the distances, it considers:

- D_s^N - distances of 'N' beats only
- D_s^{all} - distances of all beats

This allows the algorithm to model the expected distance distribution of normal beats.

The algorithm then calculates some statistics like:

- Median distance of normal beats:

$$\text{median}_s = \text{median}(D_s^N)$$

- Median absolute deviation (MAD):

$$\text{MAD}_s = \text{median}(|D_s^N - \text{median}_s|)$$

- Scaled MAD, which approximates the standard deviation:

$$\sigma_s = 1.4826 * \text{MAD}_s$$

With that the algorithm can generate candidate thresholds as:

$$\theta_s(k) = \text{median}_s + k * \sigma_s, k \in \{1.0, 1.2, \dots, 7.8\}$$

This creates a range of thresholds around the normal distribution of distances.

Next step is to predict a beat type for each threshold:

$$\hat{y}_i(k) = \begin{cases} N, & d_s(q_i, q_{\text{ref}}) \leq \theta_s(k) \\ V, & d_s(q_i, q_{\text{ref}}) > \theta_s(k) \end{cases}$$

For each candidate threshold, the classifier computes Sensitivity (Se), Positive predictivity (+P), Specificity (Sp) and F1-score.

Then it creates an optimal threshold with the equation:

$$k^* = \arg \max_k \begin{cases} F_1, & \text{if PVCs exist and } Sp \geq 0.90 \\ Sp, & \text{if no PVCs} \end{cases}$$

The final threshold is:

$$T_j = \text{median}_s + k^* * \sigma_s$$

Each strategy is then weighted for voting:

$$w_s = \begin{cases} \max(F_1, 0.05), & \text{if PVCs exist} \\ 0.5 \text{ or } 0.1, & \text{if no PVCs depending on Sp} \end{cases}$$

A higher weight indicates a strategy that better separates N from V beats.

For the standard d_2 norm classifier, the classification threshold was estimated from the distribution of distances between normal QRS complexes and a reference beat. A robust median-based estimator was used, with the threshold defined as the median distance plus a scaled median absolute deviation. This provides a data-driven yet robust default threshold for separating normal and ventricular beats.

- **Classification:** The final step is to classify each heart beat. Firstly it calculates RR intervals for the QRS fiducial points $\{s_i\}$:

$$RR_i = s_i - s_{i-1}, i = 2, \dots, n$$

and the median RR interval is computed:

$$RR_{\text{med}} = \text{median}(RR_i)$$

. These are then used to detect early beats (short RR) and pauses (long RR):

$$\text{is_early} = RR_{i-1} < 0.85 \cdot RR_{\text{med}},$$

$$\text{has_pause} = RR_i > 1.10 \cdot RR_{\text{med}}$$

Then each strategy is used to determine the distance between the QRS (q_i) and the previously calculated reference beat ($\hat{q}$). Distances are then compared to their estimated threshold (T_j), which were defined during training.

Next step is to determine a preliminary vote $v_{i,j}$:

$$v_{i,j} = \begin{cases} V, & \text{if } d_j(q_i, \hat{q}) > T_j * 1.05 \\ & \text{if is_early and has_pause and} \\ V, & d_j(q_i, \hat{q}) > 0.75 T_j \\ V, & \text{if is_early and } d_j(q_i, \hat{q}) > 0.9 T_j \\ N, & \text{otherwise} \end{cases}$$

The final step is to calculate the combined weighted score (s_i) for the 'N' and 'V' combining each strategy:

$$\text{score}_i = \sum_{j=1}^m w_j \mathbf{1}_{v_{i,j}=i}, \quad i \in \{V, N\}$$

The best score is then the final label of the QRS complex

III. EXPERIMENTS

Only N and V reference annotations were used for both training and evaluation. All other annotated heartbeat types were excluded prior to QRS extraction, in accordance with the assignment specification.

If a record did not have any 'N' beats in the first 5 minutes, the record was skipped.

To develop this algorithm, I used the MIT-BIH Arrhythmia Database [1] from PhysioNet [2]. The efficiency of the algorithm was then defined with the average sensitivity (Se), average positive predictivity (+P) and average specificity (Sp) threw the whole database. with the addition of the F1 Score to get a better feel of the balance between missed detections and false alarms.

$$Se = \frac{TP}{TP + FN}, +P = \frac{TP}{TP + FP}, Sp = \frac{TN}{TN + FP}$$

IV. RESULTS AND DISCUSSION

The results show, that the algorithm was successfully implemented.

A. Results

The multi distance classifier achieved:

Total PVCs Detected (TP):	6534
Total Missed PVCs (FN):	74
Total False Alarms (FP):	521
Global Sensitivity:	98.88%
Global Precision:	92.61%
Global Specificity:	99.31%
Global F1 Score:	95.65%

while single d_2 norm classifier achieved:

Total PVCs Detected (TP):	6414
Total Missed PVCs (FN):	194
Total False Alarms (FP):	448
Global Sensitivity:	97.06%
Global Precision:	93.47%
Global Specificity:	99.40%
Global F1 Score:	95.23%

B. Discussion

I achieved great results. High positive predictivity indicates that most classes correspond to true QRS annotations. Sensitivity indicates that some QRS complexes were falsely classified by the algorithm, possibly due to noise in the data. Overall, the high positive predictivity combined with strong sensitivity demonstrates that the implemented classifier is effective for classifying.

Compared to the single-distance d_2 classifier, the proposed multi-distance weighted voting approach significantly improves sensitivity to ventricular ectopic beats, while maintaining comparable precision and specificity. The overall improvement in F1 score confirms a better balance between missed detections and false alarms.

If we take a look at the raw statistics, the multi distance classifier detected 120 more 'V' heart beats and had a 120 decrease in missed 'V' beats, while having a 73 increase in false alarms. This trade-off is clinically meaningful, as reducing missed ventricular beats is generally more important than a moderate increase in false alarms.

V. CONCLUSION

In this work, a QRS beat classifying algorithm was implemented and evaluated on the MIT-BIH Arrhythmia Database. The algorithm successfully classifies beats with high accuracy, even tho its base calculations are linear norms which are not as advanced and complex as deep learning classifiers. The results demonstrate that the implemented approach is somewhat reliable for automated ECG analysis, providing a solid foundation for further studies in heartbeat classification.

REFERENCES

- [1] G. B. Moody and R. G. Mark, "The impact of the mit-bih arrhythmia database," *IEEE Engineering in Medicine and Biology Magazine*, vol. 20, no. 3, pp. 45–50, May–June 2001, PMID: 11446209.
- [2] A. L. Goldberger, L. A. N. Amaral, L. Glass, J. M. Hausdorff, P. C. Ivanov, R. G. Mark, J. E. Mietus, G. B. Moody, C.-K. Peng, and H. E. Stanley, "Physiobank, physiotoolkit, and physionet: Components of a new research resource for complex physiologic signals," *Circulation*, vol. 101, no. 23, pp. e215–e220, 2000, rRID:SCR_007345.