

Dial Gauge vs. Virtual Encoder Position Comparison
X Forward Frame Position
Virtual Encoder
Mean Diff: 4.6 ± 4.3 μm
Dial Gauge (±10μm)
Position (μm)
Iteration
Y Forward Frame Position
Virtual Encoder
Mean Diff: 8.5 ± 7.8 μm
Dial Gauge (±10μm)
Position (μm)
Iteration
X Backward Frame Position
Virtual Encoder
Mean Diff: 5.3 ± 2.5 μm
Dial Gauge (±10μm)
Position (μm)
Iteration
Y Backward Frame Position
Virtual Encoder
Mean Diff: 7.1 ± 4.9 μm
Dial Gauge (±10μm)
Position (μm)
Iteration