

USAGE REPORT



ALTAIR 5X - A5X-00206407

Start Date: Fri, Dec 05, 2025, 09:56:15 AM GMT+11
End Date: Fri, Dec 05, 2025, 05:13:15 PM GMT+11
Duration: 7 hours 17 minutes

Was Assigned To

Worker **Vent 407**
Company **Ventilation**
Department **VENT**

CALIBRATION

Instrument Calibration Status **✓ Compliant**

Last Calibration Test **Passed**
Wed, Dec 3, 2025, 9:33:51 AM GMT+11

Last Bump Test **Expired**
Thu, Dec 4, 2025, 8:32:46 AM GMT+11

Sensors Calibration

Status	Gas Type	Calibration Status	Bump Status
 Bump Due	Combustible	Passed Dec 3, 2025, 9:33:51 AM GMT+11	Expired Dec 4, 2025, 8:32:46 AM GMT+11
 Bump Due	O ₂	Passed Dec 3, 2025, 9:33:51 AM GMT+11	Expired Dec 4, 2025, 8:32:46 AM GMT+11
 Bump Due	CO	Passed Dec 3, 2025, 9:33:51 AM GMT+11	Expired Dec 4, 2025, 8:32:46 AM GMT+11
 Bump Due	NO ₂	Passed Dec 3, 2025, 9:33:51 AM GMT+11	Expired Dec 4, 2025, 8:32:46 AM GMT+11
 Bump Due	NH ₃	Passed Dec 3, 2025, 9:33:51 AM GMT+11	Expired Dec 4, 2025, 8:32:46 AM GMT+11

INSTRUMENT CONFIGURATION

Alert Behaviors

Alarms

LEDs, Horn, Vibrate

Instrument Alerts

Man Down Alarm

Enabled

Instrument Configuration

Firmware Version

2.24.0

Language

English

Calibration Interval

Disabled

Bump Test Interval

Disabled

Session Logging Status

Disabled

Periodic Logging Status

Disabled

INSTRUMENT ALARM SETPOINTS

Combustible

Serial Number	08220150077
Manufactured	01/2022
Deadband Enabled	Disabled
Exposure Warning	5 %LEL
Exposure Alarm	10 %LEL
Ovrrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

O₂

Serial Number	01220787326
Manufactured	07/2022
Deficiency Alarm	19.5 %VOL
Exposure Alarm	23 %VOL
Ovrrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

CO

Serial Number	11230380101
Manufactured	03/2023
Exposure Warning	30 PPM
Exposure Alarm	60 PPM
STEL Alarm Alarm	60 PPM
TWA Alarm Alarm	30 PPM
Ovrrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

NO₂

Serial Number	11230380101
Manufactured	03/2023
Exposure Warning	2.5 PPM
Exposure Alarm	5 PPM
STEL Alarm Alarm	5 PPM
TWA Alarm Alarm	2.5 PPM
Overrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

NH₃

Serial Number	06220880179
Manufactured	08/2022
Exposure Warning	25 PPM
Exposure Alarm	50 PPM
STEL Alarm Alarm	35 PPM
TWA Alarm Alarm	25 PPM
Overrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

GAS READINGS

No gas readings available

USAGE EVENTS

Dec 5, 2025, 9:56:15 AM

Instrument Powered On

Received by Grid Dec 8, 2025, 11:09:10 AM

Dec 5, 2025, 5:13:15 PM

Instrument Powered Off

Received by Grid Dec 8, 2025, 11:09:10 AM

NOTES

No notes present.