

USAGE REPORT



The Safety Company

ALTAIR 5X - A5X-00035786

Start Date: Fri, Dec 05, 2025, 09:58:30 AM GMT+11
End Date: Fri, Dec 05, 2025, 05:11:45 PM GMT+11
Duration: 7 hours 13 minutes

Was Assigned To

Company

Ore Processing

Department

Surface Fixet Plant Toolstore

CALIBRATION

Instrument Calibration Status

✓ Compliant

Last Calibration Test

Passed

Thu, Dec 4, 2025, 8:00:17 AM GMT+11

Last Bump Test

Expired

Thu, Dec 4, 2025, 8:00:17 AM GMT+11

Sensors Calibration

Status	Gas Type	Calibration Status	Bump Status
 Bump Due	Combustible	Passed Dec 4, 2025, 8:00:17 AM GMT+11	Expired Dec 4, 2025, 8:00:17 AM GMT+11
 Bump Due	O ₂	Passed Dec 4, 2025, 8:00:17 AM GMT+11	Expired Dec 4, 2025, 8:00:17 AM GMT+11
 Bump Due	CO	Passed Dec 4, 2025, 8:00:17 AM GMT+11	Expired Dec 4, 2025, 8:00:17 AM GMT+11
 Bump Due	H ₂ S	Passed Dec 4, 2025, 8:00:17 AM GMT+11	Expired Dec 4, 2025, 8:00:17 AM GMT+11
 Bump Due	SO ₂	Passed Dec 4, 2025, 8:00:17 AM GMT+11	Expired Dec 4, 2025, 8:00:17 AM GMT+11

INSTRUMENT CONFIGURATION

Alert Behaviors

Alarms

LEDs, Horn, Vibrate

Instrument Alerts

Man Down Alarm	Enabled
----------------	---------

Instrument Configuration

Firmware Version	2.24.0
Language	English
Calibration Interval	Disabled
Bump Test Interval	Disabled
Session Logging Status	Enabled
Periodic Logging Status	Disabled

INSTRUMENT ALARM SETPOINTS

Combustible

Serial Number	00240653247
Manufactured	06/2024
Deadband Enabled	Disabled
Exposure Warning	5 %LEL
Exposure Alarm	10 %LEL
Ovrrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

O₂

Serial Number	01240302422
Manufactured	03/2024
Deficiency Alarm	19.5 %VOL
Exposure Alarm	23 %VOL
Ovrrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

CO

Serial Number	04240206893
Manufactured	02/2024
Exposure Warning	25 PPM
Exposure Alarm	60 PPM
STEL Alarm Alarm	60 PPM
TWA Alarm Alarm	25 PPM
Overrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	


H₂S

Serial Number	04240206893
Manufactured	02/2024
Exposure Warning	10 PPM
Exposure Alarm	15 PPM
STEL Alarm Alarm	15 PPM
TWA Alarm Alarm	10 PPM
Overrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

SO₂

Serial Number	07240500031
Manufactured	05/2024
Exposure Warning	2 PPM
Exposure Alarm	5 PPM
STEL Alarm Alarm	5 PPM
TWA Alarm Alarm	2 PPM
Overrange Alarm	
Underrange Warning	
Replace soon Notification	
Replace now Notification	

GAS READINGS



USAGE EVENTS

Dec 5, 2025, 9:58:30 AM	Instrument Powered On
Received by Grid Dec 8, 2025, 11:05:01 AM	
Dec 5, 2025, 5:11:45 PM	Instrument Powered Off
Received by Grid Dec 8, 2025, 11:05:01 AM	

NOTES

No notes present.