

Diagnostic Medical Image Processing

Reconstruction

WS 2010/11



Joachim Hornegger, Dietrich Paulus,
Markus Kowarschik

Lehrstuhl für Mustererkennung (Informatik 5)
Friedrich-Alexander-Universität Erlangen-Nürnberg



Diagnostic Medical Image Processing

1 C-Arm Computed Tomography

■ Historical Notes

- Principle of C-arm CT
- Efficient Projection of Voxels
- Take Home Messages
- Further Readings



Historical notes

- C-arm CT was introduced as a commercial product in 1999.
- Reconstruction module was just an add on feature that time.
- Today C-arm CT became standard and a driving force of innovation in X-ray imaging.
- The major problem was the mathematical characterization and the calibration of the acquisition geometry.
- The breakthrough idea was the usage of homogeneous coordinates and of (3×4) -projection matrices that were standard in computer vision since the early 90-ies.
- In 2000 the first hardware accelerators (DSP's) for C-arm CT were shipped.
- Nowadays hardware accelerated C-arm CT is standard, and still in the major focus of research.



Diagnostic Medical Image Processing

1 C-Arm Computed Tomography

- Historical Notes
- Principle of C-arm CT
- Efficient Projection of Voxels
- Take Home Messages
- Further Readings



Basic Idea

Data Acquisition System



CT systems



C-arm systems



Mobile C-arm systems

Set of digital
projection images



Image
Reconstruction
Algorithm



Tomographic Slice Image




3D image

: Image courtesy of Holger Scherl, LME



Basic Idea




WCS World Coordinate System
DCS Detector Coordinate System

: Image courtesy of Holger Scherl, LME



Basic Idea



: Image courtesy of Holger Scherl, LME



Principle of C-arm CT

$$P_i \begin{pmatrix} x \\ y \\ z \\ 1 \end{pmatrix} = \begin{pmatrix} u \\ v \\ w \end{pmatrix} \mapsto \begin{pmatrix} u/w \\ v/w \end{pmatrix}$$

Input 2-D X-ray projection images, projection matrices.

Step 1 Perform proper weighting scheme on images.

Step 2 Filter the projection images along horizontal lines.

Step 3 Perform a weighted back-projection along projection rays.

Output Reconstructed volume



Diagnostic Medical Image Processing

1 C-Arm Computed Tomography

- Historical Notes
- Principle of C-arm CT
- **Efficient Projection of Voxels**
- Take Home Messages
- Further Readings



Volume Elements: Voxels

The back-projection is done voxel-driven, i.e. each voxel of the $(N \times N \times N)$ volume is projected into the image plane for each projection image.

- The first element of the volume has the coordinates
 $\mathbf{v} = (v_1, v_2, v_3, 1)^T$
- The increment in x -direction is denoted by $\Delta\mathbf{x} = (\Delta x, 0, 0, 1)^T$.
- The increment in y -direction is denoted by $\Delta\mathbf{y} = (0, \Delta y, 0, 1)^T$.
- The increment in z -direction is denoted by $\Delta\mathbf{z} = (0, 0, \Delta z, 1)^T$.

Back-Projection Algorithm



For a given set of projection matrices $\mathbf{P}_n \in \mathbb{R}^{3 \times 4}$, $n = 1, 2, \dots, N_v$, we get the following back-projection algorithm:

FOR $n = 1, 2, \dots, N_v$
FOR $m = 1, 2, \dots, N$
FOR $l = 1, 2, \dots, N$
FOR $k = 1, 2, \dots, N$
$\mathbf{p} = \mathbf{P}_n(\mathbf{v} + k\Delta\mathbf{x} + l\Delta\mathbf{y} + m\Delta\mathbf{z})$
$x = p_1/p_3$
$y = p_2/p_3$
$\text{Volume}[m][l][k] = \text{Volume}[m][l][k] + \frac{1}{p_3^2} \text{filtered_image}(x, y)$



Efficiency Considerations

Due to the linearity properties we can write:

$$\mathbf{p} = \mathbf{P}_n(\mathbf{v} + k \cdot \Delta \mathbf{x} + l \cdot \Delta \mathbf{y} + m \cdot \Delta \mathbf{z}) \quad (1)$$

$$= \mathbf{P}_n \mathbf{v} + k \underbrace{\mathbf{P}_n \cdot \Delta \mathbf{x}}_{\mathbf{p}_x} + l \underbrace{\mathbf{P}_n \cdot \Delta \mathbf{y}}_{\mathbf{p}_y} + m \underbrace{\mathbf{P}_n \cdot \Delta \mathbf{z}}_{\mathbf{p}_z} \quad (2)$$

We conclude:

- The vector $\mathbf{P}_n \mathbf{v}$ is constant for all iterations.
- The values of \mathbf{p}_x , \mathbf{p}_y , and \mathbf{p}_z are constant for all iterations.
- The matrix vector product can be rewritten to get a more efficient back-projection.



Back-Projection Algorithm

compute: \mathbf{v}_n , \mathbf{p}_x , \mathbf{p}_y , and \mathbf{p}_z
FOR $m = 1, 2, \dots, N$
FOR $l = 1, 2, \dots, N$
FOR $k = 1, 2, \dots, N$
$\mathbf{p} = \mathbf{v}_n + k\mathbf{p}_x + l\mathbf{p}_y + m\mathbf{p}_z$
$x = p_1/p_3$
$y = p_2/p_3$
$\text{Volume}[m][l][k] = \text{Volume}[m][l][k] + \frac{1}{p_3^2} \text{filtered_image}(x, y)$



Efficient Ratio Computation

- The most time consuming operation in the inner loop is the ratio computation.
- The projection geometry is computed in the calibration step. The values of p_3 and its range are known.
- Ratio $1/p_3$ can be efficiently computed by:
 - regression function (approx. errors known in advance)
 - look-up table



Diagnostic Medical Image Processing

1 C-Arm Computed Tomography

- Historical Notes
- Principle of C-arm CT
- Efficient Projection of Voxels
- **Take Home Messages**
- Further Readings



Take Home Messages

- Acquisition Geometry in C-arm systems is characterized by projection matrices.
- Calibration is done offline and by technicians.
- Voxel-driven back-projection more or less determines the complexity of reconstruction.
- Hardware oriented implementation required.

Diagnostic Medical Image Processing



1 C-Arm Computed Tomography

- Historical Notes
- Principle of C-arm CT
- Efficient Projection of Voxels
- Take Home Messages
- Further Readings



Further Readings

- Cone beam reconstruction was first introduced in 1984 by this paper:
L.A. Feldkamp, L.C. Davis and J.W. Kress: Practical cone-beam algorithm, J Opt Soc Am, A6, 1984, p. 612-619
- The usage of homogeneous coordinates for C-arm reconstruction is described in:
K. Wiesent, K. Barth, N. Navab, P. Durlak , T. Brunner, O. Schuetz, W. Seissler: Enhanced 3-D-reconstruction algorithm for C-arm systems suitable for interventional procedures, IEEE Transactions on Medical Imaging, 2000 May ; Vol. 19, No. 5:391-403