

Speaker Script / 演讲稿 — SNOLAB R4 Phonon Pulse Templates

约 10 分钟，13 页。每页先中文、后英文，内容一一对应，**英文可以直接照读**。原则：稿子是说的话，不是幻灯片的复读——具体参数都在屏幕上，嘴里讲思路和为什么。斜体是给自己的提示，不用念。

Slide 1 — Title (约 30 秒)

中文：大家下午好。今天讲我为 SNOLAB Run 4 做的声子脉冲模板。简单说，就是给 13 个探测器的每个声子通道做出一个“标准脉冲形状”，给 optimal filter 用。我按实际做的顺序讲：事件怎么选的，波形怎么拟合和对齐的，两个质量 cut 是怎么从数据里定出来的，最后是交付的两族模板。

English: Good afternoon everyone. Today I'll talk about the phonon pulse templates I built for SNOLAB Run 4. In short: for each phonon channel of thirteen detectors, we build a standard pulse shape for the optimal filter to use. I'll go in the order we actually did the work: how the events were selected, how each trace is fitted and aligned, how the two quality cuts were derived from the data itself, and finally the two template families we delivered.

Slide 2 — From the Ge-activation K-line to an event sample (约 65 秒)

中文：先说事件从哪来。做模板需要一大批干净、彼此相同的脉冲。我们用的是 Ge 活化 K 线：Cf 活化之后，每个探测器里都会持续出现 10.37 keV 的单能事件，这是现成的理想样本。屏幕上 13 个探测器的 PTOFamps 谱——可以把 PTOFamps 理解成官方处理链给出的能量代理量。每格里的红线，是值班分析标出来的 K 线位置，我们直接拿来用。我们做的事情非常简单：以红线为中心开一个窗，上下各 1.35 倍，窗内的事件全要，不加任何别的条件。然后把每个选中事件所有通道的原始波形，原封不动存下来，一共约 120 GB。大家看这些谱：安静的探测器比如 Z7，K 线峰和噪声峰分得很开；但不少弱探测器，红线就插在噪声堆里——窗口一开，噪声肯定跟着进来。这是故意的：我们宁愿先多收，也要让后面每一步筛选清清楚楚、随时可以回退——因为存的是原始数据，什么都没丢。

English: Let me start with where the events come from. To build a template, you need a large sample of clean, identical pulses. We use the germanium activation K-line: after the californium activation, every detector keeps producing mono-energetic events at 10.37 keV — a perfect, ready-made sample. What you see here are the PTOFamps spectra of all thirteen detectors — you can think of PTOFamps as the energy estimate from the official processing. The red line in each panel is the K-line position marked by the ops-shift analysis, and we simply take that as given. What we do is very simple: open a window around the red line, a factor of one point three five on each side, keep everything inside, and

apply no other condition. Then, for every selected event, we save the raw traces of all channels, completely untouched — about a hundred and twenty gigabytes in total. Now look at the spectra: on a quiet detector like Z7, the K-line peak is well separated from the noise peak. But on the weaker detectors, the red line sits right inside the noise population — so the window will let noise in. And that's deliberate. We'd rather collect too much, and keep every later selection explicit and reversible — nothing is lost, because what we stored is raw.

Slide 3 — Per-trace algorithm (约 80 秒)

中文：拿到原始波形后，每一条都过同样的五步。完整参数都在屏幕上，我挑重点讲。前三步是准备：低通滤波抹掉高频噪声，扣基线，按峰值归一。核心是第四步的拟合：一个双指数函数，一个指数管上升、一个管下降——这是声子脉冲的标准形状。重点在于，五个拟合参数里，连脉冲的起始时刻也是放开的。为什么？因为真实触发时刻每个事件都不一样——拟合结果显示，起点普遍比名义位置晚两百多个采样点。如果把起点钉死，这个误差就会被转嫁到上升、下降时间上，把所有参数都带歪。拟合完，每条波形记两个质量指标：一个叫 `fit_ok`，检查参数物不物理——幅度得是正的，上升得比下降快；另一个叫 `NRMSE`，全称 Normalized Root-Mean-Square Error，归一化均方根误差——就是拟合残差除以脉冲峰值，衡量拟合得好不好；`NRMSE` 等于 0.05，意思就是拟合偏差只有脉冲高度的 5%。注意，这一步只记录、不筛选。最后一步，把实测波形按拟合出的起点平移对齐——只是平移，波形本身一点没动。下面两张图就是拟合的产出：所有拟合曲线画在同一个起点上。左边是全部曲线，主束旁边有些散开的；右边是用 `NRMSE` 筛过之后，只剩一束非常一致的形状。这个筛选标准怎么定的，接下来讲。

English: Once we have the raw traces, every single one goes through the same five steps — the full parameters are on the slide, so let me just walk through the ideas. The first three steps are preparation: a low-pass filter to smooth away high-frequency noise, baseline subtraction, and normalization to the peak. The core is the fit: a two-exponential function — one exponential for the rise, one for the fall — which is the standard shape of a phonon pulse. And here's the important part: among the five fit parameters, even the pulse start time is left free. Why? Because the true trigger time is different for every event — the fits tell us the onset is typically more than two hundred samples later than the nominal position. If you pin the onset, that error gets pushed into the rise and fall times, and everything comes out biased. After the fit, each trace gets two quality numbers. One is called `fit_ok` — a sanity check that the parameters are physical: positive amplitude, rise faster than fall. The other is the `NRMSE` — the normalized root-mean-square error — which is simply the fit residual divided by the pulse height, so it tells you how good the fit actually is: an `NRMSE` of zero point zero five means the fit misses the data by five percent of the pulse height. At this stage we only record these; we don't cut on them yet. The last step: shift the measured trace by the fitted onset, so all pulses line up — a pure shift, the waveform itself is untouched. The two plots at the bottom show what the fit gives us: all the fitted curves drawn from the same starting point. On the left, everything — you can see stragglers

spreading away from the main bundle. On the right, after selecting on NRMSE, one tight, consistent shape remains. Where that selection comes from is what I'll show next.

Slide 4 — Fit quality at a glance (约 40 秒)

中文：先看单个事件层面，拟合到底靠不靠谱。网格图里一行是一个事件、一列是一个通道，蓝色是滤波后的波形，红色是拟合。规律非常清楚：左边这批是噪声触发，它在**所有**通道里同时拟合失败；右边是真正的 K 线事件，**所有**通道同时拟合得很好，NRMSE 只有百分之几。也就是说，“拟合好坏”这个指标真的能把两类事件分开——这给了我们用 NRMSE 做筛选的底气。

English: First, a sanity check at the level of individual events. In these grids, each row is one event and each column is one channel — blue is the filtered trace, red is the fit. The pattern is very clear. The left block is noise triggers: the fit fails in **every** channel at the same time. The right block is genuine K-line events: **every** channel fits well at the same time, with NRMSE at the level of a few percent. So "how well the fit went" genuinely separates the two kinds of events — and that's what gives us the confidence to select on NRMSE.

Slide 5 — Alignment result (约 40 秒)

中文：对齐之后长这样——这里直接放大在峰值附近。蓝色是上千条实测波形叠在一起，红色是它们的平均，橙色是拟合曲线的加权平均。安静探测器上这个束非常紧——脉冲形状确实高度一致，做模板的原料是好的。请注意一个细节：两个通道里，主束旁边都有一条淡淡的、错开的小束，左边这个通道特别明显。而且这两个通道分别在晶体的两个面上——所以这不是某一个通道的毛病。先记住它，一会儿回来讲它是什么。

English: This is what things look like after alignment — zoomed right in on the peak. The blue band is thousands of measured traces stacked on top of each other; red is their average; orange is the weighted average of the fitted curves. On a quiet detector this bundle is very tight — the pulse shape really is highly consistent, so the raw material for a template is good. Now, one detail I'd like you to notice: in both channels there's a faint, displaced little bundle next to the main one — very clear in the left panel. And these two channels sit on opposite faces of the crystal, so this is not a quirk of a single channel. Keep it in mind — we'll come back to what it is.

Slide 6 — Quality cut 1: $\text{NRMSE} \leq 0.4$ (约 55 秒)

中文：现在讲第一个 cut 怎么定的。把所有物理拟合的 NRMSE 画成分布，每个探测器都是同一个样子：两个峰。左边的峰是拟合好的，典型值百分之五到百分之十；右边的峰在 1 到 2 附近，就是那些根本没有脉冲的噪声触发。两峰之间有一个很深的谷，位置在 0.4 左右——阈值就取 0.4。我想强调：这个数不是调出来的，是分布自己长出来的。上图是安静的 Z7，两个峰离得

很远；下图是弱探测器 Z22，噪声峰是主体——对这些探测器来说，这一刀就是把真脉冲从噪声堆里捞出来的那一刀。

English: Now, how the first cut was set. If you histogram the NRMSE of all physical fits, every detector shows the same picture: two populations. The left one is the good fits, typically five to ten percent. The right one, around one or two, is the noise triggers — traces with no pulse in them at all. In between there's a deep valley, at about zero point four — and that's where we put the threshold. I want to stress: this number wasn't tuned, it's simply what the distribution itself shows. The top plot is quiet Z7, where the two populations are far apart. The bottom is Z22, a weak detector, where noise dominates — and for those detectors, this cut is exactly what pulls the real pulses out of the noise.

Slide 7 — The rejected population is noise (约 45 秒)

中文：定了 cut 还不够，得验证它切掉的确实是垃圾。左边是被切掉的事件的**原始**波形——可以看到，里面就没有脉冲，纯粹是噪声，只不过一条很慢的曲线恰好凑在了它上面。右边换个角度看同一件事：绿色是通过的拟合曲线，红色是被切掉的。绿的又快又一致；红的又散、中位 NRMSE 也高好几倍。结论很干净：这一刀切掉的是噪声，没有伤到物理。

English: Setting a cut isn't enough — we have to verify that what it removes really is junk. On the left are the **raw** traces of the rejected events. As you can see, there is simply no pulse there — it's pure noise, on which some slow curve happened to converge. On the right, the same thing from another angle: green curves passed the cut, red ones were removed. The green ones are fast and consistent; the red ones are scattered, with a median NRMSE several times higher. So the conclusion is clean: this cut removes noise, and it doesn't touch the physics.

Slide 8 — Quality cut 2: $\tau_{\text{rise}} \leq 0.3 \text{ ms}$ (约 50 秒)

中文：第二个 cut 针对一种漏网的情况。有些探测器的基线会慢慢漂移，这种漂移很平滑，一条很慢的双指数能贴得很好、残差很小——NRMSE 抓不到它。但它有一个藏不住的特征：慢。看上升时间的分布：真脉冲集中在 0.1 毫秒附近，非常窄；漂移拟出来的上升时间比这大得多。所以加一条上限：上升时间不超过 0.3 毫秒。在安静探测器上，这一刀只损失百分之二到五的信号。但要坦白一个代价：确实存在一小群**真实的**慢脉冲，也会被这条线切到。这个取舍还没有最终定——下一页就是这群事件。

English: The second cut targets something that slips through. On some detectors the baseline drifts slowly, and the drift is smooth — a very slow two-exponential can hug it with a tiny residual, so the NRMSE never notices. But it has one feature it can't hide: it's slow. Look at the rise-time distribution: real pulses cluster around zero point one milliseconds, very narrow, while the drift fits come out much slower. So we added a ceiling: rise time no more than zero point three milliseconds. On quiet detectors this costs only a few percent of the signal. But I should be upfront about one trade-off: there is a small

population of **genuinely** slow pulses, and this line cuts into them too. That decision isn't final yet — and that population is the next slide.

Slide 9 — A genuine slow-pulse population (约 45 秒)

中文：这就是刚才对齐图里那条"影子"。这些事件拟合得很好——NRMSE 完全达标——但就是慢，上升时间超过 0.2 毫秒。关键看原始波形：这些是真脉冲，不是噪声；起点对得齐，峰值来得晚。也就是说，探测器里真的存在第二种脉冲形状。一个模板只能描述一种形状——这种真实的形状差异，单模板方法天生盖不住。这正是我们做第二族模板、也就是多模板 NxM 方法的动机。

English: And here is that "shadow" from the alignment plot. These events fit well — their NRMSE is perfectly fine — but they're slow, with rise times above zero point two milliseconds. The key is the raw traces: these are real pulses, not noise. The onset lines up; the peak just comes late. In other words, the detector genuinely produces a second pulse shape. One template can only describe one shape — real shape variation like this is something a single-template method simply can't cover. And that is exactly the motivation for our second template family, the multi-template NxM approach.

Slide 10 — Side finding: Z7 PDS2 (约 30 秒)

(时间紧可跳过，说一句"我们还排除了一个单通道伪影"即可。)

中文：顺带一个小发现。Z7 有一个通道的下降时间分布特别宽，乍一看像慢物理。但把同样的事件在全部通道里展开看：其他 11 个通道里它们都是完全正常的快脉冲。所以这是那一个通道自己的低频干扰，不是物理，不影响任何形状结论。

English: One quick side finding. On Z7, one channel shows a much broader fall-time distribution — at first glance it looks like slow physics. But if you draw the same events across all channels, they're perfectly normal fast pulses in the other eleven. So it's a low-frequency disturbance in that one channel, not physics — and it doesn't affect any of the shape conclusions.

Slide 11 — Template family 1: 2-exp weighted (1x1) (约 45 秒)

中文：现在到产出。第一族模板走解析路线：每个通道，把所有物理拟合的曲线放到同一个起点，做加权平均——权重是 NRMSE 平方的倒数。这个设计的好处是：拟合差的事件权重自动变得极小，等于被压没了，但我们不需要人为剔除任何事件。屏幕上就是参与平均的这束曲线。因为平均的对象是解析函数，做出来的模板天然光滑、完全没有噪声。这就是标准的单模板，1x1。

English: Now to the deliverables. The first template family takes the analytic route: for each channel, we put all the physical fitted curves at the same onset and take a weighted average — the weight is one over NRMSE squared. The nice property of this design is that badly-fit events automatically get a vanishingly small weight, so they're effectively suppressed — but we never have to remove anything by hand. What's on screen is the bundle of curves that goes into that average. And because we're averaging analytic functions, the resulting template is smooth and completely noise-free by construction. That's the standard single template — the one-by-one.

Slide 12 — Template family 2: NxM PCA (约 50 秒)

中文: 第二族模板就是为刚才那种形状变化准备的。做法是对通过全部筛选的拟合曲线做主成分分析, PCA。直观理解: 黑色的 `nxm0` 是平均形状; 后面四条彩色的, 是数据里最主要的四个"变形方向"——比如上升更慢一点、下降更快一点。真实脉冲就用这五条的线性组合去拟合。效果非常好: 前两个成分就已经覆盖了 96 到 98% 的形状差异。也就是说, 那群慢脉冲不再是麻烦——它们被显式地表示进了模板空间里。

English: The second family is built precisely for the shape variation we just saw. We run a principal component analysis — PCA — on the fitted curves that pass all the selections. The intuition: the black curve, `nxm-zero`, is the average shape; the four colored ones are the four main "directions of deformation" in the data — say, a slightly slower rise, or a faster fall. A real pulse is then fitted as a linear combination of these five. And it works remarkably well: the first two components already cover ninety-six to ninety-eight percent of the shape variation. So the slow population is no longer a problem — it's explicitly represented inside the template space.

Slide 13 — Delivered, and what remains (约 45 秒)

中文: 总结一下。两族模板已经为全部 13 个探测器交付, 用的是官方 `cdmsbats` 的标准格式, 处理链可以直接读。整个分析每一步都可追溯: 原始缓存、每个拟合的检查点、所有诊断图, 全部有版本记录, 每张图上都印着自己的完整处理过程。两个 cut 都是从数据分布里读出来、又回到原始波形里验证过的。还剩两件事: 一是慢脉冲那个取舍需要拍板, 想听听大家的意见; 二是对新模板跑一遍组里的模板验证流程。谢谢大家, 欢迎提问。

English: To wrap up. Both template families are delivered for all thirteen detectors, in the official `cdmsbats` format, ready for the processing chain to read. Every step of the analysis is traceable — from the raw cache, to the per-fit checkpoints, to all the diagnostic figures: everything is under version control, and every figure carries its full processing history printed on it. Both cuts were read off the data, and then verified back in the raw traces. Two things remain: first, the slow-pulse trade-off needs a decision — I'd like to hear your thoughts on that. And second, running the group's template-validation procedure on the new templates. Thank you — I'm happy to take questions.

- **为什么不用实测波形的平均做模板？ / Why not average the measured traces?** 解析拟合曲线本身没有噪声，加权平均出来的模板天然光滑；NRMSE 权重让拟合差的事件自动降权。对齐叠加图上实测均值（红）和加权拟合均值（橙）可以互相对照。 / The analytic fitted curves carry no noise, so their weighted mean is smooth by construction; the NRMSE weight down-weights badly-fit events automatically. The overlays show the measured mean (red) and the weighted fitted mean (orange) side by side for comparison.
- **0.4 怎么定的，敏感吗？ / How was 0.4 chosen — is it sensitive?** 取在双峰分布的谷底，谷里本来就没多少事件——这正是选谷底的理由。还有交互 notebook，改阈值几秒内能重跑所有探测器。 / It sits in the valley of the bimodal distribution, where there are few events by construction — that's exactly why the valley was chosen. An interactive notebook re-runs any threshold within seconds.
- **两族模板的 cut 为什么不一样？ / Why do the two families apply cuts differently?** 1x1 不硬切，靠权重：好拟合 $\text{NRMSE} \approx 0.05$ 权重 400，噪声 $\text{NRMSE} \approx 2$ 权重 0.25，差四个数量级，噪声自动压没，不需要人为定一刀。PCA 对离群点敏感——一条噪声曲线就能污染主成分方向——所以 NxM 的输入必须硬切干净 ($\text{fit_ok} + \text{NRMSE} \leq 0.4 + \tau_{\text{rise}} \leq 0.3$ ms)。 / The 1x1 uses weighting instead of a hard cut: a good fit at NRMSE 0.05 gets weight 400, a noise trigger at 2 gets 0.25 — four orders of magnitude apart, so noise is automatically suppressed without choosing a threshold. PCA is sensitive to outliers — one noise curve can contaminate the components — so the NxM input must be hard-cut clean.
- **慢脉冲群体到底留不留？ / Keep the slow pulses or not?** 这是明确记录在案的待定事项。目前 PCA 的输入应用了 τ_{rise} cut；这个取舍要不要放开，正是想和大家讨论决定的。 / This is the documented open item. The PCA input currently applies the rise-time cut; whether to relax that trade-off is exactly what we'd like to discuss and decide.

术语自查表（只给自己，不进 PPT）

过一遍下面每一行；哪一条还觉得说不清楚，提前告诉我，我们把它从幻灯片里去掉或换个说法。

术语	一句话解释
zip / Z7	一个探测器（一块 Ge 晶体）；Run 4 一共 13 个，Z7 是最安静的一个
声子通道 PAS1...PES2	每个探测器的声子读出通道，S1/S2 是晶体的两个面；Z7 的 PES2 坏了所以只有 11 个
series	一段连续采数（约 1 – 2 小时），文件按 series 组织；每个探测器 27 – 30 个
MIDAS raw	DAQ 写出的原始波形数据格式，完全未经处理

术语	一句话解释
PTOFamps	官方处理链里"总声子 optimal filter 幅度", 可当作能量的代理量
K 线	Cf 活化后 Ge 探测器出现的 10.37 keV K 壳层活化线, 当单能参考源用
optimal filter (OF)	用"模板形状 + 噪声谱"对脉冲做加权拟合来估计幅度的标准方法
1x1 / NxM	1x1 = 每通道一个模板; NxM = 每通道多个模板 (均值 + 主成分), 能拟合形状变化
Butterworth 低通	一种平滑滤波器; 100 kHz 截止把高频噪声滤掉; "4 阶"指截止后压制的陡峭程度
steady-state init	滤波器初始条件取稳态值, 避免波形开头出现人为的启动瞬态
2-exp 模型	上升一个指数、下降一个指数: $y = A(e^{(-t/\tau_{\text{fall}})} - e^{(-t/\tau_{\text{rise}})})$; 5 个自由参数
pretrigger / onset 16050	波形里名义的触发位置 (第 16050 个采样点); 我们把它当自由参数拟合
NRMSE	拟合残差的 RMS $\div$ 拟合脉冲峰值; 无量纲的"拟合好坏"指标
fit_ok	物理性检查: 幅度为正且 $\tau_{\text{rise}} < \tau_{\text{fall}}$
亚采样插值对齐	平移量不是整数个采样点时用线性插值 (np.interp) 实现的平移
PCA	主成分分析: 把一堆曲线分解成"均值 + 按方差大小排序的正交变化方向"
explained variance	每个主成分解释的形状方差占比; PC1+PC2 = 96 – 98%
TH1D	ROOT 软件的一维直方图对象; 模板的交付格式 (32768 个 bin)
cdmsbats / PulseTemplates	官方处理软件 / 它读取模板的配置目录和文件命名规范
PT / PS1 / PS2	求和模板: 全部通道之和、S1 面之和、S2 面之和 (各自峰值归一)