Certificate of Calibration

Standards Laboratory Calibration

Certificate Number WO-01467777

Model Number	9334
Manufacturer	Guildline Instruments
Description	Resistance Standard
Serial Number	63433
Date of Calibration	23 Jul 2026
Procedure	24BR-CP185FS
Temperature	(23.0 ± 1.0) °C
Humidity	(35 to 55) %RH

Customer
Keysight Technologies
Bayan Lepas Free Industrial Zone
Malaysia Sdn Bhd (463532-M)
BAYAN LEPAS Penang 11900
Malaysia

Location of Calibration
Keysight Technologies Inc
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures and accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540.3-2006. The quality management system is registered to ISO 9001:2015.

As Received Conditions
The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

Action Taken
- No corrective actions were necessary.

As Completed Conditions
The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Remarks or Special Requirements

This calibration report shall not be reproduced, except in full. The documented results relate to the equipment calibrated only.

The test limits stated in the report correspond to the published specifications of the equipment, at the points tested.
This calibration report may refer to equipment manufactured by HP, Agilent and Keysight as being manufactured by Keysight Technologies. This calibration report is available at Keysight Support Portal http://support.keysight.com/.

Based on the customer's request, the next calibration is due on 23 Jan 2027.

Keysight Technologies Inc
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES



Dan Allen Roseville Serv. Cntr. Mgr.

Certificate of Calibration

Standards Laboratory Calibration

Certificate Number WO-01467777

Traceability Information

Technician ID 01029259

Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used

Model Number	Model Description	Equipment ID	Cal Due Date	Certificate Number
6625	Resistance Measurement System	SL903160	31 Jul 2026	WO-01126224
HPRESSET	Set of HP11XXXs series resistors maintain in a grouped bank	SLRESSETPT	31 Oct 2026	WO-01434875

KEYSIGHT TECHNOLOGIES

Keysight Technologies Primary Standards Lab · 10090 Foothills Blvd · Roseville, CA 95747

Standards Laboratory
Calibration Report

DATE:	23 July 2026	CERT#:	WO-01467777
ITEM:	100 MΩ Standard Resistor	MODEL:	9334
IDENT:	S/N 63433	MFR:	Guildline

This resistor was directly compared through a 6625A Resistance Measurement System Bridge to 10 MΩ resistor that is traceable to NIST.

Measurement voltage was 100 V.

NOMINAL VALUE	MEASURED VALUE	MU
100 MΩ	100.002 65 MΩ	8.1 μΩ/Ω

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded measurement uncertainty, which corresponds to a coverage probability of approximately 95%, is the standard uncertainty multiplied by the coverage factor k=2.

This report shall not be reproduced except in full, without written approval of the Keysight Technologies Primary Standards Lab.