# Certificate of Calibration

**Standards Laboratory Calibration**

**Certificate Number WO-01467781**

| | |
|---|---|
| **Model Number** | 9334A-1G |
| **Manufacturer** | Guildline Instruments |
| **Description** | 1G Ohm Precision Resistance Standard |
| **Serial Number** | 68820 |
| **Date of Calibration** | 22 Jul 2026 |
| **Procedure** | 24BR-CP184-BS |
| **Temperature** | (23.0 ± 1.0) °C |
| **Humidity** | (35 to 55) %RH |

**Customer**
Keysight Technologies
Bayan Lepas Free Industrial Zone
Malaysia Sdn Bhd (463532-M)
BAYAN LEPAS Penang 11900
Malaysia

**Location of Calibration**
Keysight Technologies Inc
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures and accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540.3-2006. The quality management system is registered to ISO 9001:2015.

**As Received Conditions**
The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

**Action Taken**
- No corrective actions were necessary.

**As Completed Conditions**
The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This probability corresponds to a coverage factor of k=2 for a normal distribution.

**Remarks or Special Requirements**

This calibration report shall not be reproduced, except in full. The documented results relate to the equipment calibrated only.

The test limits stated in the report correspond to the published specifications of the equipment, at the points tested.
This calibration report may refer to equipment manufactured by HP, Agilent and Keysight as being manufactured by Keysight Technologies. This calibration report is available at Keysight Support Portal http://support.keysight.com/.

Based on the customer's request, the next calibration is due on 22 Jan 2027.

Keysight Technologies Inc
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES



Dan Allen Roseville Serv. Cntr. Mgr.

# Certificate of Calibration

**Standards Laboratory Calibration**

**Certificate Number WO-01467781**

## Traceability Information

**Technician ID** 01029259

Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

### Calibration Equipment Used

| Model Number | Model Description | Equipment ID | Cal Due Date | Certificate Number |
|---|---|---|---|---|
| 1 G-1 | STANDARD RESISTOR | SL266512 | 23 Oct 2027 | WO-01074487 |
| 4339B | High Resistance Meter | SL901888 | 30 Sep 2026 | WO-01178744 |

KEYSIGHT TECHNOLOGIES

Keysight Technologies Primary Standards Lab · 10090 Foothills Blvd · Roseville, CA 95747

# Standards Laboratory
# Calibration Report

| | | | |
|---|---|---|---|
| DATE: | 22 July 2026 | CERT#: | WO-01467781 |
| ITEM: | 1 GΩ Standard Resistor | MODEL: | 9334A-1G |
| IDENT: | S/N 68820 | MFR: | Guildline |

This resistor was directly compared using a 4339B as a transfer standard and a calibrated 1 GΩ resistor that is traceable to NIST.

Measurement voltage was 10 V.

| NOMINAL VALUE | MEASURED VALUE | MU |
|---|---|---|
| 1 GΩ | 0.999 781 GΩ | 77 μΩ/Ω |

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded measurement uncertainty, which corresponds to a coverage probability of approximately 95%, is the standard uncertainty multiplied by the coverage factor k=2.



This report shall not be reproduced except in full, without written approval of the Keysight Technologies Primary Standards Lab.