KEYSIGHT TECHNOLOGIES

Keysight Technologies Primary Standards Lab · 10090 Foothills Blvd · Roseville, CA 95747

Standards Laboratory
Calibration Report

DATE:	17 December 2025	CERT#:	WO-01261128
ITEM:	Precision Resistance Standard	MODEL:	9334A-0.1
IDENT:	S/N 73137	MFR:	Guildline

This resistor was directly compared through Guildline 6625A Measurement System to a Thomas One Ohm resistor that is traceable to NIST.

Measurement current was 1 A.

NOMINAL VALUE	MEASURED VALUE	MU
0.1 Ω	.100 000 761 Ω	0.97 μΩ/Ω

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded measurement uncertainty, which corresponds to a coverage probability of approximately 95%, is the standard uncertainty multiplied by the coverage factor k=2.

This report shall not be reproduced except in full, without written approval of the Keysight Technologies Primary Standards Lab.