

Supplemental Instructions for Acton Research Corporation FC-446-010-FC Mounting Plate

This Mounting Plate is designed for ARC SpectraPro Instruments and accepts an FC light guide.

Caution

Care should be taken when handling fiber optic bundles, as excessive bending, crushing, or pulling may cause the fibers to break and become permanently damaged.

1. Remove the vertical beam baffle from the slit by removing the four binder head screws.
2. Place the Mounting Plate, with the label "TOP" or red dot at the top, against the slit of the instrument. Insert the four (4) #8-32 screws (provided) into the mounting holes and tighten them using a 9/64" allen wrench.
3. Attach the fiber bundle "FC" connector to the mating connector on the rectangular piece. **Note:** The light guide must not touch the slits as it will effect the slit setting and damage the slits.