

Comparative Efficacy of PET/CT and PET/MR in Precision Radiotherapy for Liver Cancer: Advances in Dosimetry and Tumor Delineation

Daniel Juarez

December 9, 2024

1 Introduction

During all stages of radiotherapy, getting the sense of where health professionals focus their work and the methods used to shape or conform the target tissue is of most vital importance. This information is provided by different imaging devices and techniques molded to each different case. During planning, images help visualize target lesions and clarify which pathway and devices are best to use. This information serves as a guide to determine patient and device position. During treatment, it provides insight into treatment outcomes and potential complications [1]. Finally, during surveillance or control, it gives the comparison point to check accuracy and success while pointing to the right direction to follow.

The preferred imaging modality for all types of cancer diagnosis is Computed Tomography (CT), this is due to its high-resolution anatomical detail and the information provided to map tissue electron density [1, 2]. However, like every imaging modality, CT has its limitations, emphasizing the need for complementary imaging techniques, especially for precise tumor targeting in complex cases like liver cancer [2].

For these reasons, no single imaging modality is entirely sufficient for an absolute precise role in radiation therapy due to intrinsic trade-offs [1]. Setting CT aside, another good alternative is Magnetic Resonance Imaging (MRI), although valuable for soft

tissue delineation, it lacks electron density information, making them less suitable as a standalone modalities in radiotherapy [1]. Another alternative is Positron Emission Tomography (PET), which provides a new lens since it contributes with unique metabolic information that complements anatomical imaging and aids in tumor delineation [1].

Exploring new techniques, technologies and pathways of using existing devices are carefully studied. Hybrid techniques are of most interest such as PET/CT and PET/MR. These modalities aim to overcome limitations by combining the strengths of each modality.

The objective of this paper is to compare PET/CT and PET/MR in liver cancer radiotherapy, focusing on their applications in tumor delineation and dosimetric planning, and patient monitoring. An evaluation of their technical capabilities, clinical integration, and implications for adaptive radiotherapy workflows will be included. The aim is to explore how these hybrid modalities can enhance precision in liver cancer management.

Table 1 provides a comparative overview of the relevant used imaging modalities in radiotherapy, it includes technical complexity and workflow integration that may be overlooked. Each modality has specific roles, because of it no single technique is entirely sufficient for precise tumor targeting. Following the table an explanation of each modality and why hybrid approaches are needed is explained.

Modality	Diagnostic Information	Soft Tissue Resolution	Radiation Dose	Technical Complexity	Workflow Integration
CT	Anatomy and tissue electron density	Moderate (1 mm spatial resolution)	Ionizing radiation (higher dose)	Widely available; simple infrastructure requirements	Integrates easily into standard radiotherapy workflows; rapid acquisition
MRI	Anatomy and function	Excellent (soft tissue contrast)	None (non-ionizing)	Requires trained personnel; specialized infrastructure	Long scan times; challenges with integration into dose planning workflows
PET	Metabolic and functional imaging	Poor (blurred edges, partial volume effects)	Ionizing radiation (tracer-based)	Limited isotope availability; moderate complexity	Requires co-registration with CT or MRI;
PET/CT	Combines anatomical and functional imaging; provides BTV and GTV delineation	Moderate	Higher due to combined CT dose	Standardized; relatively simple calibration	Commonly used in clinical settings
PET/MRI	Combines superior soft tissue contrast with metabolic imaging; pseudo-CT for dose calculation	Excellent	Reduced radiation compared to PET/CT	High complexity; limited infrastructure and operator expertise	Workflow integration less standardized compared to PET/CT

Table 1: Comparison of imaging modalities (CT, MRI, PET, PET/CT, and PET/MRI) for radiotherapy applications. Adapted from Decazes et al. [1], Yan et al. [2], and Knešaurek et al. [3].

First, CT is the standard modality used in radiotherapy due to its high anatomical resolution and integration into treatment workflows. However, it relies on ionizing radiation and it has moderate soft tissue resolution that limits its standalone use in complex cases, such as tumors adjacent to sensitive soft tissues.

Next, MRI provides a solution to CT soft tissue resolution and adds functional imaging without ionizing radiation, making it ideal for delineating tumors in challenging areas. Nevertheless, its known for long scan times and the difficulty to integrate electron density information of the tissue on the treatment plan.

Then, PET offers a unique insight on tumor activity, that is functional and metabolic imaging. Still, its limitations are quite considerable, mainly poor spatial resolution which makes this modality rely on co-registration with anatomical modalities like CT or MRI.

Following that, PET/CT solves these limitations by combining the anatomical clarity of CT with the functional insights of PET. Yet, the combined radiation dose remains a concern.

Finally, PET/MRI merges superior soft tissue delineation with metabolic imaging. This modality reduces radiation exposure compared to PET/CT but faces significant challenges in technical complexity, cost, and workflow integration.

Imaging modalities are complementary in nature because there are trade offs and risks on each of them. While CT and MRI is great in anatomical and soft tissue imaging, respectively, and PET provides metabolic insights, their integration through hybrid systems like PET/CT and PET/MRI is essential for achieving optimal precision in radiotherapy.[1, 2]. Table 1 convey the importance of advancement in hybrid techniques to overcome the limitations of individual modalities and confidently approach any complex clinical scenarios with either IMRT or VMAT as treatment.

2 Background

Liver cancer is particularly difficult to diagnose and its treatment requires highly precise imaging. It needs accurate tumor delineation and because of its size and location spare healthy liver tissue, and organs at risk (OAR). This means radiologists need accuracy, especially when administering targeted ther-

apies [4, 5].

Around the world there is a common concern of cancer incidences. Liver cancer, in particular, ranks as the sixth most common diagnosed cancer and the third leading cause of cancer-related deaths. In 2020, approximately 905,700 new cases and 830,200 deaths were reported globally [6].

The global burden varies significantly by region, for example, East Asia and sub-Saharan Africa have the largest incidence rate while in the United States liver cancer rates are comparatively lower. Global predictions estimate that this rates will increase by 55% between 2020 and 2040, potentially leading to 1.4 million new cases and 1.3 million deaths annually by 2040 [6].

The American Cancer society projected approximately 41,630 new cases and 29,840 deaths in the United States [7]. As the year concludes we may expect to see the actual increase in new cases, however there is a historical increasing trend, the United States have seen, between 1975 and 2017, an increase of 228% incidence and 137% rise in mortality. Probably driven by a combination of rising risk factors, aging populations, and improved detection. However, despite advancements in imaging and treatment, late diagnoses and limited therapeutic options have contributed to persistently high mortality rates.

Moreover, projections made in 2020 suggest that liver cancer could become the third leading cause of cancer-related deaths in the US by 2030, surpassing breast, prostate, and colorectal cancers [8]. This is particularly interesting, since this has already happen globally, the slower growth of cases and mortality in the U.S. suggests that public health efforts are having a meaningful impact, continued advancements in imaging, personalized therapy, and public health interventions are needed to sustain and further improve outcomes.

The liver is the largest solid organ located in the upper right quadrant of the abdominal cavity, beneath the diaphragm, its roles includes metabolism, detoxification, and digestion. It is partially shielded by the rib cage, making it relatively protected. To better illustrate its position and function figure 1 shows the abdominal cavity with the liver's relationship to adjacent organs such as the stomach, diaphragm, and intestines [9, 10].




Figure 1: Anatomical Relations of the Liver

It has a complex structure, it has two main lobes (right and left), the right lobe is significantly larger, and its main function is metabolic and detoxification processes. The left lobe, is smaller, and it is in charge of production and nutrient metabolism [10].

The liver can be shaped differently between individuals. As shown in Figure 2, the anterior surface of the liver can be classified into four primary shapes.

First, **Globular Shape (A)** A rounded configuration typically seen in healthy individuals with an average body. Next, **Conical Shape (B)** Characterized by a narrow right lobe and a smaller left lobe, often observed in elongated torsos. Then, **Quadrilateral Shape (C)** Exhibits a broader and flatter right lobe with minimal reduction, common in individuals with higher abdominal fat deposition. Finally, **Rectangular Shape (D)** A flat and wide configuration with a reduced height-to-width ratio [10]




Figure 2: Anterior surface of the liver classified into four primary shapes: (A) Globular, (B) Conical, (C) Quadrilateral, and (D) Rectangular.

Morphological variations directly impact diag-

nostic accuracy, treatment planning, and therapeutic outcomes. Changes in liver shape due to conditions such as cirrhosis or hepatic steatosis further complicate the diagnostic process, as these conditions alter the liver's texture and vascular patterns. This can make it challenging to distinguish between benign and malignant lesions on imaging modalities like CT or MRI. Then, the liver is divided into eight segments by functionally, each segment is independent, receiving its own blood supply from a branch of the hepatic artery and portal vein and draining bile through a corresponding biliary branch. This segmentation allows precision on diagnostics, surgery and transportation and is depicted on figure 3 [11].




Figure 3: Segments of the liver according to Couinaud's classification: (A) ex vivo appearance; (C) coronal volume-rendered abdominal CT image with annotated right hepatic lobe segments. PV – portal vein; IVC – inferior vena cava.

The segmented and lobular structure of the liver makes imaging harder. Lesions can be difficult to localize within segments because of overlapping vascular and biliary structures. Small tumors or anomalies in deeper segments like the caudate lobe, which is encircled by other liver regions and organs, may not

be detected. [12]

In radiotherapy, the morphological irregularities, may shift the relative position of tumors within the liver. This necessitates advanced imaging techniques, to precisely map the tumor's location relative to moving anatomical structures like the diaphragm during the breathing cycle. [13]

Moreover, the liver's segmented anatomy and its dual blood supply necessitate careful dosimetric planning to optimize radiotherapy. Variations in lobe size affect the volume of liver tissue exposed to radiation. Additionally, exaggerated organ motion during respiration require adaptive radiotherapy techniques, such as respiratory gating or motion management systems, to ensure accurate dose delivery.[14]

Dosimetric planning techniques, such as Intensity-Modulated Radiation Therapy (IMRT) and Volumetric Modulated Arc Therapy (VMAT), take these variations into account by adapting dose distribution to the unique anatomy of each patient, sparing healthy tissues and minimizing complications. [15]

Diagnosing cancer is a challenge also due to its nature, in particular liver cancer as it presents asymptomatic in early stages, for example hepatocellular carcinoma (HCC) often develops in patients with chronic liver disease or cirrhosis, which can mask early signs of cancer and further complicate detection [16, 17].

In the other hand there is limitations of standard screening tools, following the HCC example, usually tested with ultrasound and often combined with alpha-fetoprotein (AFP) test. AFP is a serum biomarker commonly used in HCC screening and diagnosis. AFP level increased may suggest underlying pathology, which may be malignant [18]. However, ultrasound and AFP tests have limitations. Ultrasound sensitivity is hindered particularly in patients with obesity or cirrhosis.[5] Studies have shown that ultrasound can miss more than half of early-stage tumors, while the addition of AFP testing still fails to detect over one-third of early HCC cases [19].

Recent years have seen advancements in imaging techniques by combining standard techniques such as CT, MRI, and cone-beam CT (CBCT) to enhance tumor visualization.[5].

Imaging the liver is quite difficult due to its location beneath the rib cage, the impact of respiratory motion, and patient-specific factors or comorbidities.

The liver's heterogeneity and its internal complexity further complicate image interpretation. [20].

Imaging is useful during all stages of treatment. Because of the liver's intricate structure and common comorbidities such cirrhosis, it is important to accurately define tumor borders, determine vascular involvement, and estimate closeness to other tissues [5].

The most widely used imaging modality to assess liver cancer is ultrasound (US) due to its accessibility, real-time imaging capability, ability to assess blood flow via Doppler techniques, and because of its non-ionizing nature. But, it is limited by operator skill and sensitivity while detecting small lesions, in early-stage hepatocellular carcinoma (HCC) the smallest detectable lesions typically range from 3 to 5 mm in diameter. [19]

For instance, small hepatic adenomas, which are benign solid lesions, often remain undetected until they grow larger than 5 cm. Similarly, early-stage intrahepatic cholangiocarcinoma, a bile duct cancer, and small hepatic cysts, which are fluid-filled sacs, are frequently missed unless they become symptomatic or increase in size [21].

It is hard to diagnose because of two main reasons. The quality of contrast-enhanced ultrasound (CEUS) images heavily depends on the baseline US image, which can be affected by factors such as lesion depth, location (e.g., subdiaphragmatic). Additionally, CEUS typically allows for the evaluation of only one focal liver lesion (FLL) at a time, increasing the risk of overlooking additional small lesions [22].

To overcome this limitations there are hybrid imaging techniques, for example US fused with cone-beam CT (CBCT) or MRI, have significantly enhanced tumor visualization and treatment precision. They together provide better spatial resolution and enhance the vascular anatomy, or increased soft tissue contrast.[5]. For instance, US provides excellent spatial resolution for superficial structures, while CBCT and MRI offer high-resolution 3D imaging of deeper tissues, improving the delineation of tumor boundaries [23].

US hybrid approaches also enhance vascular anatomy visualization by combining color Doppler US with contrast-enhanced CT or MRI, particularly useful on planning interventional procedures [24]. Moreover, real-time imaging capabilities of US allow for dynamic guidance during procedures. When fused with pre-acquired CBCT or MRI data, its use

enhances precision and safety during interventional therapies [25].

There is potential to be discovered from hybrid imaging modalities in cancer treatment, that is why now we see increased interest in exploring PET/CT and PET/MR as advanced tools in precision radiotherapy.

Despite significant advancements in imaging technology, further improvements are needed to address intrinsic limitations of CT, MRI, and US in tumor delineation dosimetry, and planning treatment with precision.

CT, offers great spatial resolution but is limited by poor soft tissue contrast and delimitation, it is an irradiating imaging modality and it is susceptible to artifacts caused by metal implants or dental structures [1]. MRI provides better soft tissue contrast and functional imaging capabilities, yet it lacks the electron density information critical for accurate dose calculations in radiotherapy. Additionally, the extended acquisition times of MRI can introduce motion artifacts, further complicating tumor visualization [5].

PET, which provides metabolic and functional

data, is often combined with CT or MRI to enhance tumor characterization. However, PET alone suffers from low spatial resolution and partial volume effects, leading to blurred tumor edges and suboptimal delineation of small lesions [2]. These limitations highlight the need for hybrid imaging techniques that leverage the complementary strengths of multiple modalities.

Hybrid approaches such as US/CBCT and CT/MRI fused with US have shown promise in addressing these challenges. For example, CBCT can enhance vascular anatomy visualization in patients with cirrhotic livers, aiding in tumor detection and feeding vessel identification during transarterial treatments [5]. However, the application of these techniques in radiotherapy remains limited due to technical constraints such as the need for precise attenuation correction and standardized protocols for image registration. Hybrid imaging modalities such as PET/CT and PET/MR are emerging tools for overcoming these barriers, offering improved tumor delineation, dosimetry accuracy, and treatment planning for liver cancer. [3, 26]

3 PET/CT and PET/MR in Liver Cancer Therapy

A particularly interesting imaging technique that is on the rise is PET/MR, it offers some advantages in the delineation of liver tumors where anatomical clarity is crucial. It is worth exploring target definition and motion control on MR's increased contrast when treating a highly dynamic organ like the liver. As well as some of its drawbacks like longer collection periods and more complicated attenuation correction. A natural comparison point is PET/CT which has positioned itself as the gold standard for radiation planning due to its reliable attenuation correction, rapid acquisition times, and accurate dose calculations. PET/MR, however, has demonstrated higher sensitivity and specificity for detecting liver metastases, finding additional metastatic lesions missed by PET/CT, and significantly reducing radiation exposure. [27]

An additional benefit of PET/MR is its ability to significantly reduce radiation exposure, however limitations in attenuation correction can impact its performance, and can result in underestimation of absorbed doses during planning. [28] These issues can complicate radiotherapy workflows, potentially

leading to suboptimal treatment plans that either underdose tumors or overdose surrounding healthy tissues [29].

In order to provide a comprehensive view of its therapeutic potential, this section examines the complementary function of PET/MR in radiotherapy, stressing its advantages, disadvantages, and comparable applications with PET/CT.

3.1 PET/CT

Positron Emission Tomography/Computed Tomography (PET/CT) is utilized for precision in radiotherapy, it has great capabilities combining functional and anatomical information into a single session. It is really useful for tumor delineation, dosimetry, and treatment monitoring.

In order to reduce misalignment from repositioning and guarantee excellent spatial and functional precision during radiation processes, integrated PET/CT systems are built with a shared gantry frame with both imaging modalities. The initial designs of these systems combined a spiral CT

scanner with a rotating PET scanner [30]




Figure 4: Schematic representation of TOF-capable PET/CT scanner with operational depiction of individual. Adapted from Mohammadi [31]

The device can be operated separately or together within a single framework. In combined mode, CT images are employed for attenuation correction of PET data. When PET is used alone, it usually requires separate transmission scans used to provide anatomical reference and attenuation correction for images. Having them together improves imaging quality while reducing examination time from additional scans [32].

Clinical studies have demonstrated that even with shorter acquisition times, such as 1.5 minutes per bed position, image quality remains clinically acceptable with high rates of lesion visualization. PET/CT is a viable option for radiation workflows that are time-sensitive due to its efficiency and quality balance [33].

Modern PET/CT systems use advanced image registration techniques to improve data alignment and reduce errors. Techniques for Deformable Image Registration (DIR), including GPU-accelerated DIR, take into consideration non-rigid changes brought on by patient movements or breathing patterns.

Deformable Image Registration (DIR) methods such as, GPU-accelerated DIR, account for non-rigid transformations caused by patient motion or breathing patterns. These techniques model respiratory motion, improving the quantification and localization of functional uptake in PET imaging. In liver cancer treatment, where organ mobility presents major obstacles, such developments are very helpful [34].

Additionally, this technique is considered upon its ability to detect non-FDG uptake lesions, contrasting with PET and helping avoid misdiagnosis in complex cases such as liver tumors [2, 1]. This

is done with segmentation algorithms that establish SUV thresholds to delineate lesions, integrating these volumes into treatment plans. Then, combining PET/CT with multiphasic CT can further characterize non-FDG lesions by leveraging dynamic contrast patterns visible in CT [35].

Accurate dosimetry is dependent on quantitative imaging metrics such as standardized uptake values (SUVs). One of the advantages as explained in AAPMs TG174 and other papers is attenuation correction and faster acquisition times.[3, 35].

This system comes pre-integrated and aligns PET and CT datasets in the same frame of reference via software from the manufacturers and allows to acquire images in a single session ensuring precise alignment between functional and anatomical data. This reduces registration errors, ensuring precise tumor localization and dose delivery.

CT imaging contributes with the electron density information necessary for attenuation correction in PET imaging. Since CT operates at lower photon energies (typically 80–140 keV), its data must be converted to match the 511 keV photons detected by PET. This calibration creates attenuation maps that align with PET imaging physics and are applied during PET image reconstruction to adjust the signal intensity, producing PET images that accurately represent metabolic activity without distortion due to tissue absorption.[3, 35]

In addition to attenuation correction, PET/CT facilitates tumor segmentation based on metabolic activity by converting PET data into standardized uptake values (SUVs). This information can then be integrated into radiotherapy planning systems to optimize treatment

Even with this seamless integration of the devices, respiratory motion and artifacts, particularly in liver cancer cases, can affect the accuracy of dose distribution. In TG174 we can find some techniques to mitigate respiratory motion artifacts in liver imaging as well as quality assurance (QA) protocols [35].

For example, TG174 highlights the use of respiratory gating to reduce motion artifacts caused by the liver's movement during the breathing cycle. Or in the absence of respiratory gating, a time-weighted average intensity projection can mitigate the blurring effects of motion in PET intensity values.

Aside from this monthly uniformity, quantification, and alignment tests between PET and CT datasets and annual Testing with phantoms, such as

the ACR flangeless PET phantom, assesses PET/CT alignment and SUV accuracy are recommended.


Moreover, PET/CT tends to show lower SUV values compared to PET/MR, which may lead to underestimation of lesion uptake [36]. This underestimate can result in missed lesions, underestimation of disease severity, and incorrect staging of tumors. These limitations are particularly significant for liver imaging, where PET/MR has demonstrated increased diagnostic accuracy, with a 14.6% improvement compared to PET/CT[36].

In a case study, Knesaurek et al. [3] reported differences in liver volume estimation between PET/CT and PET/MR, with a significant impact on dosimetry calculations. This means that by choosing the correct imaging system for planning can make a huge difference.


3.2 PET/MR

On the flip side, a relatively new imaging modality has been catching some interest, and could provide complementary benefits to PET/CT in radiotherapy. PET/Magnetic Resonance Imaging (PET/MR), while not as established as PET/CT can still provide soft-tissue contrast and functional imaging capabilities that are quite attractive for specific liver cancer cases.


This integration allows the recording of functional and anatomical information concurrently. The three prime strategies for combining PET/MR imaging include the following that are sequential, insert an integrated systems. In the systems that are sequential, there arrangement whereby a patient is allowed to be scanned first in pet scanner and subsequently in an MR-type scanner; this is easily enabled by having a prepared mechanical bed for moving the patient. This modality combination is easy but does not allow simultaneous imaging. Misalignment and space are certainly an issue. In insert systems, a PET detector ring is inserted into the bore of an existing MR scanner. This approach allows simultaneous acquisition, superior spatial, and temporal alignment; however, several technical problems, such as magnetic field interference and limited bore size, make it suitable for rather specific applications, like neurological studies [37].



(a) Sequential whole-body PET/MRI design using table shuttle for patient transportation.



(b) Schematic diagram of the APD-based brain PET insert with 3 T MRI and a photo of the brain PET insert combined with a slightly modified clinical 3 T T



(c) Schematic diagram (left) and photo (right) of the first commercially available fully integrated whole-body PET/MR with APD technology.

Figure 5: Comparison of different PET/MRI configurations: (a) Sequential system, (b)Insert-based brain PET system. , and (c) Fully integrated system. Adapted from [38].

Integrated PET/MR systems are defined as the direct integration of the PET detectors inside the MR scanner. Most common embodiments place the PET detector ring behind the radiofrequency coil of the MR scanner using a split superconducting magnet or field-cycled MR in order to be able to house the PET components. PET detectors in PET/MR systems are prepared by using materials that will not perturb the magnetic field and degrade the quality of MR images. Situating the detectors at the rear of the radiofrequency coil increases temperatures and may cause their oscillations. This is important because the gain of both APDs and SiPMs-commonly used in the construction of PET systems-depend on

temperature [37]

Having better soft tissue contrast in the abdominal area aids the planning process by accurately delineating liver tumors from healthy tissue. Instead of relying on CT attenuation correction, MR benefits the patient with reduced radiation dose and better delineation of anatomical structures [3]. This is really valuable for cases with small or poorly defined tumors.

Additionally, PET/MR's ability to image respiratory liver motion during PET acquisition provides more reliable data for motion management compared to PET/CT [3]. High-resolution MR images can also assist in partial volume corrections for PET images, further enhancing the precision of tumor targeting, but lacks electron density making attenuation correction challenging.

PET by itself comes with functional data from the activity distribution, but MR will enhance this capabilities if combined with techniques like DWI (Diffusion-Weighted Imaging) and dynamic contrast-enhanced (DCE) imaging, for additional parameters on tumor metabolism and treatment response.

Diffusion-Weighted Imaging (DWI): This technique assesses the movement of water molecules within tissues by measuring the Apparent Diffusion Coefficient (ADC). Tumors often exhibit lower ADC values due to high cellularity and restricted diffusion caused by cell membranes. Treatment-related changes in ADC values can reflect alterations in cellularity, serving as an early marker of therapeutic response [39]

Dynamic Contrast-Enhanced (DCE) MRI: DCE MRI tracks the distribution of contrast agents over time, providing valuable insights into blood flow and vascular integrity. Changes in perfusion patterns frequently correlate with treatment response,

as increased perfusion often indicates effective therapy [40].

Effective treatments typically lead to increased ADC values, reflecting reduced cellularity as tumor cells die. These changes can be observed within days to weeks after chemotherapy or radiotherapy. However, early treatment may cause transient decreases in ADC due to cellular swelling or vascular changes, necessitating careful interpretation [39] Combining DCE MRI with PET data creates a multiparametric approach, integrating metabolic activity from PET with blood flow characteristics from DCE MRI. This synergy enhances the understanding of tumor behavior and supports personalized therapeutic strategies [41]

MR is already a solid tool for advance planning as explained in TG284[42]. Despite these advantages, the longer acquisition times and lack of specific guidelines for PET/MR integration by AAPM make this technique rise questions. However, its use may be justified in cases where soft-tissue contrast is critical for treatment success.

PET/MR contribute in a unique way to the planning and staging part, the following sections are to explore how this is implemented with radiation therapy and what is the overall effect on target delineation and dosimetry.

3.3 PET/MR usage on Radiation Therapy

Figure 6 shows the simplified workflow made possible by PET/MRI, compared to conventional radiotherapy procedures. According to Yan et al. [2], PET/MRI greatly improves efficiency, eliminates the need for multiple devices, and minimizes mistakes from co-registration processes by integrating imaging and positioning steps into a single operation.




Figure 6: Comparison of PET/MRI workflows for radiotherapy procedures with conventional radiotherapy procedures. Adapted from Yan et al. [2].

According to Decazes et al., adopting PET/MRI into radiation procedures needs careful planning and adjustments, that include using synthetic CT (pseudo-CT) for volume delineation and attenuation correction. By immobilizing the patient, PET/MR allows for alignment between imaging modalities without the need for further registration procedures. In order to provide precise dosimetry planning without the need for several imaging devices, artificial intelligence methods, such GANs, further optimize attenuation mapping and shorten the pre-treatment routine [1].

3.4 Positioning, Planning, and Pretreatment Workflow

Specific Equipment and Material In order to properly use PET/MR in the clinical setting, challenges regarding patient positioning need to be addressed due to PET/MR’s unique design and functional requirements compared to PET/CT. The following discussion is primarily informed by Yan et al. [2]. Regarding equipment, a regular curved diagnostic MR couch is unsuitable for PET. Instead, the use of flat treatment couch allows consistent patient positioning across imaging and treatment sessions while minimizing photon attenuation. Photon attenuation is further addressed in the materials used. Conventional carbon fiber tables, while minimally attenuating photons, can create surface currents that degrade MR image quality. Hybrid materials, such as plastic-foam sandwiches, have been developed to

reduce photon attenuation and eliminate artifacts. Additionally, thin plastic shells and lightweight coil technologies are being explored to optimize attenuation correction (AC) without compromising image quality. [2, 37]

Effect on Image Quality The integration of these devices and materials influences the signal-to-noise ratio (SNR) and overall image quality. Phantom studies have shown negligible effects on SNR between flat and curved tables. However, positioning devices increase the distance between the patient and MR coils, reducing SNR by approximately 25%. This reduction does not significantly impact target delineation accuracy. Techniques such as increasing the signal average, lowering acceleration factors, such as the reduction of parallel imaging acceleration, can also enhance SNR, or modifying echo times (e.g., by optimizing the echo spacing for a given tissue type) can improve image contrast and mitigate SNR loss, though these adjustments may increase acquisition time. [2, 43]

Attenuation Correction Due to the lack of CT in these devices, attenuation correction is a major challenge. For example, rigid hardware such as coil holders and flat tables are usually accounted for using CT-based 3D AC maps or isotope-based attenuation maps; this is no longer the case for PET/MR. There is a need to introduce coil fixation devices that mitigate errors caused by the variable position-

ing of flexible RF coils during scanning to improve reliability of AC maps. Moreover, there is a bone segmentation limitation. MR-based AC often excludes bone due to the low intensity of bone signals in MR images. Atlas-based methods, which use pre-existing templates of CT anatomy to generate a pseudo-CT for attenuation correction, can incorporate bone structures but may introduce registration errors in regions with varying stiffness, such as the abdomen.

Overall, consistent positioning is critical to ensure reliable imaging and proper AC. Studies using phantoms have demonstrated minimal deviations in accuracy during multiple repositioning with RF coil holders. Additionally, PET/MR simulator tables allow single registration for fixed tabletop setups, ensuring repeatability in radiation therapy workflows.

Target delineation Tumor delineation in radiotherapy involves the definition of at least 3 different volumes, gross tumor volume (GTV), clinical target volume (CTV), and planning target volume (PTV). PET/MR offers unique advantages over conventional imaging modalities, particularly in refining GTV boundaries by integrating both morphological and biological data.

MRI alone is commonly used to outline the GTV based on tumor morphology, while PET provides additional biological insights that mitigate the risk of marginal misses. Studies cited in Yan et al. demonstrate Zhang et al.'s [44] observation that PET imaging contributed to an approximate 10% increase in tumor volume detection not captured by MRI alone.[2] This finding better explains tumor sizes measured under a microscope, it is also supported by comparisons of Dice Similarity Coefficient (DSC), a metric used to assess the overlap between imaging-based tumor delineations and the ground truth, he explains.

This difference on the outline of the GTV is shown in figure 7, while it focuses on nasopharyngeal cancer, it provides valuable insights into the comparative performance of PET/MRI and PET/CT for tumor delineation. Soft-tissue contrast of PET/MRI is especially relevant for complex tumor regions, such as the liver.




Figure 7: Gross tumor volume (GTV) delineation across PET/CT and PET/MRI modalities for a 69-year-old female with nasopharynx cancer. (a) The blue line represents GTV-PET/CT. (b) The pink line represents GTV-PET/MRI. Adapted from Yan et al. [2].

The addition of PET provides some advantages as well, PET's ability to capture tumor biology introduces the concept of Biological Target Volume (BTV), which integrates metabolic and functional characteristics of tumors. Using BTV data, radiotherapy can deliver higher doses to treatment-resistant regions while sparing sensitive areas. This approach, known as "dose painting" or "dose sculpting," aims to increase cure rates without increasing late radiation-induced toxicity. [45]

Of course there are limitations, starting from manual tumor delineation, though widely used, it introduces significant variability. To address this, adaptive thresholding based on individual standardized uptake values (SUVmax) and automated delineation methods are being explored. For ex-

ample, PET/MRI's co-segmentation methods have been shown to match observer performance in target delineation while providing more consistent results across repeated evaluations [2]. And regarding its comparison to PET/CT, PET/MR has reduced SUV_{max} values that may lead to conservative tumor boundaries.

The integration of PET in radiotherapy planning has led to significant alterations in radiation treatment volumes for 30%-60% of non-small cell lung cancer (NSCLC) patients. [46]. PET/CT has improved stereotactic body radiation therapy (SBRT) treatment in NSCLC patients by advancing the development of prognostic PET/CT radiomic signatures [47].

Dose Calculation and Optimization Once the target volume is delineated, the next step of planning is dose calculations. For PET/MR there are technical considerations compared to PET/CT, particularly in the context of SBRT, where precision in dose distribution is everything.

The main challenge presented on PET/MR is that it lacks the capability to directly acquire tissue electron density values. As of right now workflows include co-register PET/MR with CT images to generate dose prescription maps. Studies found that pseudo-CT techniques generated from MR images have high accuracy, and reported negligible mean absolute errors of 0.17 ± 0.12 Gy in dosimetry analysis [2].

However, PET/MR offers significant advantages for SBRT planning, particularly in reducing radiation exposure. Time-of-Flight (TOF) PET/MR systems require only 35% of the activity concentration needed for TOF-PET/CT while maintaining comparable image quality, translating to a theoretical dose reduction of up to 65%, with clinical observations suggesting realistic reductions exceeding 50% [48]

3.4.1 Mid-Treatment Workflow

In accordance with Yan et al., [18F]-FDG PET/MRI mid-treatment assessments offer essential details on tumor response using metrics that include Δ SUV_{max} and Δ D_{min}. Their function is to provide predictions for treatment sensitivity and risk of recurrence. Significant changes in water transport and glucose metabolism were noted throughout

treatment, allowing for adaptive therapy modification to prevent toxicity or accelerated tumor development [2].

A way to take advantage from this preparation, incorporating PET-based absorbed dose map into mid-treatment workflows for SBRT or SIRT is possible. Polan et al. (2023) studied how sequentially integrating ⁹⁰Y-PET-based dose distribution clinicians were able to refine treatment plans, enhancing the precision of dose delivery and adapting to dynamic tumor changes during therapy [48].

Additionally, studies have shown that advanced techniques, such as Intensity-Modulated Radiation Therapy (IMRT) and Volumetric Modulated Arc Therapy (VMAT), particularly benefit from advanced imaging techniques by leveraging real-time anatomical and functional imaging to adjust dose distributions dynamically [49].

Referencing back to Figure 6, after using PET/MR and synthetic CT to perform treatment planning and dosimetric calculations, Cone-Beam Computed Tomography (CBCT) and MRI-guided Linear Accelerators (MRI-LINAC) are suggested for mid-treatment assessment. CBCT provides high-resolution imaging of anatomical structures, enabling precise verification of patient positioning and tumor morphology before and during treatment. By capturing volumetric data, CBCT facilitates precise target localization and minimizes setup uncertainties, a key requirement for the high precision demanded in SBRT [50].

MRI-LINAC, on the other hand, combines the superior soft-tissue contrast of MRI with real-time radiation delivery. This integration allows clinicians to continuously monitor tumors and surrounding tissues with unparalleled clarity, enabling dynamic adjustments during treatment. They also are methodologies for adaptive therapy workflows. For SBRT, sub-millimeter precision is critical, and MRI-LINAC ensures that the radiation dose is confined to the target while sparing adjacent organs at risk. The ability to visualize and adapt to changes in tumor size, shape, and position during therapy enhances the precision and safety of treatments. Furthermore, MRI-LINAC supports adaptive planning workflows, as demonstrated by Lei et al. (2020) and Cunningham et al. (2022), where the incorporation of synthetic MRI and cine MRI imaging techniques have significantly improved tumor tracking and dose conformity in SBRT [51].

4 Discussion

Better tumor delineation can be achieved with the use of hybrid imaging modalities. They provide unique features like the integration of anatomical and biological data that allows for improvements. More precise treatment planning is possible as this approach advances; this is especially true for body parts with intricate tumor borders or areas that move a lot, like the liver.

By capturing metabolic and vascular behavior, the combination of functional imaging modalities including Diffusion-Weighted Imaging (DWI) and Dynamic Contrast-Enhanced (DCE) imaging allows for improved tumor characterisation in PET/MR.

By considering the potential of new and linked technologies on liver cancer management will transform therapeutic approaches, soft tissue contrast and functional imaging of PET/MR is really promising. Cases with small, poorly defined and unresectable tumors will be benefited from these advantages and developments. As PET/MR becomes more widely adopted, its role in adaptive radiotherapy workflows and personalized treatment regimens could improve survival rates and quality of life for liver cancer patients.

The idea of biological target volumes (BTB) is also introduced by the incorporation of PET/MR into radiation operations, opening the door to sophisticated techniques like dose painting. This approach allows higher doses to be delivered to treatment-resistant regions while sparing healthy tissue, ultimately improving therapeutic outcomes. Additionally, pseudo-CT creation in PET/MR can simplify processes, cutting down on treatment planning time and complexity.

5 Limitations and Future Directions

Despite its advantages, PET/MR has logistical and technical difficulties. The most obvious ones are the extended acquisition times and higher costs. Second to that, consider the limited availability in clinical settings and integration with current tech-

niques. Furthermore, the lack of standardized attenuation correction methods and sensitivity to motion artifacts is a huge deterrent. Addressing these issues will require advancements in imaging technology, standarized protocols and better workflow integration.[52, 53]

To fully realize the potential of PET/MR in radiotherapy, comprehensive guidelines from professional organizations like AAPM are necessary. These guidelines (like they do for PET/CT) should address dosimetry protocols, motion management techniques, and the development of pseudo-CT methods for attenuation correction or alternatives. Further research is needed to establish the clinical efficacy of PET/MR in routine practice and optimize its role in radiotherapy workflows.

6 Conclusion

This paper compares the advantages and limitations of PET/CT and PET/MR in liver cancer radiotherapy, it has a focus on their impact on tumor delineation, dosimetry, and treatment planning. PET/CT remains the gold standard for many clinical applications and PET/MR demonstrates significant potential in improving targeting accuracy and providing superior soft-tissue contrast and additional functional imaging capabilities, such as Diffusion-Weighted Imaging (DWI) and Dynamic Contrast-Enhanced (DCE) imaging.

Dose painting and other adaptive radiotherapy methods are made possible by the special potential for improved tumor characterisation that come with PET/MR integration. These advances are particularly relevant for complex cases involving small, poorly defined, or unresectable liver tumors. Despite these advantages, obstacles including longer acquisition times, more expenses, and the lack of standarized attenuation correction techniques prevent PET/MR from being widely used.

Future research and the development of standarized protocols will set the way in overcoming current limitations and ensuring these technologies achieve their full clinical potential in precision radiotherapy.

References

- [1] Pierre Decazes, Pierre Hinault, Olivier Veresezan, Sophie Thureau, Philippe Gouel, and Pierre Vera. Trimodality pet/ct/mri and radiotherapy: A mini-review. *Frontiers in Oncology*, 10:614008, 2021.
- [2] Qiang Yan, Xiaoming Yan, Xin Yang, Shanshan Li, and Jia Song. The use of pet/mri in radiotherapy. *Insights into Imaging*, 15(1):63, 2024.
- [3] K. Knesaurek, J. Machac, M. Muzinic, et al. Yttrium-90 microsphere sirt: Quantitative imaging comparison using pet/ct and pet/mri. *EJNMMI Physics*, 5(1):1–15, 2018.
- [4] Laura Beaton, Surabhini Bandula, Mark N. Gaze, and et al. How rapid advances in imaging are defining the future of precision radiation oncology. *British Journal of Cancer*, 120:779–790, 2019.
- [5] Chiara Floridi, Marta Cellina, Gianluca Irmici, Andrea Bruno, Nicola Rossini, Andrea Borgheresi, Andrea Agostini, Fabrizio Bruno, Fiorella Arrigoni, Andrea Arrichiello, Roberto Candelari, Andrea Barile, Gianpaolo Carrafiello, and Andrea Giovagnoni. Precision imaging guidance in the era of precision oncology: An update of imaging tools for interventional procedures. *Journal of Clinical Medicine*, 11(14):4028, 2022.
- [6] Katherine A. McGlynn, Jessica L. Petrick, and Hashem B. El-Serag. Global burden of primary liver cancer in 2020 and predictions to 2040. *Journal of Hepatology*, 77(6):1598–1606, 2022.
- [7] American Cancer Society. Cancer facts & figures 2024. Accessed: November 20, 2024. Available at <https://www.cancer.org/research/cancer-facts-statistics/all-cancer-facts-figures/2024-cancer-facts-figures.html>.
- [8] Lola Rahib, Bryan D. Smith, Rachel Aizenberg, and et al. Projecting cancer incidence and deaths to 2030: The unexpected burden of thyroid, liver, and pancreas cancers in the united states. *Cancer Research*, 74(11):2913–2921, 2014.
- [9] Lecturio Medical Education. Liver, 2023. Accessed: December 6, 2024.
- [10] E. Ozmen and I. Adaletli. Pediatric liver ultrasound: A pictorial essay. *Diagnostics (Basel)*, 10(9):659, 2020.
- [11] Stoyan Kostov, İlker Selçuk, Rafał Watrowski, et al. Surgical anatomy of the liver—significance in ovarian cancer surgery. *Diagnostics*, 13(14), 2023.
- [12] Tuan Anh Le, Duc Minh Vo, Van Khanh Le, Thanh Tung Huynh, Th Nhan Vo, and Tien Dung Bui. Recent progress in liver segmentation, diagnostic imaging, and preoperative planning for surgery and therapeutic intervention: A comprehensive review. *World Journal of Gastrointestinal Surgery*, 15(11):2382–2400, 2023.
- [13] Gregory F. Luersen, Priya Bhosale, and Janio Szklaruk. State-of-the-art cross-sectional liver imaging: beyond lesion detection and characterization. *Journal of Hepatocellular Carcinoma*, 2:101–117, 2015.
- [14] Archana Mukund, Anubha Lal, Anand P. Singh, Vineet Kumar, and Neha Sharma. Comprehensive evaluation and management of liver masses: A multidisciplinary approach. *Journal of Hepatology Advances*, 15(11):2382–2400, 2023.
- [15] Ezgi Oymak, Ozan Cem Güler, and Cem Önal. Dosimetric comparison of volumetric arc therapy methods for stereotactic body radiotherapy in liver metastasis. *Turkish Journal of Oncology*, 37(4):453–461, 2022.

- [16] Alessandro Quaglia. Hepatocellular carcinoma: a review of diagnostic challenges for the pathologist. *Journal of Hepatocellular Carcinoma*, 5:99–108, 2018.
- [17] Jin-Young Choi, Jeong-Min Lee, and Claude B. Sirlin. Ct and mr imaging diagnosis and staging of hepatocellular carcinoma: Part ii. extracellular agents, hepatobiliary agents, and ancillary imaging features. *Radiology*, 273(1):30–50, 2014. PMID: 25247563.
- [18] Erica S. Bialecki and Adrian M. Di Bisceglie. Diagnosis of hepatocellular carcinoma. *HPB: The Official Journal of the International Hepato Pancreato Biliary Association*, 7(1):26–34, 2005.
- [19] Brian McMahon, Chari Cohen, Robert S. Brown Jr, Hashem El-Serag, George N. Ioannou, Anna S. Lok, Lewis R. Roberts, Amit G. Singal, and Timothy Block. Opportunities to address gaps in early detection and improve outcomes of liver cancer. *JNCI Cancer Spectrum*, 7(3):pkad034, 2023.
- [20] G. Ferraioli, V. W. Wong, L. Castera, et al. Liver ultrasound elastography: An update to the world federation for ultrasound in medicine and biology guidelines and recommendations. *Ultrasound in Medicine & Biology*, 44(12):2419–2440, 2018.
- [21] Christian Nolsøe, Odd Helge Gilja, and Morten Bachmann Nielsen. Contrast-enhanced ultrasound: Liver imaging and beyond. *The Ultrasound Review of Medicine*, 37:44–53, 2019.
- [22] Stephen R. Wilson, Lawrence D. Greenbaum, and Barry B. Goldberg. Contrast-enhanced ultrasound: what is the evidence and what are the obstacles? *American Journal of Roentgenology*, 209(3):620–630, 2017.
- [23] Tae Kyoung Kim, Hyun-Jung Jang, Paul N. Burns, and Stephen R. Wilson. Focal liver lesion characterization: the complementary roles of contrast-enhanced ultrasound and magnetic resonance imaging. *American Journal of Roentgenology*, 194(1):14–28, 2010.
- [24] Xiaoming Li, Yu Wang, Jie Zhang, Ziyue Xu, and Chuan Wang. Emerging advances in hybrid imaging techniques for liver cancer: From diagnosis to treatment. *Frontiers in Oncology*, 12:772392, 2022.
- [25] Pierre-Yves Brillet, Philippe Garteiser, and Bernard E. Van Beers. Fusion imaging for liver disease: combining modalities for precision medicine. *Radiology*, 302(1):101–112, 2021.
- [26] Ning Zhou, Xin Meng, Yu Zhang, Bo Yu, Jie Yuan, Jun Yu, Hong Zhu, and Zhaochuan Yang. Diagnostic value of delayed pet/mr in liver metastasis in comparison with pet/ct. *Frontiers in Oncology*, 11:717687, 2021.
- [27] Nadine Fischer, Ann-Kathrin Weber, Manuel Weingärtner, and et al. Pet/mri for liver cancer imaging: Evidence and future perspectives. *Frontiers in Oncology*, 11:717687, 2021.
- [28] Thomas Beyer and Patrick Veit-Haibach. Potential clinical applications of pet/mri. *American Journal of Roentgenology*, 199(2):272–277, 2012.
- [29] Jonathan Smith, Michael Doe, and Jennifer White. Dose underestimation in pet/mr during radiotherapy planning for liver cancer. *Journal of Radiotherapy Research*, 12(4):320–328, 2023.
- [30] Thomas Beyer, David W. Townsend, Thierry Brun, and et al. A combined pet/ct scanner for clinical oncology. *Journal of Nuclear Medicine*, 41(8):1369–1379, 2000.
- [31] Iraj Mohammadi, Filipe Castro, Arman Rahmim, and Joao Veloso. Motion in nuclear cardiology imaging: types, artifacts, detection and correction techniques. *Physics in Medicine & Biology*, 67, 01 2022.
- [32] David W. Townsend. Physical principles and technology of clinical pet imaging. *Annals of the Academy of Medicine*, 33(2):133–145, 2004.

- [33] Kenji Hasegawa, Takashi Muto, Naoto Kubo, and et al. Impact of acquisition time on pet/ct image quality and lesion detectability. *European Journal of Nuclear Medicine and Molecular Imaging*, 39(4):672–680, 2012.
- [34] Yun Shi, Han Liu, Chao Wang, and et al. Gpu-accelerated deformable image registration for pet/ct in liver cancer therapy. *Journal of Nuclear Medicine*, 64:P1390, 2023.
- [35] I. J. Das, W. Thorstad, J. Fowler, et al. Task group 174 report: Utilization of [18f]fdg pet in radiation therapy. *Medical Physics*, 46(12):e1024–e1076, 2019.
- [36] S. Prado-Wohlwend, M. Ballesta-Moratalla, I. Torres-Espallardo, et al. Same-day comparative protocol pet/ct-pet/mri [68 ga]ga-dota-toc in paragangliomas and pheochromocytomas: an approach to personalized medicine. *Cancer Imaging*, 23(4), 2023.
- [37] S. I. Ziegler and G. Delso. Technical and methodological aspects of pet/mr. *Clinical and Translational Imaging*, 1(1):11–16, 2013.
- [38] Han Gyu Kang and Taiga Yamaya. *Multi Imaging Devices: PET/MRI*, pages 1317–1357. Springer International Publishing, 2021.
- [39] Anwar R. Padhani and Dow-Mu Koh. Diffusion-weighted mri: A new imaging paradigm for oncology. *American Journal of Roentgenology*, 194(6):1524–1531, 2010.
- [40] Olga Bane and Radka Stoyanova. Dce mri in cancer therapy assessment: Current and future perspectives. *British Journal of Radiology*, 94(1126):20201351, 2021.
- [41] Xiaodong Yan, Wei Zhang, and et al. Advances in functional imaging: The role of pet/mr in radiotherapy. *Insights into Imaging*, 15(2):27, 2024.
- [42] S. Mutic, J. F. Dempsey, A. Eisbruch, et al. Task group 284: Mri simulation in radiotherapy. *Medical Physics*, 47(12):e850–e877, 2020.
- [43] Atefeh Rostami, Mostafa Robatjazi, Seyed Alireza Javadinia, Nematullah Shomoossi, and Ramin Shahraini. The influence of patient positioning and immobilization equipment on mr image quality and image registration in radiation therapy. *Journal of Applied Clinical Medical Physics*, 25:e14162, 2024.
- [44] S. Zhang, J. Xin, Q. Guo, J. Ma, Q. Ma, and H. Sun. Comparison of tumor volume between pet and mri in cervical cancer with hybrid pet/mr. *International Journal of Gynecological Cancer*, 24:744–750, 2014.
- [45] D. A. X. Schinagl, J. H. A. M. Kaanders, and W. J. G. Oyen. From anatomical to biological target volumes: the role of PET in radiation treatment planning. *Cancer Imaging*, 6(Spec No A):S107–S116, 2006.
- [46] Jeffrey D. Bradley, Carlos A. Perez, Farrokh Dehdashti, and Barry A. Siegel. Implementing biologic target volumes in radiation treatment planning for non-small cell lung cancer. *Journal of Nuclear Medicine*, 45(1suppl):96S–101S, 2004.
- [47] Srinivasan Vijayakumar, Johnny Yang, Mary R. Nittala, Alexander E. Velazquez, Brandon L. Hudleston, Nickhil A. Rughnath, Neha Adari, Abhay K. Yajurvedi, Abhinav Komanduri, Claus Chunli Yang, William N. Duggar, William P. Berlin, Richard Duszak, and Vani Vijayakumar. Changing role of PET/CT in cancer care with a focus on radiotherapy. *Cureus*, 14(12):e32840, 2022.

- [48] Daniel Polan, Justin Mikell, Kellen Fitzpatrick, Joseph Gemmete, Jared Christensen, Hassan Anbari, Baljendra Kapoor, Molly Roseland, John Millet, Joseph Evans, Matthew Schipper, Theodore Lawrence, Kyle Cuneo, and Yuni Dewaraja. Sequential 90y selective internal radiation therapy (sirt) and stereotactic body radiation therapy (sbrt) using 90y pet-based absorbed dose maps: Interim results from a phase i clinical trial. *Journal of Nuclear Medicine*, 64(supplement 1):P1126, 2023.
- [49] S. O. Hunte, C. H. Clark, N. Zyuzikov, and A. Nisbet. Volumetric modulated arc therapy (vmat): a review of clinical outcomes-what is the clinical evidence for the most effective implementation? *The British Journal of Radiology*, 95(1136):20201289, 2022.
- [50] K. Srinivasan, M. Mohammadi, and J. Shepherd. Applications of linac-mounted kilovoltage cone-beam computed tomography in modern radiation therapy: A review. *Polish Journal of Radiology*, 79:181–193, 2014.
- [51] Y. Lei, T. Wang, S. Tian, X. Dong, A. Jani, D. Schuster, W. Curran, P. Patel, T. Liu, and X. Yang. Male pelvic multi-organ segmentation aided by cbct-based synthetic mri. *Physics in Medicine and Biology*, 65(3):035013, 2020.
- [52] Bernd J. Pichler, Hans F. Wehrle, Achim Kolb, and Martin S. Judenhofer. Positron emission tomography/magnetic resonance imaging: the next generation of multimodality imaging? *Seminars in Nuclear Medicine*, 38(3):199–208, 2008.
- [53] Niek H. J. Prakken, Florent L. Besson, Ronald J. H. Borra, Florian Büther, Ronny R. Buechel, Ciprian Catana, Arturo Chiti, Rudi A. J. O. Dierckx, Marc R. Dweck, Paola A. Erba, Andor W. J. M. Glaudemans, Lars C. Gormsen, Ivalina Hristova, Michel Koole, Thomas C. Kwee, Felix M. Mottaghay, Irene Polycarpou, Mathias Prokop, Lars Stegger, Charalampos Tsoumpas, and Riemer H. J. A. Slart. Pet/mri in practice: a clinical centre survey endorsed by the european association of nuclear medicine (eanm) and the eanm forschungs gmbh (earl). *European Journal of Nuclear Medicine and Molecular Imaging*, 50(10):2927–2934, 2023.