



| AA | No. | Old value | New value | Area |
|----|-----|------------|--------------|------|
| | 1.1 | 22,5-0,1/0 | 22,0 +0,15/0 | 3B |
| | 1.2 | 2x45° | 30° | 3B |
| | 1.3 | 2x45° | 3 | 3B |

| V. | Change | Date | Name |
|----|------------------------|------------|----------|
| 01 | Ø+Phase chuck mounting | 29.09.2021 | G. Gaus |
| 02 | Transfer to Inventor | 30.01.2025 | Z.Gabric |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Unless otherwise specified:
Dimensions are in millimeters.
General tolerances:
- Plastic moulded parts: DIN ISO 20457:2018
- TG7
- Metal and other plastic parts: DIN ISO 2769-fH
Material to comply with
RoHS Directive 2011/65/EU incl. 2015/863/EU

ERGOSANA
SCHILLER GROUP
Ganshorn Medizin Electronic GmbH
Industriestraße 6-8, D-97618 Niederlauer, Germany

Proprietary and confidential.
The information contained in this drawing is the sole property of Ganshorn Medizin Electronic GmbH. Any reproduction in part or as a whole without the written permission of Ganshorn Medizin Electronic GmbH is prohibited.

| | | |
|------------------------------|----------------|---------------------------------|
| Material 1.0711 (9s20k) | Scale 2 : 1 | |
| Title Saddle mounting pin | | Approved by |
| | | ERG-Article-Nr. 21-10-803-A2 |
| | | Dept.R&D Abteilung |
| | | Sheet 1 |