APPENDIX FULL PAIRWISE ANALYSIS TABLES

TABLE I
PAIRWISE DIFFERENCES OF SELECTED SYSTEMS. CREDIBILITY (CREDIB.) IS PROBABILITY OF IMPROVEMENT (OF THE BETTER SYSTEM) MEASURED BY BOOTSTRAP ESTIMATE [1]. SENTENCE DIFFERENCE (S-DIFF.) IS THE PERCENTAGE OF TRANSCRIPTIONS THAT WERE DIFFERENT BETWEEN A AND B.

	A			В			WEF	R Difference	(B-A)	S-Diff. WER Rank		Rank $ au$
	Family	Encoder	LM Data	Family	Encoder	LM Data	[pp.]	Relative	Credib.	[%]	Utt.	Spk.
	HMM	CRDNN	Transcript	AED	CRDNN	-	+2.18	+21%	100%	63.85	0.54	0.87
	HMM	CRDNN	Parl30M	AED	CRDNN	Parl30M	+0.72	+11%	100%	48.18	0.59	0.73
	FS HMM	CRDNN	Transcript	AED	CRDNN	_	-0.16	-1.6%	79%	72.67	0.47	0.80
	FS HMM	CRDNN	Parl30M	AED	CRDNN	Parl30M	+0.13	+2.0%	79%	53.48	0.54	0.83
FP Test16	HMM	w2v2.0	Transcript	AED/CTC	w2v2.0	-	+0.32	+4.8%	99%	51.09	0.59	0.84
	HMM	w2v2.0	Parl30M	AED/CTC	w2v2.0	Parl30M	+0.09	+1.7%	79%	42.84	0.65	0.87
F ₂	AED	CRDNN	Parl30M	AED	CRDNN	-	+3.72	+36%	100%	53.32	0.62	0.70
	AED/CTC	w2v2.0	_	AED	CRDNN	_	+3.60	+35%	100%	64.83	0.51	0.73
\uparrow	AED/CTC	w2v2.0	Parl30M	AED	CRDNN	Parl30M	+0.63	+9.6%	100%	45.49	0.60	0.78
Train20	AED/CTC	w2v2.0	Parl30M	AED/CTC	w2v2.0	_	+0.76	+11%	100%	37.81	0.71	0.82
rair	HMM	CRDNN	Parl30M	HMM	CRDNN	Transcript	+2.27	+28%	100%	50.67	0.63	0.68
FP T	HMM	w2v2.0	Transcript	HMM	CRDNN	Transcript	+1.75	+21%	100%	52.96	0.59	0.78
	HMM	w2v2.0	Parl30M	HMM	CRDNN	Parl30M	+0.01	+0.2%	54%	40.04	0.66	0.88
	HMM	w2v2.0	Parl30M	HMM	w2v2.0	Transcript	+0.53	+8.3%	100%	40.04	0.71	0.85
	HMM	CRDNN	Transcript	FS HMM	CRDNN	Transcript	+2.34	+22%	100%	60.74	0.57	0.86
	HMM	CRDNN	Parl30M	FS HMM	CRDNN	Parl30M	+0.59	+9.0%	100%	42.32	0.65	0.83
	FS HMM	CRDNN	Parl30M	FS HMM	CRDNN	Transcript	+4.02	+38%	100%	63.07	0.51	0.78
	HMM	CRDNN	Transcript	AED	CRDNN	-	+0.98	+11%	100%	38.11	0.60	0.84
	HMM	CRDNN	Parl30M	AED	CRDNN	Parl30M	-0.83	-11%	100%	36.52	0.63	0.79
	FS HMM	CRDNN	Transcript	AED	CRDNN	-	-0.82	-9.5%	100%	46.28	0.53	0.81
	FS HMM	CRDNN	Parl30M	AED	CRDNN	Parl30M	-2.23	-30%	100%	42.37	0.58	0.77
0	HMM	w2v2.0	Transcript	AED/CTC	w2v2.0	-	+0.12	+1.8%	82%	28.31	0.68	0.85
FP Test20	HMM	w2v2.0	Parl30M	AED/CTC	w2v2.0	Parl30M	-0.93	-13%	100%	34.28	0.66	0.80
Te	AED	CRDNN	Parl30M	AED	CRDNN	-	+1.17	+13%	100%	30.95	0.70	0.78
	AED/CTC	w2v2.0	-	AED	CRDNN	-	+1.97	+23%	100%	38.69	0.57	0.77
\uparrow	AED/CTC	w2v2.0	Parl30M	AED	CRDNN	Parl30M	+0.30	+4.1%	96%	32.22	0.66	0.84
120	AED/CTC	w2v2.0	Parl30M	AED/CTC	w2v2.0	-	-0.49	-7.4%	100%	22.85	0.77	0.87
rair	HMM	CRDNN	Parl30M	HMM	CRDNN	Transcript	-0.64	-8.5%	100%	36.91	0.64	0.86
FP Train20	HMM	w2v2.0	Transcript	HMM	CRDNN	Transcript	+1.11	+15%	100%	30.60	0.66	0.81
	HMM	w2v2.0	Parl30M	HMM	CRDNN	Parl30M	+0.21	+2.5%	92%	32.15	0.70	0.79
	HMM	w2v2.0	Parl30M	HMM	w2v2.0	Transcript	-1.54	-24%	100%	31.49	0.68	0.83
	HMM	CRDNN	Transcript	FS HMM	CRDNN	Transcript	+1.80	+19%	100%	36.83	0.64	0.80
	HMM	CRDNN	Parl30M	FS HMM	CRDNN	Parl30M	+1.40	+15%	100%	39.00	0.63	0.82
	FS HMM	CRDNN	Parl30M	FS HMM	CRDNN	Transcript	-0.24	-2.6%	91%	45.47	0.56	0.72

REFERENCES

^[1] M. Bisani and H. Ney, "Bootstrap estimates for confidence intervals in asr performance evaluation," in 2004 IEEE International Conference on Acoustics, Speech, and Signal Processing, vol. 1, 2004, pp. I–409.

TABLE II

PAIRWISE DIFFERENCES OF SELECTED SYSTEMS. CREDIBILITY (CREDIB.) IS PROBABILITY OF IMPROVEMENT (OF THE BETTER SYSTEM) MEASURED BY BOOTSTRAP ESTIMATE [1]. SENTENCE DIFFERENCE (S-DIFF.) IS THE PERCENTAGE OF TRANSCRIPTIONS THAT WERE DIFFERENT BETWEEN A AND B.

THE SUPERSCRIPTS MW AND FT MEAN THE POST-10-MODEL-AVERAGING MINIMUM WORD ERROR RATE FINETUNING AND HMM CONTINUED TRAINING, RESPECTIVELY.

	A			В			WER Difference (B-A)			S-Diff. WER Rank $ au$		Rank $ au$
	Family	Encoder	LM Data	Family	Encoder	LM Data	[pp.]	Relative	Credib.	[%]	Utt.	Spk.
	HMM	CRDNN	Transcript	AED	CRDNN	-	+0.55	+11%	100%	47.60	0.51	0.69
u	HMM	CRDNN	Orig. 800M	AED	CRDNN	-	+1.22	+24%	100%	47.71	0.45	0.53
	FS HMM	CRDNN	Transcript	AED	CRDNN	-	-0.24	-4.7%	97%	51.15	0.49	0.65
	FS HMM	CRDNN	Orig. 800M	AED	CRDNN	-	+0.70	+14%	100%	51.18	0.42	0.47
	HMM	w2v2.0	Transcript	AED/CTC	w2v2.0	-	+0.17	+7.0%	100%	26.95	0.60	0.60
	HMM	Conformer _S	Transcript	AED/CTC	Conformer _S	-	+0.05	+1.4%	72%	35.95	0.58	0.64
	HMM	Conformer _S	Orig. 800M	AED/CTC	Conformer _S	-	+0.28	+7.7%	98%	38.85	0.50	0.52
	HMM	Conformer FT	Transcript	AED/CTC	Conformer $_{S}^{MW}$	-	+0.04	+1.2%	69%	36.45	0.58	0.67
	HMM	$\begin{array}{c} Conformer_S^{FT} \\ Conformer_S^{FT} \end{array}$	Orig. 800M	AED/CTC	$\begin{array}{c} \text{Conformer}_S^{MW} \\ \text{Conformer}_S^{MW} \end{array}$	-	+0.41	+11%	100%	39.24	0.48	0.52
	HMM	$Conformer_L$	Transcript	AED/CTC	$Conformer_L$	-	+0.18	+6.3%	99%	29.31	0.60	0.73
lea	HMM	$Conformer_L$	Orig. 800M	AED/CTC	$Conformer_L$	-	+0.21	+7.2%	99%	32.25	0.53	0.62
st C	HMM	$Conformer_L$	Transcript	AED/CTC	Conformer $_{L}^{MW}$	-	-0.03	-1.2%	67%	27.44	0.61	0.71
Tes	HMM	$Conformer_L$	Orig. 800M	AED/CTC	Conformer $_{L}^{\widetilde{M}W}$	-	-0.01	-0.2%	52%	30.95	0.53	0.61
sch	FS HMM	Conformer $_{L}^{FT}$	Transcript	AED/CTC	Conformer $_{L}^{\widetilde{M}W}$	-	-0.36	-14%	100%	29.20	0.62	0.70
bec	FS HMM	Conformer $_{L}^{FT}$ Conformer $_{L}^{FT}$	Orig. 800M	AED/CTC	$\begin{array}{c} \operatorname{Conformer}_L^{MW} \\ \operatorname{Conformer}_L^{MW} \\ \operatorname{Conformer}_L^{MW} \\ \operatorname{Conformer}_L^{MW} \end{array}$	-	-0.27	-10%	100%	32.33	0.52	0.54
Librispeech Test Clean	AED/CTC	$Conformer_L$	-	AED/CTC	Conformer _S	-	+0.78	+21%	100%	31.76	0.64	0.70
	AED/CTC	w2v2.0	-	AED/CTC	$Conformer_L$	-	+0.44	+16%	100%	30.92	0.59	0.71
	AED/CTC	w2v2.0	-	AED/CTC	Conformer $_L^{MW}$	-	+0.23	+8.8%	100%	29.85	0.60	0.69
	AED/CTC	w2v2.0	-	AED/CTC	Conformer _S	- - -	+1.22	+34%	100%	35.65	0.57	0.67
	AED/CTC	w2v2.0	-	AED/CTC	Conformer $_{S}^{MW}$	-	+1.20	+33%	100%	35.73	0.57	0.66
	AED/CTC	w2v2.0	-	AED	CRDNN	-	+2.70	+53%	100%	45.92	0.47	0.58
	AED/CTC	Conformer $_L$	-	AED	CRDNN	-	+2.26	+44%	100%	43.21	0.51	0.58
	AED/CTC	Conformer _S	-	AED	CRDNN	-	+1.48	+29%	100%	43.05	0.54	0.65
	HMM	$Conformer_L$	Transcript	HMM	$Conformer_S$	Transcript	+0.91	+25%	100%	32.06	0.60	0.70
	HMM	$Conformer_L$	Orig. 800M	HMM	$Conformer_S$	Orig. 800M	+0.70	+21%	100%	26.26	0.67	0.72
	HMM	w2v2.0	Transcript	HMM	Conformer $_L$	Transcript	+0.43	+16%	100%	28.02	0.58	0.67
	HMM	w2v2.0	Transcript	HMM	$Conformer_S$	Transcript	+1.34	+37%	100%	35.42	0.51	0.58
	HMM	w2v2.0	Transcript	HMM	CRDNN	Transcript	+2.32	+51%	100%	42.98	0.45	0.46
	HMM	Conformer $_L^{FT}$	Transcript	HMM	CRDNN	Transcript	+1.87	+41%	100%	40.92	0.51	0.62
	HMM	$Conformer_S$	Transcript	HMM	CRDNN	Transcript	+0.98	+22%	100%	38.66	0.59	0.72
	HMM	CRDNN	Orig. 800M	HMM	CRDNN	Transcript	+0.67	+15%	100%	35.23	0.62	0.67
	HMM	CRDNN	Transcript	FS HMM	CRDNN	Transcript	+0.79	+15%	100%	43.44	0.58	0.75
	HMM	CRDNN	Orig. 800M	FS HMM	CRDNN	Orig. 800M	+0.52	+12%	100%	33.85	0.65	0.70
	HMM	$Conformer_L$	Transcript	FS HMM	$\begin{array}{c} \text{Conformer}_L^{FT} \\ \text{Conformer}_L^{FT} \end{array}$	Transcript	+0.33	+11%	100%	27.82	0.63	0.71
	HMM	$Conformer_L$	Orig. 800M	FS HMM	Conformer $_{L}^{\overline{F}T}$	Orig. 800M	+0.27	+9.2%	100%	26.30	0.66	0.73
	FS HMM	CRDNN	Orig. 800M	FS HMM	CRDNN	Transcript	+0.94	+18%	100%	42.14	0.57	0.55
	FS HMM	Conformer $_L^{FT}$	Orig. 800M	FS HMM	$Conformer_L^{FT}$	Transcript	+0.09	+3.0%	85%	25.34	0.63	0.58
	HMM	Conformer $_{S}^{FT}$	Transcript	HMM	$Conformer_S$	Transcript	+0.02	+0.4%	69%	6.87	0.94	0.92
	FS HMM	Conformer $_{L}^{\widetilde{F}T}$	Transcript	FS HMM	$Conformer_L$	Transcript	+0.08	+2.7%	95%	17.44	0.80	0.86
	AED/CTC	Conformer $_{S}^{\overline{M}W}$	-	AED/CTC	$Conformer_S$	-	+0.02	+0.6%	75%	9.16	0.92	0.90
	AED/CTC	$\operatorname{Conformer}_L^{\widetilde{M}W}$	-	AED/CTC	$Conformer_L$	-	+0.21	+7.4%	100%	13.47	0.84	0.81

TABLE III

PAIRWISE DIFFERENCES OF SELECTED SYSTEMS. CREDIBILITY (CREDIB.) IS PROBABILITY OF IMPROVEMENT (OF THE BETTER SYSTEM) MEASURED BY BOOTSTRAP ESTIMATE [1]. SENTENCE DIFFERENCE (S-DIFF.) IS THE PERCENTAGE OF TRANSCRIPTIONS THAT WERE DIFFERENT BETWEEN A and B. The superscripts MW and FT mean the post-10-model-averaging Minimum Word Error Rate finetuning and HMM finetuning, respectively.

	A			В			WER Difference (B-A)			S-Diff.	WER	Rank $ au$
	Family	Encoder	LM Data	Family	Encoder	LM Data	[pp.]	Relative	Credib.	[%]	Utt.	Spk.
	HMM	CRDNN	Transcript	AED	CRDNN	-	+2.46	+16%	100%	75.88	0.53	0.80
	HMM	CRDNN	Orig. 800M	AED	CRDNN	-	+4.58	+30%	100%	76.62	0.48	0.74
	FS HMM	CRDNN	Transcript	AED	CRDNN	-	+1.35	+9.0%	100%	77.99	0.53	0.81
	FS HMM	CRDNN	Orig. 800M	AED	CRDNN	-	+3.93	+26%	100%	79.55	0.43	0.70
	HMM	w2v2.0	Transcript	AED/CTC	w2v2.0	-	+1.02	+17&	100%	47.29	0.60	0.78
	HMM	Conformer _S	Transcript	AED/CTC	Conformer _S	-	+0.61	+6.9%	100%	59.65	0.57	0.75
	HMM	$Conformer_S$	Orig. 800M	AED/CTC	$Conformer_S$	-	+2.08	+23%	100%	61.01	0.52	0.71
	HMM	Conformer $_{S}^{FT}$	Transcript	AED/CTC	Conformer $_{S}^{MW}$	-	+0.66	+7.4%	100%	59.24	0.58	0.78
Librispeech Test Other	HMM	Conformer $_{S}^{FT}$	Orig. 800M	AED/CTC	Conformer $_S^{MW}$ Conformer $_S^{MW}$	-	+2.04	+23%	100%	61.18	0.52	0.71
	HMM	Conformer _{L}	Transcript	AED/CTC	$Conformer_L$	-	-0.17	-2.8%	94%	45.93	0.64	0.79
	HMM	$Conformer_L$	Orig. 800M	AED/CTC	Conformer _L		+0.32	+5.5%	100%	47.91	0.58	0.71
	HMM	Conformer $_L$	Transcript	AED/CTC	$\begin{array}{c} \operatorname{Conformer}_{L}^{MW} \\ \operatorname{Conformer}_{L}^{MW} \end{array}$	-	-0.08	-1.3%	77%	46.55	0.63	0.78
Tes	HMM	Conformer $_L$	Orig. 800M	AED/CTC	Conformer MW	-	+0.41	+6.8%	100%	48.79	0.57	0.72
ç	FS HMM	Conformer $_{\tau}^{FT}$	Transcript	AED/CTC	Conformer _L	-	-0.69	-12%	100%	49.17	0.61	0.76
ebee.	FS HMM	Conformer $_{I}^{E}T$	Orig. 800M	AED/CTC	Conformer _I	-	-0.02	-0.3%	55%	51.75	0.52	0.67
bris	FS HMM	$\begin{array}{c} \text{Conformer}_{L}^{ET} \\ \text{Conformer}_{L}^{FT} \end{array}$	Transcript	AED/CTC	Conformer $_{I}^{MW}$	-	-0.60	-10%	100%	50.09	0.59	0.76
I	FS HMM	Conformer $_{L}^{FT}$	Orig. 800M	AED/CTC	$\begin{array}{c} \overset{L}{\text{Conformer}}^{MW}_L \\ \text{Conformer}^{MW}_L \end{array}$	-	+0.07	+1.2%	72%	52.67	0.51	0.67
	AED/CTC	$Conformer_L$	-	AED/CTC	Conformer _S	-	+2.95	+33%	100%	52.88	0.61	0.80
	AED/CTC	w2v2.0	-	AED/CTC	$Conformer_L$	-	-0.13	-2.2%	59%	51.72	0.57	0.76
	AED/CTC	w2v2.0	-	AED/CTC	$\begin{array}{c} Conformer_L \\ Conformer_L^{MW} \end{array}$	-	-0.04	-0.7%	50%	51.82	0.57	0.74
	AED/CTC	w2v2.0	-	AED/CTC	Conformer _S		+2.82	+32%	100%	60.67	0.54	0.74
	AED/CTC	w2v2.0	-	AED/CTC	Conformer $_{S}^{\stackrel{\circ}{M}W}$	-	+2.78	+31%	100%	60.97	0.53	0.75
	AED/CTC	w2v2.0	-	AED	CRDNN	-	+8.96	+60%	100%	74.48	0.43	0.65
	AED/CTC	$Conformer_L$	-	AED	CRDNN	-	+2.26	+44%	100%	43.21	0.51	0.58
	AED/CTC	Conformer _S	-	AED	CRDNN	-	+1.48	+29%	100%	43.05	0.54	0.65
	HMM	Conformer $_L$	Transcript	HMM	$Conformer_S$	Transcript	+2.17	+26%	100%	50.63	0.64	0.81
	HMM	$Conformer_L$	Orig. 800M	HMM	$Conformer_S$	Orig. 800M	+1.19	+18%	100%	44.27	0.66	0.81
	HMM	w2v2.0	Transcript	HMM	$Conformer_L$	Transcript	+1.06	+17%	100%	49.17	0.56	0.76
	HMM	w2v2.0	Transcript	HMM	Conformer _S	Transcript	+3.23	+39%	100%	57.23	0.53	0.73
	HMM	w2v2.0	Transcript	HMM	CRDNN	Transcript	+7.53	+60%	100%	69.45	0.43	0.66
	HMM	$Conformer_L$	Transcript	HMM	CRDNN	Transcript	+6.46	+51%	100%	67.57	0.50	0.70
	HMM	Conformer $_{S}^{FT}$	Transcript	HMM	CRDNN	Transcript	+4.37	+35%	100%	65.09	0.55	0.69
	HMM	CRDNN	Orig. 800M	HMM	CRDNN	Transcript	+2.12	+17%	100%	62.61	0.63	0.81
	HMM	CRDNN	Transcript	FS HMM	CRDNN	Transcript	+1.11	+8.1%	100%	69.48	0.61	0.81
	HMM	CRDNN	Orig. 800M	FS HMM	CRDNN	Orig. 800M	+0.65	+5.9%	100%	62.78	0.60	0.81
	HMM	Conformer $_L$	Transcript	FS HMM	$\begin{array}{c} \text{Conformer}_L^{FT} \\ \text{Conformer}_L^{FT} \end{array}$	Transcript	+0.52	+7.9%	100%	46.51	0.63	0.79
	HMM	$Conformer_L$	Orig. 800M	FS HMM	ConformerFT	Orig. 800M	+0.34	+5.7%	100%	43.11	0.62	0.81
	FS HMM	CRDNN	Orig. 800M	FS HMM	CRDNN	Transcript	+2.58	+19%	100%	68.49	0.58	0.74
	FS HMM	Conformer $_L^{FT}$	Orig. 800M	FS HMM	Conformer $_{L}^{FT}$	Transcript	+0.68	+10%	100%	40.63	0.66	0.73
	HMM	Conformer $_{S}^{\overline{F}T}$	Transcript	HMM	Conformer _S	Transcript	+0.08	+1.0%	96%	16.40	0.91	0.93
	FS HMM	Conformer $_{I}^{FT}$	Transcript	FS HMM	$Conformer_L$	Transcript	-0.00	-0.1%	52%	34.60	0.75	0.85
	AED/CTC	Conformer $_{S}^{\widetilde{M}W}$	-	AED/CTC	$Conformer_S$	-	+0.04	+0.5%	75%	21.61	0.89	0.93
	AED/CTC	$\begin{array}{c} \operatorname{Conformer}_{S}^{LW} \\ \operatorname{Conformer}_{L}^{MW} \end{array}$	-	AED/CTC	$Conformer_L$	-	-0.09	-1.5%	94%	19.60	0.87	0.91
	'	L	. '	1	•	1		'		•		1