

Analyte	N	Events	Per_{RI}				$NORMA_{RI}$			
			Precision	Sensitivity	Specificity	Accuracy	Precision	Sensitivity	Specificity	Accuracy
A1C	—	—	—	—	—	—	—	—	—	—
ALB	866	154	0.19	0.26	0.77	0.60	0.32	0.21	0.91	0.53
ALP	9,746	3,242	0.32	0.58	0.40	0.47	0.40	0.28	0.79	0.56
ALT	7,367	2,411	0.32	0.55	0.42	0.47	0.37	0.26	0.78	0.54
AST	8,005	2,553	0.30	0.48	0.46	0.47	0.36	0.23	0.81	0.54
BUN	19,861	3,354	0.15	0.41	0.53	0.45	0.23	0.25	0.83	0.58
CA	16,054	2,944	0.20	0.35	0.68	0.54	0.28	0.25	0.85	0.58
CL	26,551	5,172	0.15	0.29	0.60	0.45	0.29	0.22	0.87	0.59
CO2	29,320	5,515	0.12	0.21	0.66	0.42	0.27	0.21	0.87	0.60
CRE	27,793	5,022	0.19	0.27	0.75	0.52	0.24	0.22	0.85	0.59
DBIL	284	133	0.56	0.21	0.85	0.52	0.55	0.27	0.80	0.53
GLU	9,515	724	0.06	0.18	0.75	0.52	0.12	0.24	0.86	0.59
HCT	2,912	466	0.16	0.42	0.57	0.50	0.20	0.13	0.90	0.53
HDL	—	—	—	—	—	—	—	—	—	—
HGB	3,367	486	0.13	0.33	0.63	0.52	0.17	0.13	0.89	0.53
K	57,667	11,159	0.20	0.35	0.66	0.51	0.29	0.28	0.84	0.63
LDL	—	—	—	—	—	—	—	—	—	—
MCH	32,228	7,289	0.25	0.51	0.55	0.54	0.31	0.16	0.90	0.60
MCHC	22,827	4,454	0.18	0.35	0.60	0.48	0.21	0.11	0.90	0.56
MCV	43,766	9,763	0.22	0.57	0.43	0.48	0.32	0.21	0.87	0.60
MPV	34,458	7,956	0.24	0.55	0.48	0.52	0.25	0.66	0.41	0.56
NA	40,655	8,501	0.17	0.36	0.55	0.45	0.33	0.25	0.86	0.62
PLT	33,483	7,573	0.22	0.67	0.30	0.47	0.33	0.49	0.71	0.63
RBC	4,587	776	0.18	0.29	0.73	0.55	0.22	0.15	0.89	0.54
RDW	16,261	2,917	0.19	0.45	0.57	0.51	0.24	0.20	0.86	0.58
TBIL	8,962	2,978	0.35	0.32	0.71	0.52	0.40	0.19	0.86	0.55
TC	11	1	0.00	0.00	0.30	—	0.00	0.00	0.70	—
TGL	170	114	0.65	0.54	0.41	0.44	0.65	0.23	0.75	0.45
TP	4,842	1,496	0.33	0.36	0.67	0.55	0.40	0.20	0.86	0.56
WBC	25,815	5,037	0.17	0.38	0.57	0.46	0.25	0.20	0.85	0.59