

ROUTINE TEST RECORD FOR BELT SCALE

JOB No. : BS-CHECK-V10
CAPACITY : 1000

CUSTOMER NAME: ABEL STEEL CORP
CONVEYOR NO: CV-V10

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BS-1000 Serial No: BS-SN-V10 REMOTE: Model No: RM-1000 Serial No: RM-SN-V10	Model No: BR3200H0 Serial No: 1) S1 2) S2 3) S3 4) S4	Sensor: Model No: TS-100 Serial No: TS-SN-V10	Model No: JB-1 Serial No: JB-SN-V10 Model No: Serial No:

Check for short between Line & Neutral, Neutral & Earth, Line & Earth :

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Display and Keypad Functionality	OK	OK
4.	Belt Load as specified in G.A. or MPL	185.19	185.19 Kg/m
5.	Belt Speed as specified in G.A. or MPL	1.50	1.50 m/s
6.	Number of PF Contacts	4	4
7.	Communication Output (Specify)	RS-485	RS-485
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	OK	OK

Resolution L: 0.1 S: 0.01 r: 1 t: 0.1

Load on Load cell Kg / mV	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA AO1	O/p current mA AO2
3	0.0	1.50	0	0.0	4.00	4.00
3	37.0	1.50	200	20.0	7.20	7.20
4	74.1	1.50	400	40.0	10.40	10.40
4	111.1	1.50	600	60.0	13.60	13.60
5	148.2	1.50	800	80.0	16.80	16.80
5	185.2	1.50	1000	100.0	20.00	20.00

Instrument Used: Fluke Multimeter
Tested by: System Admin
Date: 14-Sep-2026

Approved by:
(HOD - PDN)