

ROUTINE TEST RECORD FOR BELT SCALEJOB No. : BS-10001
CAPACITY : 1000 TPHCUSTOMER NAME: ABEL STEEL CORP
CONVEYOR NO: BC-01

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BS-10000 Serial No: BS-SN-001 S/W Version: 3.1.3 REMOTE: Model No: RM-100 Serial No: RM-SN-001	Model No: LC-500 Serial No: 1) LC-01 2) LC-02 3) LC-03 4) LC-04	Model No: SS-100 Serial No: TS-SN-01	Model No. : JB-101 Serial No. : JB-SN-101
		ANGLE SENSOR	Model No. : JB-102 Serial No. : JB-SN-102
		Model No: AS-300 Serial No: AS-SN-101	Model No. : JB-103 Serial No. : JB-SN-103

Check for short between Line & Neutral, Neutral & Earth, Line & Earth : OK (100M Ohm)

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Display and Keypad Functionality	Display OK	Display OK
4.	Belt Load as specified in G.A. or MPL	27.78	27.78 Kg/m
5.	Belt Speed as specified in G.A. or MPL	2.50	2.50 m/s
6.	Number of PF Contacts	4	4
7.	Communication Output (Specify)	RS-485	RS-485
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	Wiring OK	Wiring OK

Resolution: L: 0.1 kg/m S: 0.01 m/s r: 1 tph t: 0.1 tonnes

Measurement to be done from 0 to 110% of the Belt Scale capacity.

Load on Load cell Kg/Mv	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA

Instrument Used: Fluke 87V Multimeter

Date: 14-Sep-2026

(HOD - PDN)

Date: 14-Sep-2026

(HOD - PDN)