

ROUTINE TEST RECORD FOR BELT SCALE

JOB No. : BS-FINAL-1PAGE-TEST
CORP
CAPACITY : 1500

CUSTOMER NAME: ABEL STEEL

CONVEYOR NO: CV-01

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BS-2000 Serial No: BS-SN-001 REMOTE: Model No: RM-2000 Serial No: RM-SN-001	Model No: BR3200H0 Serial No: 1) SN1 2) SN2 3) SN3 4) SN4	Sensor: Model No: TS-100 Serial No: TS-SN1	Model No: JB-1 Serial No: JB-SN1 Model No: Serial No:

Check for short between Line & Neutral, Neutral & Earth, Line & Earth :

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Display and Keypad Functionality	OK	OK
4.	Belt Load as specified in G.A. or MPL	208.33	208.33 Kg/m
5.	Belt Speed as specified in G.A. or MPL	2.00	2.00 m/s
6.	Number of PF Contacts	4	4
7.	Communication Output (Specify)	RS-485	RS-485
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	OK	OK

Resolution

L: 0.1

S: 0.01

r: 1

t: 0.1

Load on Load cell Kg / mV	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA AO1	O/p current mA AO2
3	0.0	2.00	0	0.0	4.00	4.00
3	41.7	2.00	300	30.0	7.20	7.20
4	83.3	2.00	600	60.0	10.40	10.40
4	125.0	2.00	900	90.0	13.60	13.60
5	166.7	2.00	1200	120.0	16.80	16.80
5	208.3	2.00	1500	150.0	20.00	20.00

Instrument Used: Fluke 87V Multimeter & 50kg Class F Weights

Tested by: System Admin

Date: 14-Sep-2026

Approved by:

(HOD - PDN)