

ROUTINE TEST RECORD FOR BELT SCALE

JOB No. : JOB-2026-812

CAPACITY : 1000

CUSTOMER NAME: Tata Steel Mining

CONVEYOR NO: CV-39B

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BS-MODEL-812	Model No: LC-250KG	Sensor: Proximity	Model No: JBOX-IP67 Serial No: JB-SN-912
Serial No: SN-BS-7812	Serial No: 1) LC1-002 2) LC2-002 3) LC3-002 4) LC4-002	Model No: TACHO-SENS-PRO Serial No: TS-8812	Model No: JBOX-IP67 Serial No: JB-SN-1012
REMOTE: Model No: RM-UNIT-01 Serial No: SN-RM-5812			

Check for short between Line & Neutral, Neutral & Earth, Line & Earth : OK (Checked Line/Neutral, Neutral/Earth, Line/Earth)

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Display and Keypad Functionality	OK	OK
4.	Belt Load as specified in G.A. or MPL	141.0	141.0 Kg/m
5.	Belt Speed as specified in G.A. or MPL	1.97	1.97 m/s
6.	Number of PF Contacts	4	4
7.	Communication Output (Specify)	RS-485	RS-485
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	OK	OK

Resolution

L: 0.1

S: 0.01

r: 1

t: 0.1

Load on Load cell Kg / mV	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA AO1	O/p current mA AO2
3	0.0	1.97	0	0.0	4.00	4.00
3	28.2	1.97	200	20.0	7.20	7.20
4	56.4	1.97	400	40.0	10.40	10.40
4	84.6	1.97	600	60.0	13.60	13.60
5	112.8	1.97	800	80.0	16.80	16.80
5	141.0	1.97	1000	100.0	20.00	20.00

Instrument Used: Fluke 87V Digital Multimeter & 50kg Calibrated Class F Weights

Tested by: A

Date: 14-Sep-2026

Approved by: HOD-PDN

(HOD - PDN)