

ROUTINE TEST RECORD FOR BELT SCALE

JOB No: JOB-2026-886

CUSTOMER NAME: Adani Power Thermal

CAPACITY : 1200

CONVEYOR NO : CV-30B

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BS-MODEL-886 Serial No: SN-BS-7886 S/W Version: 3.1.3 REMOTE: Model No: RM-UNIT-01 Serial No: SN-RM-5886	Model No: LC-300KG Serial No: 1) LC1-006 2) LC2-006 3) LC3-006 4) LC4-006	Model No: TACHO-SENS-PRO Serial No: TS-8886 ANGLE SENSOR Model No: AS-300-TILT Serial No: SN-AS-9886	Model No: JBOX-IP67 Serial No: JB-SN-986 Model No: JBOX-IP67 Serial No: JB-SN-1086 Model No: JBOX-IP67 Serial No: JB-SN-1186

Check for short between Line & Neutral,
Neutral & Earth, Line & Earth : OK (Checked Line/Neutral, Neutral/Earth, Line/Earth)

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Display and Keypad Functionality	OK	OK
4.	Belt Load as specified in G.A. or MPL	170.1 kg/m	170.1 Kg/m
5.	Belt Speed as specified in G.A. or MPL	1.96	1.96 m/s
6.	Number of PF Contacts	4	4
7.	Communication Output (Specify)	RS-485	RS-485
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	OK	OK

Resolution: L: 0.1 S: 0.01 r: 1 t: 0.1

Measurement to be done from 0 to 110% of the Belt Scale capacity.

Load on Load cell Kg/Mv	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA AO1 AO2	
3	0.0	1.96	0	0.0	4.00	4.00
3	34.0	1.96	240	24.0	7.20	7.20
4	68.0	1.96	480	48.0	10.40	10.40
4	102.0	1.96	720	72.0	13.60	13.60
5	136.1	1.96	960	96.0	16.80	16.80
5	170.1	1.96	1200	120.0	20.00	20.00

Instrument Used: Fluke 87V Digital Multimeter & 50kg Calibrated Class F Weights

Tested by: A

Approved by: HOD-PDN

Date: 15-Sep-2026

(HOD - PDN)