

**ROUTINE TEST RECORD FOR LOAD INDICATOR / DIGITAL INDICATOR / TANK /
HOPPER / SILO WEIGHING SYSTEM/CRANE WEIGHING SYSTEM**

SYSTEM : DIGITAL INDICATOR

JOB No. : JOB-2026-709

CAPACITY: 1500 Kg

CUSTOMER NAME: UltraTech Minerals

TAG. NO: CV-74B

Load/Digital Indicator.	SENSOR	REMOTE (IF ANY)	JUNCTION BOX
Model No: DI-3000 Serial No: SN-DI-8821 S/W Version: v3.2.0	Model No: LC-CZ50T Serial No: 1) LC1-009 5) LC-SN-005 2) LC2-009 6) LC-SN-006 3) LC3-009 7) LC-SN-007 4) LC4-009 8) LC-SN-008	Model No: N/A Serial No: N/A	Model No: JB-8P Serial No: JB-SN-441

Check for short between Line & Neutral , Neutral & Earth , Line & Earth : Checked & found OK

SL No.	Test Performed	Specified Values	Actual Values
1.	Input Supply Voltage (+/-10%)	230V	230V
2.	Derived DC Voltages (+/-5%)	24V	24V
3.	Check for the power on display and the Keyboard function.	OK	OK
4.	Load on the Load cell as specified in G.A. or MPL	Full Capacity Span	Passed
5.	Tare weight as specified in G.A. or MPL	Zero Balance OK	Passed

Resolution: L: kg/T

Measurement to be done from 0 to 110% of the System capacity.

Load on Load cell Kg/T or mV	Measured Load kg/T (1 to 4 Load cell system)				Current output mA	Measured Load kg/T (5 to 8 Load cell system)			
	1	2	3	4		5	6	7	8
3.00	0.0	1.66	0	0.0		0	0	0	0
3.40	50.2	1.66	300	30.0		100	100	100	100
3.80	100.4	1.66	600	60.0		200	200	200	200
4.20	150.6	1.66	900	90.0		300	300	300	300
4.60	200.8	1.66	1200	120.0		400	400	400	400
5.00	251.0	1.66	1500	150.0		500	500	500	500
60%	600	600	600	600		600	600	600	600
70%	700	700	700	700		700	700	700	700
80%	800	800	800	800		800	800	800	800
90%	900	900	900	900		900	900	900	900
100%	1000	1000	1000	1000		1000	1000	1000	1000
110%	1100	1100	1100	1100		1100	1100	1100	1100

**ROUTINE TEST RECORD FOR LOAD / DIGITAL INDICATOR / TANK / HOPPER / SILO
WEIGHING SYSTEM**

SL No	Test Performed	Checked: YES/NO
6.	PF/ Relay Contacts: Check for the relay change over for the set points.	Checked: YES
7.	Display Annunciation: (As per the software Details)	Checked: YES
8.	Communication: RS 232 / RS 485 / BCD / Other Communication Protocols (If any specify) The RS232 / RS485 to be checked with interfacing to PC	Checked: YES
9.	Any other parameters (If Any)	N/A
10.	Rating of the fuse used in the system:	2A Fast Blow

REMARKS (If ANY): - Calibration completed within acceptable industrial limits.

Instrument Used: Fluke 87V Digital Multimeter & 50kg Calibrated Class F Weights

Tested by: A

Approved by: HOD-PDN

Date: 16-Sep-2026

(HOD - PDN)