

**ROUTINE TEST RECORD FOR LOAD INDICATOR / DIGITAL INDICATOR / TANK /
HOPPER / SILO WEIGHING SYSTEM/CRANE WEIGHING SYSTEM**

JOB No. :

CAPACITY: Kg/Tonnes

CUSTOMER NAME:**TAG. NO:**

Load/Digital Indicator.	SENSOR	REMOTE (IF ANY)	JUNCTION BOX
Model No:	Model No:	Model No:	Model No:
Serial No:	Serial No:	Serial No:	Serial No:
S/W Version:	1) 5) 2) 6) 3) 7) 4) 8)		Model No: Serial No:

Check for short between Line & Neutral , Neutral & Earth , Line & Earth :
--

SL · No ·	Test Performed	Specified Values	Actual Values
1.	Input Supply Voltage (+/-10%)		
2.	Derived DC Voltages (+/-5%)		
3.	Check for the power on display and the Keyboard function.		
4.	Load on the Load cell as specified in G.A. or MPL		Kg/T
5.	Tare weight as specified in G.A. or MPL		Kg/T

Resolution: L: kg/T

Measurement to be done from 0 to 110% of the System capacity.

[illegible]

**ROUTINE TEST RECORD FOR LOAD / DIGITAL INDICATOR / TANK / HOPPER / SILO
WEIGHING SYSTEM**

SL No	Test Performed	Checked: YES/NO
6.	PF/ Relay Contacts: Check for the relay change over for the set points.	
7.	Display Annunciation: (As per the software Details)	
8.	Communication: RS 232 / RS 485 / BCD / Other Communication Protocols (If any specify) The RS232 / RS485 to be checked with interfacing to PC	
9.	Any other parameters (If Any)	
10.	Rating of the fuse used in the system:	

REMARKS (If ANY): -

	LC 1	LC 2	LC 3	LC 4	LC 5	LC 6	LC 7	LC 8
SPAN								
TARE								
EXC V								

Instrument Used:

Tested by:

Date:

Approved by:

(HOD – PDN)