

Test Report for Miscellaneous items: Production Systems

Job No: JOB-2026-902

Model No: LC-5000

Customer: JSW STEEL

Sl. No: LC-101 to LC-110

Item Description: Load Cell Testing

Qty: 10 Nos

Test Readings and Result:

SL NO	LOAD CELL SERIAL NO	INPUT RESISTANCE(Ω)	OUTPUT RESISTANCE(Ω)	INITIAL UNBALANCE(Mv)
1	LC-101	350.2	350.0	0.02
2	LC-102	350.2	350.0	0.02
3	LC-103	350.2	350.0	0.02
4	LC-104	350.2	350.0	0.02
5	LC-105	350.2	350.0	0.02
6	LC-106	350.2	350.0	0.02
7	LC-107	350.2	350.0	0.02
8	LC-108	350.2	350.0	0.02
9	LC-109	350.2	350.0	0.02
10	LC-110	350.2	350.0	0.02

Tested by: Tester A

Approved by: HOD-PDN

Date: 18-Sep-2026

(HOD-PDN)