

TEST REPORT FOR ACCELEROMETER: MODEL NO PG109M0/PG114M0

Job No:

Customer Name:

Serial No:

Test 1. Axial Sensitivity

Linearity test

Frequency: 50Hz

Gain of the standard charge amplifier (G1) = 1.022mV/pC

Acceleration from Std A (g)	Charge amplifier O/p of Test accelerometer V1 mV (Pk-Pk)	Charge O/p of Test accelerometer $V2 = \frac{V1}{G1}$ (pC)	Sensitivity of Test accelerometer $S = \frac{V2}{A}$ (pC/g)
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			

Mean: pC/g

Frequency test

Gain of the standard charge amplifier(G1)= 1.022mv/pC

Frequency (Hz)	Acceleration from Standard A (g)	Charge Amplifier output of Test Accelerometer V1 mV(Pk-Pk)	Charge output of test accelerometer $V2 = \frac{V1}{G1}$ (pC)	Sensitivity of test accelerometer $S = \frac{V2}{A}$ (pC/g)
50	1			
100	1			
200	1			
300	1			
400	1			
50	10			
100	10			
200	10			
300	10			
400	10			

Mean: pC/g

Test 2. Post Sealing Axial Sensitivity

Linearity test

Frequency: 50Hz

Gain of the standard charge amplifier (G1) = 1.022mV/pC

Acceleration from Std A (g)	Charge amplifier O/p of test accelerometer V1 mv (pk-pk)	Charge O/p of test accelerometer $V2 = \frac{V1}{G1}$ (pC)	Sensitivity of test accelerometer $S = \frac{V2}{A}$ (pC/g)
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			

Mean: pC/g

A) CHARGE SENSITIVITY :

B) CAPACITANCE :

C) INSULATION RESISTANCE :

TESTED BY :

DATE :

APPROVED BY:

HOD-PDN