

ROUTINE TEST RECORD FOR BELT SCALE

JOB No: BS-001
CAPACITY : 500 TPH

CUSTOMER NAME: ACME Ltd
CONVEYOR NO : CV-01

BELT SCALE	LOAD CELL	SPEED SENSOR	JUNCTION BOX
Model No: BSW-4000 Serial No: A001 S/W Version: V2.5 REMOTE: Model No: RMT-100 Serial No: R001	Model No: CAS-500 Serial No: 1) 101 5) 105 2) 102 6) 106 3) 103 7) 107 4) 104 8) 108	Model No: SPD-200 Serial No: SP01 ANGLE SENSOR Model No: Serial No:	Model No: JB-100 Serial No: J001 Model No: Serial No: Model No: Serial No:

Check for short between Line & Neutral, Neutral & Earth, Line & Earth : Checked and found ok

SL. No.	Test Performed	Specified	Actual
1.	Input Supply Voltage (+/-10%)	230V	231V
2.	Derived DC Voltages (+/-5%)	+5V	+5.1V
3.	Display and Keypad Functionality	OK	OK
4.	Belt Load as specified in G.A. or MPL	190	190.5 Kg/m
5.	Belt Speed as specified in G.A. or MPL	1.75	1.76 m/s
6.	Number of PF Contacts	OK	OK
7.	Communication Output (Specify)	OK	OK
8.	Analog Output (Specify number)	4-20mA	4-20mA
9.	Wiring, TB and Component Layout	OK	OK

Resolution: L: 0.01 S: 0.01 r: 0.1 t: 0.1

Measurement to be done from 0 to 110% of the Belt Scale capacity.

Load on Load cell Kg/Mv	Belt load Kg/m	Speed in m/s	Rate in tph	Totalizer in Tones (6 min)	O/p current mA
0	0	0.00	0	0	4.0
50	25.0	1.75	157	15.7	8.0
100	50.0	1.75	315	31.5	12.0
150	75.0	1.75	472	47.2	16.0
200	100.0	1.75	630	63.0	20.0

Instrument Used: FLUKE 87V
Tested by: John
Date: 20/09/2026

Approved by: Smith
(HOD - PDN)