

TEST REPORT FOR ACCELEROMETER: MODEL NO PG109M0/PG114M0

Job No:

Customer Name:

Serial No:

Test 1. Axial Sensitivity

Linearity test

Frequency: 50Hz

Gain of the standard charge amplifier (G1) = 1.022mV/pC

Acceleration from Std A (g)	Charge amplifier O/p of Test accelerometer V1 mV (Pk-Pk)	Charge O/p of Test accelerometer $V2 = \frac{V_1}{G_1}$ (pC)	Sensitivity of Test accelerometer $S = \frac{V_2}{A}$ (pC/g)
1		{{spec_1}}	{{act_1}}
2		{{spec_2}}	{{act_2}}
3		{{spec_3}}	{{act_3}}
4		{{spec_4}}	{{act_4}}
5		{{spec_5}}	{{act_5}}
6		{{spec_6}}	{{act_6}}
7		{{spec_7}}	{{act_7}}
8		{{spec_8}}	{{act_8}}
9		{{spec_9}}	{{act_9}}
10		{{spec_10}}	{{act_10}}

Mean: pC/g

Frequency test

Gain of the standard charge amplifier(G1)= 1.022mv/pC

Frequency (Hz)	Acceleration from Standard A (g)	Charge Amplifier output of Test Accelerometer V1 mV(Pk-Pk)	Charge output of test accelerometer $V2 = \frac{V_1}{G_1}$ (pC)	Sensitivity of test accelerometer $S = \frac{V_2}{A}$ (pC/g)
50	1	{{spec_1}}	{{act_1}}	
100	1	{{spec_2}}	{{act_2}}	
200	1	{{spec_3}}	{{act_3}}	
300	1	{{spec_4}}	{{act_4}}	
400	1	{{spec_5}}	{{act_5}}	
50	10	{{spec_6}}	{{act_6}}	
100	10	{{spec_7}}	{{act_7}}	
200	10	{{spec_8}}	{{act_8}}	
300	10	{{spec_9}}	{{act_9}}	
400	10	{{spec_10}}	{{act_10}}	

Mean: pC/g

Test 2. Post Sealing Axial Sensitivity

Linearity test

Frequency: 50Hz

Gain of the standard charge amplifier (G1) = 1.022mV/pC

Acceleration from Std A (g)	Charge amplifier O/p of test accelerometer V1 mv (pk-pk)	Charge O/p of test accelerometer $V2 = \frac{V1}{G1}$ (pC)	Sensitivity of test accelerometer $S = \frac{V2}{A}$ (pC/g)
1		{{spec_1}}	{{act_1}}
2		{{spec_2}}	{{act_2}}
3		{{spec_3}}	{{act_3}}
4		{{spec_4}}	{{act_4}}
5		{{spec_5}}	{{act_5}}
6		{{spec_6}}	{{act_6}}
7		{{spec_7}}	{{act_7}}
8		{{spec_8}}	{{act_8}}
9		{{spec_9}}	{{act_9}}
10		{{spec_10}}	{{act_10}}

Mean: pC/g

A) CHARGE SENSITIVITY :

B) CAPACITANCE :

C) INSULATION RESISTANCE :

TESTED BY :

APPROVED BY:

DATE {{cal_in_1}}

{{cal_out_1}}