

ROUTINE TEST RECORD FOR CRANE SCALE

JOB No. :
CAPACITY : kg/ Tonnes
Load cell. Model No. :

CUSTOMER:
STAMP NO:
Serial No. :

Crane scale	Battery charger	Remote	Hook	Bow shackle
Model No.	Model No:	Model No:	Sl. No:	Sl. No:
Serial No.	Sl. No:	Sl. No:		

SL. No.	Test Performed	Specified Values	Actual Values
1.	Open Battery Voltage	6.5 V $\pm$ 1 V	
2.	Check for function of all the switches on Main unit and Remote		
3.	Check the Wi-Fi/ Communication Function		

Resolution : L: kg/tonnes

Measurement to be done from 0 to 100% of the Crane Scale capacity.

Load on Load cell Kg/T or mv	Measured load Kg/T

- 1) Battery Charger voltage
When the battery connected:
- 2) Low battery voltage set (CS):
- 3) Check for the remote operation:
(For the distance of min. 15 meters)

Remarks (if any)

Instrument Used:

Checked by:
Date:

Approved by:
(HOD – PDN)