

TEST RECORD FOR TRIP SAFE

JOB ORDER No. : JOB-TS-008
CUSTOMER: Refinery Safety Unit

TRIPSAFE

LOAD CELL: LC-S-BEAM-500

MODEL NO.: Trip Safe TS-300
SERIAL NO.: SN-TS-1234

MODEL NO.: LC-S-BEAM-500
SERIAL NO.: SN-LC-5501

Linearity Test:

1	LOAD ON LOAD CELL		0.00 mV	LED 1 (Green)
	0 Kg	0.0 mV		
2	200 Kg	4.0 mV	4.01 mV	LED 2 (Green)
3	400 Kg	8.0 mV	7.99 mV	LED 4 (Yellow)
4	600 Kg	12.0 mV	12.02 mV	LED 6 (Yellow)
5	800 Kg	16.0 mV	15.98 mV	LED 8 (Red Trip)
6	1000 Kg	20.0 mV	20.01 mV	LED 10 (Full Trip)
6				

Percentage of error for the Trip Level: 0.01 %

PF Contact for Healthy / Over Load: Healthy Relay Contacts OK

Lamp indication / LED annunciation as applicable: LED Bar Graph Annunciation OK

Instrument Used: Fluke 87V & Load Cell Calibrator

NOTE: Ensure that the Jumper/Shorting as per the Required Underload/Overload

Tested by: Safety Officer
Date: 2026-08-22

Approved by: HOD-PDN
(HOD - PDN)