

**ROUTINE TEST RECORD FOR LOAD INDICATOR / DIGITAL INDICATOR / TANK /
HOPPER / SILO WEIGHING SYSTEM/CRANE WEIGHING SYSTEM**

SYSTEM :

JOB No. :

CAPACITY: Kg/Tonnes

CUSTOMER:

TAG. NO.:

Load/Digital Indicator.	SENSOR	REMOTE (IF ANY)	JUNCTION BOX
Model No.	Model No.	Model No.	Model No.
Serial No.	Serial No.	Serial No.	Serial No.
	1)		Model No.
	2)		Serial No.
	3)		
	4)		

Check for short between Line & Neutral ,
Neutral & Earth , Line & Earth :

SL · No ·	Test Performed	Specified Values	Actual Values
1.	Input Supply Voltage (+/-10%)		
2.	Derived DC Voltages (+/-5%)		
3.	Check for the power on display and the Keyboard function.		
4.	Load on the Load cell as specified in G.A. or MPL		Kg/T
5.	Tare weight as specified in G.A. or MPL		Kg/T

Resolution: L: kg/T

Measurement to be done from 0 to 110% of the System capacity.

Load on Load cell Kg/T or mV	Measured Load kg/T (1 to 4 Load cell system)				Current output mA	Measured Load kg/T (5 to 8 Load cell system)			
	1	2	3	4		5	6	7	8

**ROUTINE TEST RECORD FOR LOAD / DIGITAL INDICATOR / TANK / HOPPER / SILO
WEIGHING SYSTEM**

SL · No ·	Test Performed	Checked: YES/NO
6.	PF/ Relay Contacts: Check for the relay change over for the set points.	
7.	Display Annunciation: (As per the software Details)	
8.	Communication: RS 232 / RS 485 / BCD / Other Communication Protocols (If any specify) The RS232 / RS485 to be checked with interfacing to PC {Attach a printout along with the test report for the Printer option}	
9.	Any other parameters (If Any)	
10.	Rating of the fuse used in the system:	

REMARKS (If ANY): -

Instrument Used:

Tested by:

Date:

Approved by:

(HOD – PDN)