

Test Report for Miscellaneous items: Production Systems

Job No: **JOB-MISC-9901**

Model No: **As Detailed Below**

Customer: **JSW Steel Works**

Sl. No: **As Detailed Below**

Item Description: **As Detailed Below**

Qty: **7**

Test Readings and Result:

A. LOAD CELL RESISTANCE & INITIAL UNBALANCE TEST READINGS: (Model: LC-350-S-TYPE)

Sl. No	Serial Number	Input Res (Ω)	Output Res (Ω)	Initial Unbal (mV/V)	Status
1	SN-LC-101	350.2	350.5	0.02	PASSED
2	SN-LC-102	350.1	350.4	0.01	PASSED
3	SN-LC-103	350.4	350.6	-0.03	PASSED
4	SN-LC-104	350.0	350.3	0.02	PASSED

Result: The above mentioned load cells were checked for input resistance, output resistance, and initial unbalance using Fluke 87V Digital Multimeter and were found OK.

B. JUNCTION BOX INSPECTION:

Model No: JB-04-SS-IP67 | Type: 4-in 1-out Junction Box | Qty: 1 | SI No: SN-JB-8821

Result: The above mentioned junction boxes were checked for cross-continuity, continuity, and output functionality and were found OK.

C. SPEED SENSOR / TACHO INSPECTION:

Model No: BCTWDS-SPD-PRO | Indicator Model: LH32220 | Qty: 1 | SI No: SN-SPD-1029

Result: The above mentioned speed sensor(s) were checked for speed with indicator model LH32220 and were found OK.

D. REMOTE DISPLAY / INDICATOR INSPECTION:

Type: 100mm Remote Display | Model No: RD-100-LARGE | Master Indicator: LS 32200 | SI No: SN-RD-901

Result: The above mentioned 100mm Remote Display unit(s) were checked for communication with indicator model LS 32200 and were found OK.

E. ADDITIONAL INSPECTION & TEST FINDINGS:

Tested by: **System Technician**

Approved by: **Quality Manager**

Date: **2026-08-22**

(HOD-PDN)