

BIS-Referenced Refrigerator Tender Validation Demo

This is a software demonstration data pack. It contains metadata and procurement test requirements based on BIS-published listings. It does not reproduce the text of any Indian Standard and is not an official BIS document.

Standard	BIS-listed title	Use in demo
IS 17550 : Part 1 (2024)	Household Refrigerating Appliances - Characteristics and Test Methods Part 1 General Requirements (First Revision)	Registry match
IS 17550 : Part 2 (2024)	Household Refrigerating Appliances - Characteristics and Test Methods Part 2 Performance Requirements (First Revision)	Registry match
IS 17550 : Part 3 (2025)	Household Refrigerating Appliances - Characteristics and Methods Of Test Part 3 Energy Consumption and Volume (First Revision)	Registry match
IS 18689 (2024)	Household and Similar Electrical Appliances - Safety - Particular Requirements for Commercial Refrigerating Appliances	Registry match

Demo tender requirements

- R-01: Supply commercial laboratory refrigerators.
- R-02: Minimum usable capacity of 300 litres per unit.
- R-03: Controlled temperature range of 2 °C to 8 °C.
- R-04: Address applicable general requirements with IS 17550 Part 1:2024.
- R-05: Address performance requirements with IS 17550 Part 2:2024.
- R-06: Provide energy-consumption and volume evidence with IS 17550 Part 3:2025.
- R-07: Address applicable electrical safety requirements with IS 18689:2024.
- R-08: Provide manufacturer/accredited-laboratory test reports.
- R-09: Installation and commissioning.
- R-10: Two-year warranty.

Expected application behavior: R-04 to R-07 should receive explicit-reference matches when the four corresponding verified Standard Registry records are present. R-01, R-02, R-03, R-08, R-09 and R-10 may remain missing or require review because they are procurement/product requirements rather than direct standard references.

Authoritative references: BIS Product Manual for IS 17550 (Part 1):2024; BIS LIMS listings for IS 17550 Parts 1-3 and IS 18689.