

SYNTHETIC TENDER — LABORATORY REFRIGERATORS

Tender Reference: TEST-LAB-REF-2026-002
Purpose: Controlled test of the Tender Validation workflow.

Testing disclaimer: This is a synthetic tender for software testing. It is not an official government tender. Standard references below are test references and must be verified for applicability before any real procurement.

1. Scope of Procurement

The purchaser invites bids for the supply, installation, commissioning and warranty support of 20 commercial laboratory refrigerators for a government research laboratory. Units shall be new, unused and suitable for continuous operation in a laboratory environment.

2. Minimum Technical Requirements

Req. ID	Requirement
R-01	Supply of 20 commercial laboratory refrigerators.
R-02	Usable capacity shall be minimum 300 litres per unit.
R-03	The refrigerator shall maintain a controlled temperature range of 2 °C to 8 °C.
R-04	The unit shall provide a digital electronic temperature controller and external digital temperature display.
R-05	The unit shall provide high-temperature, low-temperature and door-open alarms.
R-06	Electrical supply shall be 230 V AC, single phase, 50 Hz.
R-07	The offered commercial refrigerating appliance shall address applicable electrical safety requirements with reference to IS 18689:2024, subject to applicability verification.
R-08	Performance and storage characteristics shall be supported by test evidence and shall identify the applicable edition of IS 17550 Part 2:2024 or another authoritative applicable standard.
R-09	Energy consumption and rated volume evidence shall identify the applicable edition of IS 17550 Part 3:2025 or another authoritative applicable standard.
R-10	General refrigerating-appliance requirements shall identify the applicable edition of IS 17550 Part 1:2024 or another authoritative applicable standard.
R-11	The bidder shall provide manufacturer or accredited laboratory test reports supporting electrical safety and performance characteristics.
R-12	Supplier shall install and commission each unit at site.
R-13	Minimum warranty shall be two years from successful commissioning.

3. Standards and Compliance

The bidder shall clearly identify applicable Indian Standards and BIS requirements. The tender deliberately contains explicit references to IS 17550 Part 1:2024, IS 17550 Part 2:2024, IS 17550 Part 3:2025 and IS 18689:2024 so that the validation engine can exercise explicit-reference matching. The bidder shall provide current certificates, licences and test reports where applicable.

4. Bidder Compliance Matrix

Req. ID	Bidder Response	Evidence
R-01	Comply / Deviate	Document reference
R-02	Comply / Deviate	Document reference
R-03	Comply / Deviate	Document reference
R-04	Comply / Deviate	Document reference

Req. ID	Bidder Response	Evidence
R-05	Comply / Deviate	Document reference
R-06	Comply / Deviate	Document reference
R-07	Comply / Deviate	Document reference
R-08	Comply / Deviate	Document reference
R-09	Comply / Deviate	Document reference
R-10	Comply / Deviate	Document reference
R-11	Comply / Deviate	Document reference
R-12	Comply / Deviate	Document reference
R-13	Comply / Deviate	Document reference

Expected software test outcome

If the four matching Standard Registry records are present and processed/indexed, the Tender Validation workflow should detect the explicit standard references and produce matched or partially matched requirements rather than a 0% coverage result. Requirements such as quantity, installation and warranty may remain unmatched if no registry standard addresses them; that is an expected outcome, not necessarily a software failure.