# Meridian Labs - Equipment Maintenance Manual

Document ID: MM-2025-07
Revision: 3
Issue date: 12 March 2025
The pressure valve must be replaced after 4500 operating hours.
The inspection interval is every 90 days.

# Spare Parts and Costs

Filter Cartridge, part number FC-101, costs 45 USD.
Seal Ring, part number SR-202, costs 18 USD.
Pressure Valve, part number PV-303, costs 320 USD.
The recommended spare-parts budget for 2025 is 3800 USD.

# Warranty and Support

The warranty period is 24 months from the issue date.
Support is available on weekdays only.
The service hotline operates from 9 AM to 5 PM.
Support email: support@meridianlabs.example