

Sigma- 20

VITAL SIGN-SPO2 & NIBP P- ULTRA (Pulse Oximeter)





- Lasting working capability with built-in rechargeable lithium battery
- 400 groups of NIBP list, 60 alarm event record
- Central monitoring system applicable to ICU/CCU/OT
- Support solution: wall mounting, mobile trolley





Sigma- 20

VITAL SIGN-SPO2 & NIBP

P- ULTRA (Pulse Oximeter)



Features:

- 8 inch high definition color TFT display
- Standard 3-parameter : NIBP, SP02 & PR
- 2-4 hours battery power supply
- Sound, light & text alarm, adjustable volume
- Connected to central monitoring System

NIBP:

Method: Oscillometry

Modes: Manual/Auto/Continuous Auto measure: time Adjustable

Measurement range: 10-270mmHg

Size: L9.5 * W6.5 * H8.5 inch

Weight: 1.700 kg

Electrical Specification:

100-240v AC, 50/60Hz

Max Input Power: 70VA

Battery:

11.1V 4000mAh rechargeable lithium battery

Standard configuration:

Sp02, PR, NIBP

SP02:

Measurement range: 0-100%

Accuracy: 70-100%, ±2% 0-69%,

upspecified

Pulse rate:

Range 20-254bpm

Accuracy ±3bpm

Respiration:

Method RA-LL impedance

Range

Adult: 0-120bpm

pediatric/neonatal:0-150bpm

Apnea alarm: Yes

Accure Medical Pvt. Ltd.

Ground Floor Plot no.-183, Pocket N, Sector-2, DSIIDC Bawana Industrial Area, North West Delhi-11003 Contact Number: +91 8077818101 Mail id: - ampl2207@gmail.com, Web: www.accuremedical.in

