AFRICA CENTRE OF EXCELLENCE FOR SUSTAINABLE POWER AND ENERGY DEVELOPMENT
UNIVERSITY OF NIGERIA, NSUKKA
OFFICE OF THE ACE-SPED TESTING FACILITY
☎: (+234) 8028433576

Ref: UNN/ACE-SPED/ILO

Date: 25th JULY, 2024

Certification documents from relevant agencies

In order to maintain standards, ACE-SPED has partnered with Katchey Company Limited, situated at 26 Odanye Close, Harmony Enclave, Adeniyi Jones, Ikeja, Lagos, to certify their equipment for SON importation and ISO. Additionally, the University of Nigeria's Nsukka's National Equipment Maintenance has certified other equipment. The following list showcases some of the certifications that Katchey Company Limited has carried out:

Declaration of Conformity

We, Thermo Fisher Scientific Brno s.r.o., part of Thermo Fisher Scientific
Vlastimila Pecha 1282/12, Černovice, 627 00 Brno
Czech Republic

an ISO9001 certified company

declare under our sole responsibility that the electron microscope

Product Name: **Axia ChemiSEM LoVac**

Serial Number: 9961704

Meets the provisions of the directives:

2006/42/EC	Machinery directive
2014/30/EU	EMC directive
2011/65/EU	RoHS directive
2013/59/EURATOM	Ionizing radiation directive

using to the following standards and normative documents:

EN 61326-1	Electrical equipment for measurement, control and laboratory use – EMC requirements
EN 61010-1	Safety measurements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements

and that the X-ray emission is below 1 µSv/h at 10 cm distance from the surface

Peter Walk
Sr. Regulatory Compliance Manager

Brno, 5.2.2024

ThermoFisher SCIENTIFIC
The world leader in serving science

Declaration of Conformity

We,

Thermo Fisher Scientific
Molecular and Microanalysis Products
5225 Verona Rd.
Madison, WI 53711
Ph: +1 608 276 6100

declare under our sole responsibility that the product

Nicolet™ iS™20 FT-IR spectrometer
Nicolet™ iZ10 module

To which this declaration relates is in conformity with the following standard(s) or other normative document(s)

EN 61010-1:2010 (Safety requirements for electrical equipment for measurement, control, and laboratory use)
EN 60825-1:2014 (Safety of laser products)
EN 61326-1:2013 (Electrical equipment for measurement, control, and laboratory use. EMC requirements)
EN 50581:2012 (Assessment of electrical and electronic products with respect to the RoHS)

following the provisions of Directive(s) 2014/30/EU (Electromagnetic Compatibility), 2014/35/EU (Low Voltage), and 2011/65/EU (RoHS).

The Technical Construction file is maintained at:

Thermo Fisher Scientific
Molecular and Microanalysis Products
5225 Verona Rd.
Madison, WI 53711
Ph: +1 608 276 6100

The authorized representative located within the Community is:

Thermo Fisher Scientific
Dieselstr. 4
76227 Karlsruhe, Germany
Ph: +49 (0) 721 4094 389

Date of issue: 04 (DD) DEC (MMM) 2018 (YYYY)

Place of issue: Madison, Wisconsin, USA

Signed:

Carlo Nardini
Director of Operations

CE

ThermoFisher SCIENTIFIC

All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries

ThermoFisher
SCIENTIFIC

Shipment Authorization/ Certification of Conformance

Customer Name: **AT - Vienna**

Job Number: **4501007902** **System Serial Number:** **2080-2037**

☑ The undersigned certifies this tested instrument meets or exceeds applicable Product Acceptance Criteria.

Ken Compton
Ken Compton
Validating Technician

26/Apr/2021
Date

Year of Manufacturing: 2021

Country of Origin: USA:

☑ All components of your order are present unless otherwise noted on the Notification of Partial Shipment form included in this envelope.

I authorize shipment of this order:

Ken Compton
System Configuration Specialist

26/Apr/2021
Date

If you have any problems conserning this order, please contact your nearest Sales/Service Office or Representative.

Thermo Fisher Scientific 5225 Verona Road Madison, WI 53711 (608) 276-6100
www.thermofisher.com

Multiskan SkyHigh Performance Test Results

Date:	26/11/2021
Instrument serial number:	1550-801287C
Instrument internal software version:	2.00.35
Instrument UI software version:	1.0.70.0
Skanlt software version:	6.1
Factory performance calculations file version:	1.0
Spectrophotometric verification plate serial number:	1911-04
Cuvette standard set	34152
Testing technician:	Eddie Ho

Linearity

96-well plate, 0 - 2.5 Abs, < 2 % @ 450 nm	PASSED

Precision

Fast mode, SD < 0.003 Abs or CV % < 1 % @ 450 nm	PASSED
Precision mode, SD < 0.003 Abs or CV % < 0.5 % @ 450 nm	PASSED

Accuracy

Plate, 0 - 2 Abs, < (1 % + 0.003 Abs), 2 - 2.5 Abs, <2%	PASSED
Cuvette, 0 - 2 Abs, < (1 % + 0.003 Abs), 2 - 2.5 Abs, <2%	PASSED

Spectral stray light

Spectral stray light, < 0.05 % @ 230 nm	PASSED

Wavelength accuracy

Plate, < 2 nm	PASSED
Cuvette, < 2 nm	PASSED

Incubator

Plate accuracy, ± 1.0°C @ 37°C	PASSED
Cuvette accuracy, ± 1.0°C @ 37°C	PASSED

Shaker

Operational test	PASSED

Name: EDDIE HO **Date:** 26 NOV 2021

For internal use only. This document contains confidential, proprietary information of Thermo Fisher Scientific Inc. It may not be reproduced or copied without prior written permission from Thermo Fisher Scientific Inc. Hardcopy versions of this document must be verified to the current revision level.

Thanks

Engr. Prof. V.S. Aigbodion **FNMS, SA/NRF-C3**
Industrial Liaison Officer
Head-ACE-SPED TESTING FACILITY
+2348028433576
victor.aigbodion@unn.edu.ng