

Sea-Bird Scientific 13431 NE 20th Street Bellevue, WA 98005 +1 425-643-9866 seabird@seabird.com www.seabird.com

SENSOR SERIAL NUMBER: 21145 CALIBRATION DATE: 28-Aug-19 SBE 37 V2 TEMPERATURE CALIBRATION DATA ITS-90 TEMPERATURE SCALE


COEFFICIENTS:

a0 = -1.797573e-004 a1 = 3.200483e-004 a2 = -5.298528e-006 a3 = 2.218687e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP· (° C)	RESIDUAL (° C)
1.0000	569731.8	1.0001	0.0001
4.5000	487756.3	4.4999	-0.0001
15.0000	312103.7	15.0000	0.0000
18.5000	270627.0	18.5001	0.0001
24.0000	217598.4	23.9999	-0.0001
29.0000	179556.3	28.9999	-0.0001
32.4999	157481.5	32.5000	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a0 + a1[ln(n)] + a2[ln^2(n)] + a3[ln^3(n)]\}$ - 273.15 Residual (°C) = instrument temperature - bath temperature



Page 2



SBE37-SMP-ODO MicroCAT

Instrument Configuration

Instrument Serial Number:

37-21145

Instrument Firmware Version:

6.1.2

Zero Conductivity Frequency:

2647.35 RS232

Communications Format: Communications Settings:

9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

Data Format	Measurement	Sensor Type	Serial Number	Rating
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure Sensor	Kistler	5044127	2000m(2000 dBar)
RS232	Oxygen	SBE 63	63-2318	7000m

Maximum Depth:

2000m

CAUTION - The maximum deployment depth will be limited by the measurement range of the pressure sensor, if installed, an attached sensor, if installed, or the housing.

