



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005515096

03-MAR-2022

320049500

SERVICE REPORT

Service Request

Date

Sales Order

CUSTOMER INFORMATION

Name: UNIVERSITY OF ALASKA FAIRBANKS
Account : 40299784
HANK STATSCEWICH
HANK.STATS@ALASKA.EDU
9073223470

PO Number:
(Hank Statscewich)

Bill To Address

UNIVERSITY OF ALASKA FAIRBANKS
9150 KOYUKUK DRIVE, STE 245 O'NEILL BLDG
FAIRBANKS, AK, 99775, US

Ship To Address

UNIVERSITY OF ALASKA FAIRBANKS
9150 KOYUKUK DRIVE, STE 245 O'NEILL BLDG
FAIRBANKS, AK, 99775, US

PRODUCT INFORMATION

Item: 37SMP.36203

Item Description: SBE37SMP-ODO V4, TI, 2000M, MCBH, 5000M ODO

Serial: 37-21147

Special Notes

Services Requested:
O-RINGS_PT
SERVICE37

No problems found.

Services Performed:
Performed initial diagnostic evaluation.
Performed pressure calibration.
Performed "POST" cruise calibration.
Replaced the O-rings.
Performed a hydrostatic pressure test.
Performed complete system check and full diagnostic evaluation.
Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
SERVICE37	Complete External Inspection. Test All Functions. Calibrate C & T, C&P and recal if necessary. Replace end cap O-rings and Hydrostatic pressure test if necessary. Includes typical repairs. (Does not include: Major electronic board replacement. Instruments deemed "Damaged Beyond Repair". Missing/damaged mounts or cables. Upgrades or modifications. Spares or antifoulant replacement.)	1
REPLACEAF	Extra charge to install one antifoulant device, includes one 801542.1.	2
PCAL37	Extra charge to calibrate SBE 37 pressure channel	1

Unbilled Items

Item	Item Description	Qty
------	------------------	-----



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005515096

03-MAR-2022

320049500

SERVICE REPORT

Service Request

Date

Sales Order

801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	2
----------	---	---



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005515096

03-MAR-2022

320049500

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 63.21

Item Description: SBE63, OPTICAL OXYGEN, MICROCAT, 7000m

Serial: 63-2323

Special Notes

Services Requested:

Evaluate/Repair Instrumentation.

Perform Routine Calibration Service.

No problems found.

Services Performed:

Performed initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_63	SBE 63 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty