Complementary classification of solid oral dosage forms in ambient conditions by desorption electrospray ionization mass spectrometry and transmission Raman spectroscopy

This manuscript (<u>permalink</u>) was automatically generated from <u>adamjtaylor/tablet-classification@7e88c16</u> on October 14, 2021.

Authors

Adam J Taylor

Dimitrios Tsikritsis

National Physical Laboratory, Teddington, UK

Alex Dexter

National Physical Laboratory, Teddington, UK

Amy Burton

National Physical Laboratory, Teddington, UK

Josephine Bunch

National Physical Laboratory, Teddington, UK; Imperial College London, London, UK; The Rosalind Frankilin Insitute, Harwell, UK

Natalie Belsey

National Physical Laboratory, Teddington, UK

Abstract

Discrepancies or defects in active ingredients, excipients and coatings that form solid oral dosage forms can both impact product quality and provide hallmarks of off-brand or counterfeit products. There is therefore a need for rapid and continuous analytical techniques that can assess and classify product differences of intact samples at- or near the production line, or in analytical labs, ideally without resorting to product dissolution.

Here we test the ability of two rapid ambient chemical characterization methods to discriminate between solid dosage forms: desorption electrospray ionization mass spectrometry and transmission Raman spectroscopy. These two techniques are highly complementary, offering greater sensitivity to the analysis of the surface and the tablet bulk, respectively. The data sets generated were then used to test a variety of classification algorithms including linear discriminate analysis, tree-based methods, a simple neural network, and support vector machines (SVM). The highest performing algorithms for DESI-MSI were the SVM, with an additional performance boost when used with a polynomial kernel. For transmission Raman data, a linear discriminant analysis (LDA) model was found to be the most effective.

Introduction

Inconsistencies in active ingredients, excipients, the thickness and integrity of coatings and the presence of impurities in solid oral dosage forms all negatively affect their performance. Inferior quality attributes can be useful to identify off-brand or counterfeit products. There is a need for rapid and continuous analytical techniques that can assess and classify product differences of intact samples at- or near the production line, or in analytical labs, ideally without resorting to product dissolution [1]. Rapid measurement tools are particularly important to enable continuous monitoring, necessary to support the change from batch to continuous manufacturing. Analytical methods are required to monitor both the actives, coatings and consistency of the product: For example, in addition to the total API content, insight is also needed on degradation products, impurities, (co-) crystallinity/presence of polymorphs, and content uniformity. The ability to monitor tablet coating thickness and integrity is of great importance, particularly for functional coatings, such gastroresistance, which would be compromised by insufficient thickness, or the occurrence of cracks in the film [2].

Quantitative analysis of pharmaceutical tablets is routinely performed by HPLC which offers accurate and sensitive measurements of the active ingredient(s) and excipients, in addition to the presence of any contaminants. However, solution-based analytical methods are destructive and labor-intensive.

Mass spectrometric methods can provide unlabeled identification, both of expected ingredients in known samples and of contaminants or components of unknown formulations. Ambient ionization mass spectrometry approaches including DESI (desorption electrospray ionisation) and DART (direct analysis in real time) facilitate the desorption and ionization from the surface of samples at atmospheric conditions, without dissolution or additional sample preparation. They are therefore potentially useful tools for rapid assessment of solid oral dosage forms

Optical spectroscopy techniques offer rapid, non-destructive analysis, including polymorphic identification [3], and are also able to measure insoluble ingredients. They have consequently been exploited for in-line process analytical testing and as quality control tools [1]. For example, infra-red-based techniques (FTIR and DESI [4]; and NIR and Mid-IR spectroscopy classification of MDMA containing tablets [5] Near infrared is the most commonly used process analytical tool [1], however Raman spectroscopy provides complementary information and has grown in popularity in recent

years, since it provides more distinct spectral features, and is better-suited to analysis in aqueous environments owing to the relatively weak strength of the Raman O-H band. Technological advancements have facilitated miniaturization, increased speed and reduced cost, resulting in more widespread implementation. [6]

Ambient ionisation mass spectrometry of tablets

Desorption electrospray ionization uses a charged electrospray of organic solvent which, when directed at the sample surface in proximity to the mass spectrometry inlet, desorbs ions from the sample which may be taken up into a mass spectrometer [7]. As this process takes place at ambient pressure and with a flexible geometry, the technique is suited for the analysis of a wide range of samples including explosives on surfaces [8], fingerprints [9], plants [10] and tissues [11,12]

One of the early descriptions of DESI-MSI was in the profiling of tablets [13]. Chen et al demonstrated the use of DESI-MS to profile tablets containing loratadine, folic acid, acetaminophen (paracetamol), aspirin, melatonin or caffeine. Optimization of DESI parameters including voltage, solvent delivery and capillary temperature facilitated analysis at up to three scans per second. Subsequent studies using DESI-MS of tablets have focused on targeted analysis for active ingredients. For example, the identification MDMA and amphetamine derivatives in ecstasy tablets [14], counterfeit artesunate antimalarial tablets [15,16] and antiviral capsules [17].

For ambient mass spectrometry to be deployable in the field for counterfeiting applications, or in manufacturing environments for QA/QC, the mass spectrometer must be compact. Several designs for small field-deployable mass spectrometers have been demonstrated with DESI MS sources [18,19].

Each of these applications has targeted expected components of the tablet of interest, predominantly active ingredients or excipients. However, in manufacturing QA/QC and counterfeit-detection applications, additional information on unexpected changes in active or excipient source or quality, as well as the introduction of contaminants may be of importance. Untargeted multivariate and machine learning approaches are therefore of interest to determine differences between samples using all spectral information.

Classification approaches for mass spectrometry applications are proving powerful in a range of applications. The two most widespread applications of classification in mass spectrometry are in disease diagnosis and determination of bacterial type [20]. A range of classification algorithms have been applied to mass spectrometry and spectroscopy data. PLS-DA is most commonly reported, although a range of algorithms including neural networks, and support vector machines [21] have been reported. Several publications have evaluated different classification algorithms but unsurprisingly the optimal algorithm depends greatly on the nature of the input data. A summary of classification and other data analysis for proteomics can be found here [22] and for metabolomics here [23]. Classification approaches are becoming more accessible through modeling tools with consistent grammar and data structure, and their integration into mass spectrometry software [ScilsLab,Waters software] [23].

Notably, classification of rapid evaporative ionization MS enables real-time classification of tissue types during surgery [24]. Classification of REIMS data has also found applications in food security [25] and bacterial speciation [26]. Classification approaches have also been widely employed in mass spectrometry imaging data, particularly in the classification of cancerous tissue [27]

Raman spectroscopy analysis of tablets

Raman spectroscopy exploits the inelastic scattering of light by the sample to reveal valuable chemical and structural information. Information can be obtained from the sample in a non-destructive manner, making it a popular process analytical technology tool. Raman spectroscopy can be performed in a variety of sampling configurations/geometries, the most appropriate depending on the application. Confocal Raman microscopy can provide detailed chemical mapping with high spatial resolution, however this is generally reserved for forensic investigation rather than continuous monitoring, since it requires lengthy acquisition times. Sub-sampling issues associated with conventional backscattered Raman can be overcome by strategies such as sample rotation in conjunction with spectral averaging, or simultaneous wide angle illumination [28].

Matousek et al demonstrated the ability of transmission Raman spectroscopy to probe deep into turbid materials such as pharmaceutical tablets and provide information on their bulk properties [29,30]. In contrast to conventional backscattered Raman, in transmission Raman spectroscopy the beam passes through the full thickness of the tablet, sampling a much larger volume of the material, and consequently provides more representative sampling [31] Although Raman scattering intensity is linear with concentration within the same confocal plane, transmission Raman signal intensity is slightly biased towards the bulk of the tablet relative to the exterior due to internal scattering [32] . In contrast, DESI-MSI sampling is biased towards the surface/coating composition. Therefore, in combination, these two techniques should provide a powerful toolkit with which to assess compositional differences between pharmaceutical tablet formulations.

Raman spectra of complex mixtures such as solid dosage forms often have complicated spectra with overlapping peaks. For this reason, multivariate techniques are often applied to help identify the components of interest and changes in chemistry. The selection and use of unsupervised and/or supervised techniques on Raman spectra rely on factors such as prior knowledge of the raw component spectra, and the quantity and complexity of the spectra.[33]

As with mass spectrometry, classification of Raman spectroscopy data has been primarily focused on disease diagnostics [34,35,36] and bacterial analysis [37,38]. Other noteworthy examples of the use of classification in Raman spectroscopy include differentiation of narcotics [39], pharmaceuticals [40], [doi:https://www.sciencedirect.com/science/article/abs/pii/S0731708506000926?], and counterfeit tablets [41].

There have been relatively few comparisons of different classification methods for Raman spectroscopy data. Zheng et al. compared SVM, LDA and k-nearest neighbours (KNN) methods to classify renin hypertension from Raman data from serum [42]. They found that SVM and LDA performed similarly, and both outperformed the KNN algorithm. Partial least squares (PLS) and PLS discriminant analysis are also commonly used methods in characterizing tablets, however care is required depending on the data quantity and the pre-preprocessing performed [43]. Qun et al. tested the classification of expired drugs using PLS-DA, SVM and KNN, and reported that SVM gave the strongest performance [44]. Fransson et al tested the performance of multivariate methods including PLS, classical least squares (CLS) and multivariate curve resolution (MCR) for classification of pharmaceutical tablets [45].

Objective

In this study we set out to explore the potential of DESI-MSI and transmission Raman spectroscopy to distinguish commercially available pharmaceutical tablets with similar or different formulations. Pairing DESI with transmission Raman was of particular interest due to their complementarity and relative abilities to sensitively probe the surface vs the bulk of the tablets. Classification of tablets based on both active ingredients and excipients has the potential to be used for in-line quality control measures during pharmaceutical manufacturing, and for rapid counterfeit testing. As such we have

tested a range of classification algorithms on their capability to differentiate these tablets using a range of pre-processing methods to determine the best approaches to use in different applications.

Experimnental

Samples

Samples were selected from commercially available of-the-shelf products and purchased from a local supplier. Their names, active ingredients, listed excipients and MHRA product license numbers are included in Table $\underline{1}$

 Table 1: Details of the tablet types analysed. Letter codes for each type are used throughout

Туре	Product name	Active ingredients	Listed excipients	MHRA licence	
Α	Anadin Extra Tablets	300 mg Aspirin, 200 mg Paracetamol, 45 mg Caffeine	Maize starch, microcrystalline cellulose (E460), hydrogenated vegetable oil, hydroxypropyl methylcellulose (E464), polyethylene glycol, pregelatinised starch and povidone	PL 00165/5013R	
В	Tesco Paracetamol Extra Tablets	500 mg Paracetamol, 65 mg Caffeine	tarch, povidone k-30, povidone k-90, croscarmellose sodium, talc, stearic acid and magnesium stearate	PL 08977/0025	
С	Tesco Paracetamol Tablets	500 mg Paracetamol	Potato Starch, pregelatinised starch, magnesium stearate, povidone, stearic acid and talc	PL 08977/0014	
D	Tesco Extra Power Pain Control Tablets	300 mg Aspirin, 200 mg Paracetamol, 45 mg Caffeine	Povidone, hydroxypropylcellulose, stearic acid, microcrystalline cellulose, maize starch, pregelatinised starch, hydroxypropyl methylcellulose 5cPs, hydroxypropyl methylcellulose 15cPs, macrogol 4000	PL 29831/0164	

DESI MS

DESI-MS measurements were performed on a Synapt G2-Si Q-IM-ToF mass spectrometer (Waters Corp). The instrument was operated in 'resolution mode'. The ion mobility cell was not used. Positive ion mode spectra were collected with a scan time of 1 second across a mass range of m/z 50 to m/z 1200. The instrument was fitted with a prototype DESI source (Waters Corp), with the sprayer configured for electroflow focusing with a fused silica capillary sitting approximately 1 mm behind a 200 μ m steel orifice. Methanol with 5% water by volume was delivered at 2 μ l/m by a pressure pump (Dolomite). Nitrogen gas was delivered at 0.2 MPa. The spray voltage was set at 5 kV. A heated inlet capillary was set to a calibrated temperature of 400°C using a PID (Waters Research Centre, Hungary). Tablets were sampled by holding the tablet 1-2 mm away from the DESI spray head using plastic

tweezers. For training data, acquisition was started with the tablet already under the spray head, such that only data from the tablet surface was acquired, while validation data was collected continuously.

Transmission Raman spectroscopy

Transmission Raman spectra were acquired using a Renishaw InVia Qontor Raman microscope equipped with a 830 nm excitation source fibre-coupled to an InVia transmission Raman accessory (Renishaw plc, Wotton-under-Edge, Gloucestershire), in a temperature controlled environment. Light was collected in transmission with the x5 air objective lens (0.12 NA, N-PLAN, Leica). Tablets were carefully placed onto a flat silicon sample support with a hole just smaller than the tablet dimensions, so that the excitation beam was able to pass through the tablet but not the sample support. Six tablets were analysed of each type in pseudo-random order. Three measurement replicates were acquired, each complete data set was collected on three separate days.

For all tablets, extended spectra were acquired using Renishaw Wire (version 5.3) software for the spectral range of 50 to 1800 cm⁻¹, with an acquisition time of 30 seconds, and 5 accumulations. Laser power was set to 100% which has been measured at the sample to be approximately 117 mW. An internal silicon calibration reference spectrum was acquired each day to correct the Raman shift of the data.

Data analysis

All data were analyzed in R version 3.6.2 (2019-12-12) "Dark and Stormy Night" and RStudio Server version 1.2.5019. Analysis was conducted using the tidyverse [46] and tidymodels [47] metapackages. Raman data preprocessing was conducted in MATLAB 2020a. All analysis was performed on a Linux workstation (Intel Core i9-7900X CPU with 10 cores @ 3.30 GHz, 128G RAM, Ubuntu 16.04.6 LTS.

DESI preprocessing

For model development and comparison, data were converted from Waters raw format to mzML format using ProteoWizard MSConvert version 3.0.19239-0ae547798. These were read into R using the mzR package. All spectra were re-binned onto the same mass axis with a bin width of m/z 0.01. A mean spectrum of all training data was peak picked using the findPeaks function from the prcma package with a peak intensity threshold of three times the median intensity of the spectrum. 1217 peaks were found. Each spectrum was then individually integrated across the found peak widths to form a datacube. Each scan of the validation dataset was similarly integrated across the peak widths from the training dataset. Reduced peak datacubes were generated by filtering for the top n most intense peaks. Down-binning to simulate reduced mass resolving power was performed by rounding m/z values and summing intensities within each rounded m/z bin.

Transmission Raman spectroscopy preprocessing

Cosmic ray removal was performed automatically by Renishaw Wire (version 5.3) and spectra exported to .txt format. The Raman spectra were baseline corrected using the msbackadj() Matlab function. The baseline was estimated within multiple shifted windows of width 20 separation units, then a spline approximation was used to regress the varying baseline to the window points. The estimated baseline for each spectrum was then subtracted from the corresponding original. The background subtracted spectra were read in R for subsequent processing and analysis. The data were normalized to total spectrum intensity and the Raman shift recalibrated using the weighted-mean centroid to the 520.7 cm-1 peak from the daily Si wafer sample spectrum as a reference. Extended spectra were truncated to a wavenumber range between 250 cm-1 and 1700 cm-1. Due to the limited

number of wavenumber bins and the challenges of peak-picking Raman data, the continuous data were taken forward for classification.

Classification

Spectra were collated into a 10-fold cross validation 1 with 10 repeats in a 4/1/5 (train/test/total) split. To remove highly co-variate polymer peak sequences leading to overfitting, highly correlating variables (Pearson correlation > 0.9) were removed from DESI MSI data. Data were centered around the arithmetic mean and scaled to have a standard deviation of one. Underrepresented classes (for DESI MSI, the background class) were up sampled. Each training fold was applied to a range of classification algorithms using the tidymodels package. All models were implemented with their default parameters beyond setting to classification mode. The functions, engines and default parameters used for each model are provided in supplementary table 1. These models were then used to predict each testing fold. For each fold the F1 score was calculated. For DESI MSI the algorithm with the highest F1 score, a support vector machine with a polynomial kernel was selected for further model tuning on a single 4/1/5 validation split of the training data. For transmission Raman data a LDA model was selected. A final model was fitted on all the training data. These models were used to predict the test independent test sets. Cosine similarity between spectra were calculated using the cosine function from the coop package [48, =coop]. Considering the angle between vectors, rather than magnitude, cosine similarity provides a useful and robust measure of spectral similarity for highly multivariate datasets [49]

This manuscript is a template (aka "rootstock") for <u>Manubot</u>, a tool for writing scholarly manuscripts. Use this template as a starting point for your manuscript.

The rest of this document is a full list of formatting elements/features supported by Manubot. Compare the input (.md files in the /content directory) to the output you see below.

Basic formatting

Bold text

Semi-bold text

Centered text

Right-aligned text

Italic text

Combined italics and bold

Strikethrough

- 1. Ordered list item
- 2. Ordered list item
 - a. Sub-item
 - b. Sub-item
 - i. Sub-sub-item
- 3. Ordered list item
 - a. Sub-item

- List item
- · List item
- List item

subscript: H₂O is a liquid

superscript: 2¹⁰ is 1024.

unicode superscripts⁰¹²³⁴⁵⁶⁷⁸⁹

unicode subscripts₀₁₂₃₄₅₆₇₈₉

A long paragraph of text. Lorem ipsum dolor sit amet, consectetur adipiscing elit, sed do eiusmod tempor incididunt ut labore et dolore magna aliqua. Ut enim ad minim veniam, quis nostrud exercitation ullamco laboris nisi ut aliquip ex ea commodo consequat. Duis aute irure dolor in reprehenderit in voluptate velit esse cillum dolore eu fugiat nulla pariatur. Excepteur sint occaecat cupidatat non proident, sunt in culpa qui officia deserunt mollit anim id est laborum.

Putting each sentence on its own line has numerous benefits with regard to <u>editing</u> and <u>version</u> <u>control</u>.

Line break without starting a new paragraph by putting two spaces at end of line.

Document organization

Document section headings:

Heading 1

Heading 2

Heading 3

Heading 4

Heading 5

Heading 6



Horizontal rule:

Heading 1's are recommended to be reserved for the title of the manuscript.

Heading 2's are recommended for broad sections such as Abstract, Methods, Conclusion, etc.

Heading 3's and Heading 4's are recommended for sub-sections.

Links

Bare URL link: https://manubot.org

<u>Long link with lots of words and stuff and junk and bleep and blah and stuff and other stuff and more stuff yeah</u>

Link with text

Link with hover text

Link by reference

Citations

Citation by DOI [50].

Citation by PubMed Central ID [51].

Citation by PubMed ID [52].

Citation by Wikidata ID [53].

Citation by ISBN [54].

Citation by URL [55].

Citation by alias [56].

Multiple citations can be put inside the same set of brackets [50,54,56]. Manubot plugins provide easier, more convenient visualization of and navigation between citations [51,52,56,57].

Citation tags (i.e. aliases) can be defined in their own paragraphs using Markdown's reference link syntax:

Referencing figures, tables, equations

Figure 1

Figure 2

```
Figure 3

Figure 4

Table 2

Equation 1

Equation 2
```

Quotes and code

Quoted text

Quoted block of text

Two roads diverged in a wood, and I—I took the one less traveled by, And that has made all the difference.

Code in the middle of normal text, aka inline code.

Code block with Python syntax highlighting:

```
from manubot.cite.doi import expand_short_doi

def test_expand_short_doi():
    doi = expand_short_doi("10/c3bp")
    # a string too long to fit within page:
    assert doi == "10.25313/2524-2695-2018-3-vliyanie-enhansera-copia-i-
        insulyatora-gypsy-na-sintez-ernk-modifikatsii-hromatina-i-
        svyazyvanie-insulyatornyh-belkov-vtransfetsirovannyh-geneticheskih-
        konstruktsiyah"
```

Code block with no syntax highlighting:

```
Exporting HTML manuscript
Exporting DOCX manuscript
Exporting PDF manuscript
```

Figures



Figure 1: A square image at actual size and with a bottom caption. Loaded from the latest version of image on GitHub.



Figure 2: An image too wide to fit within page at full size. Loaded from a specific (hashed) version of the image on GitHub.



Figure 3: A tall image with a specified height. Loaded from a specific (hashed) version of the image on GitHub.



Figure 4: A vector .svg image loaded from GitHub. The parameter sanitize=true is necessary to properly load SVGs hosted via GitHub URLs. White background specified to serve as a backdrop for transparent sections of the image.

Tables

Table 2: A table with a top caption and specified relative column widths.

Bowling Scores	Jane	John	Alice	Bob
Game 1	150	187	210	105
Game 2	98	202	197	102
Game 3	123	180	238	134

Table 3: A table too wide to fit within page.

	Digits 1-33	Digits 34-66	Digits 67-99	Ref.
pi	3.14159265358979323 846264338327950	28841971693993751 0582097494459230	78164062862089986 2803482534211706	piday.org
е	2.71828182845904523 536028747135266	24977572470936999 5957496696762772	40766303535475945 7138217852516642	nasa.gov

 Table 4: A table with merged cells using the attributes plugin.

	Colors	
Size	Text Color	Background Color
big	blue	orange
small	black	white

Equations

A LaTeX equation:

$$\int_0^\infty e^{-x^2} dx = \frac{\sqrt{\pi}}{2} \tag{1}$$

An equation too long to fit within page:

$$x = a + b + c + d + e + f + g + h + i + j + k + l + m + n + o + p + q + r + s + t + u + v + w + x + y + z + 1 + 2 + 3 + 4 + 5 + 6 + 7 + 8 + 9$$
(2)

Special

▲ WARNING The following features are only supported and intended for .html and .pdf exports. Journals are not likely to support them, and they may not display correctly when converted to other formats such as .docx.

LINK STYLED AS A BUTTON

Adding arbitrary HTML attributes to an element using Pandoc's attribute syntax:

Manubot Manubot Manubot Manubot Manubot. Manubot Manubot Manubot Manubot. Manubot Manubot Manubot. Manubot Manubot. Manubot.

Adding arbitrary HTML attributes to an element with the Manubot attributes plugin (more flexible than Pandoc's method in terms of which elements you can add attributes to):

Manubot Manubot.

Available background colors for text, images, code, banners, etc:

white lightgrey grey darkgrey black lightred lightyellow lightgreen lightblue lightpurple red orange yellow green blue purple

Using the Font Awesome icon set:



Light Grey Banner
useful for general information - manubot.org

1 Blue Banner

useful for important information - manubot.org

♦ Light Red Banner useful for *warnings* - <u>manubot.org</u>

References

1. A Review of PAT Strategies in Secondary Solid Oral Dosage Manufacturing of Small Molecules

Stephan Laske, Amrit Paudel, Otto Scheibelhofer, Stephan Sacher, Theresa Hoermann, Johannes Khinast, Adrian Kelly, Jukka Rantannen, Ossi Korhonen, Fanny Stauffer, ... Brenda Colegrove

Journal of Pharmaceutical Sciences (2017-03) https://doi.org/f965kb

DOI: 10.1016/j.xphs.2016.11.011 · PMID: 28017464

2. PAT-tools for process control in pharmaceutical film coating applications

Klaus Knop, Peter Kleinebudde

International Journal of Pharmaceutics (2013-12) https://doi.org/f5hg6m

DOI: 10.1016/j.ijpharm.2013.01.062 · PMID: 23380626

3. Transmission Raman spectroscopy as a tool for quantifying polymorphic content of pharmaceutical formulations

Adeyinka Aina, Michael D Hargreaves, Pavel Matousek, Jonathan C Burley

The Analyst (2010) https://doi.org/fsqqdt

DOI: 10.1039/c0an00352b · PMID: 20614090

4. Combined Fourier-transform infrared imaging and desorption electrospray-ionization linear ion-trap mass spectrometry for analysis of counterfeit antimalarial tablets

Camilla Ricci, Leonard Nyadong, Facundo M Fernandez, Paul N Newton, Sergei G Kazarian *Analytical and Bioanalytical Chemistry* (2006-11-29) https://doi.org/ctdcrs

DOI: 10.1007/s00216-006-0950-z · PMID: 17136340

5. Combining attenuated total reflectance- infrared spectroscopy and chemometrics for the identification and the dosage estimation of MDMA tablets

E Deconinck, R Van Campenhout, C Aouadi, M Canfyn, JL Bothy, L Gremeaux, P Blanckaert, P Courselle

Talanta (2019-04) https://doi.org/gm4tg4

DOI: 10.1016/j.talanta.2018.11.027 · PMID: 30625524

6. Raman spectroscopy in pharmaceutical product design

Amrit Paudel, Dhara Raijada, Jukka Rantanen

Advanced Drug Delivery Reviews (2015-07) https://doi.org/f7pkr2

DOI: 10.1016/j.addr.2015.04.003 · PMID: 25868453

7. Mass Spectrometry Sampling Under Ambient Conditions with Desorption Electrospray Ionization

Zoltán Takáts, Justin M Wiseman, Bogdan Gologan, RGraham Cooks

Science (2004-10-15) https://doi.org/fsf82c

DOI: 10.1126/science.1104404 · PMID: 15486296

8. Direct, trace level detection of explosives on ambient surfaces by desorption electrospray ionization mass spectrometry

Zoltán Takáts, Ismael Cotte-Rodriguez, Nari Talaty, Huanwen Chen, RGraham Cooks

Chem. Commun. (2005) https://doi.org/bw9fq2

DOI: 10.1039/b418697d · PMID: 15834468

9. Rapid detection of cocaine, benzoylecgonine and methylecgonine in fingerprints using surface mass spectrometry

Melanie J Bailey, Robert Bradshaw, Simona Francese, Tara L Salter, Catia Costa, Mahado Ismail, Roger P. Webb, Ingrid Bosman, Kim Wolff, Marcel de Puit

The Analyst (2015) https://doi.org/f728fc

DOI: 10.1039/c5an00112a · PMID: 25977942

10. Characterization of barley leaf tissue using direct and indirect desorption electrospray ionization imaging mass spectrometry

Bin Li, Nanna Bjarnholt, Steen H Hansen, Christian Janfelt Journal of Mass Spectrometry (2011-12) https://doi.org/fzhf26

DOI: 10.1002/jms.2010 · PMID: 22223414

Imaging drugs, metabolites and biomarkers in rodent lung: a DESI MS strategy for the 11. evaluation of drug-induced lipidosis

Alex Dexter, Rory T Steven, Aateka Patel, Lea Ann Dailey, Adam J Taylor, Doug Ball, Jan Klapwijk, Ben Forbes, Clive P Page, Josephine Bunch

Analytical and Bioanalytical Chemistry (2019-11-27) https://doi.org/gm4tg3

DOI: 10.1007/s00216-019-02151-z · PMID: 31776643 · PMCID: PMC6920235

Desorption electrospray ionization mass spectrometry for lipid characterization and 12. biological tissue imaging

Livia S Eberlin, Christina R Ferreira, Allison L Dill, Demian R Ifa, RGraham Cooks Biochimica et Biophysica Acta (BBA) - Molecular and Cell Biology of Lipids (2011-11) https://doi.org/cm2xk8

DOI: 10.1016/j.bbalip.2011.05.006 · PMID: 21645635 · PMCID: PMC3205348

13. **Desorption Electrospray Ionization Mass Spectrometry for High-Throughput Analysis of Pharmaceutical Samples in the Ambient Environment**

Huanwen Chen, Nari N Talaty, Zoltán Takáts, RGraham Cooks Analytical Chemistry (2005-09-27) https://doi.org/b45j8g

DOI: 10.1021/ac050989d · PMID: 16255590

14. Desorption electrospray ionization mass spectrometry: direct toxicological screening and analysis of illicit Ecstasy tablets

Luc Alexis Leuthold, Jean-François Mandscheff, Marc Fathi, Christian Giroud, Marc Augsburger, Emmanuel Varesio, Gérard Hopfgartner

Rapid Communications in Mass Spectrometry (2006-01-30) https://doi.org/cnt8ps

DOI: http://dx.doi.org/10.1002/rcm.2280

15. Combined Fourier-transform infrared imaging and desorption electrospray-ionization linear ion-trap mass spectrometry for analysis of counterfeit antimalarial tablets

Camilla Ricci, Leonard Nyadong, Facundo M Fernandez, Paul N Newton, Sergei G Kazarian Analytical and Bioanalytical Chemistry (2006-11-29) https://doi.org/ctdcrs DOI: https://doi.org/10.1007/s00216-006-0950-z

Direct quantitation of active ingredients in solid artesunate antimalarials by 16. noncovalent complex forming reactive desorption electrospray ionization mass spectrometry

Leonard Nyadong, Sameer Late, Michael D Green, Ajay Banga, Facundo M Fernández Journal of the American Society for Mass Spectrometry (2011-11-22) https://doi.org/d3f6ch DOI: https://doi.org/10.1016/j.jasms.2007.11.016

17. Desorption electrospray ionization reactions between host crown ethers and the influenza neuraminidase inhibitor oseltamivir for the rapid screening of Tamiflu® Leonard Nyadong, Edward G Hohenstein, Kristin Johnson, CDavid Sherrill, Michael D Green, Facundo M Fernández

The Analyst (2008) https://doi.org/cz4gnp
DOI: 10.1039/b809471c · PMID: 18936828

18. Desorption electrospray ionization with a portable mass spectrometer: in situ analysis of ambient surfaces

Christopher C Mulligan, Nari Talaty, RGraham Cooks *Chemical Communications* (2006) https://doi.org/dps53b

DOI: https://doi.org/10.1039/b517357d

19. Imaging of Polar and Nonpolar Species Using Compact Desorption Electrospray Ionization/Postphotoionization Mass Spectrometry

Chengyuan Liu, Keke Qi, Lei Yao, Ying Xiong, Xuan Zhang, Jianye Zang, Changlin Tian, Minggao Xu, Jiuzhong Yang, Zhenkun Lin, ... Yang Pan

Analytical Chemistry (2019-03-25) https://doi.org/gm4thf

DOI: https://doi.org/10.1021/acs.analchem.9b00520

20. A comparative investigation of modern feature selection and classification approaches for the analysis of mass spectrometry data

Piotr S Gromski, Yun Xu, Elon Correa, David I Ellis, Michael L Turner, Royston Goodacre *Analytica Chimica Acta* (2014-06) https://doi.org/f54wmh

DOI: https://doi.org/10.1016/j.aca.2014.03.039

21. Establishing chemical profiling for ecstasy tablets based on trace element levels and support vector machine

Camila Maione, Vanessa C de Oliveira Souza, Loraine R Togni, Jose L da Costa, Andres D Campiglia, Fernando Barbosa, Rommel M Barbosa

Neural Computing and Applications (2016-11-30) https://doi.org/gdw85d

DOI: <u>10.1007/s00521-016-2736-3</u>

22. Computational advances of tumor marker selection and sample classification in cancer proteomics

Jing Tang, Yunxia Wang, Yongchao Luo, Jianbo Fu, Yang Zhang, Yi Li, Ziyu Xiao, Yan Lou, Yunqing Qiu, Feng Zhu

Computational and Structural Biotechnology Journal (2020) https://doi.org/gg5hkb

DOI: https://doi.org/10.1016/j.csbj.2020.07.009

23. Navigating freely-available software tools for metabolomics analysis

Rachel Spicer, Reza M Salek, Pablo Moreno, Daniel Cañueto, Christoph Steinbeck *Metabolomics* (2017-08-09) https://doi.org/gdz62k

DOI: https://doi.org/10.1007/s11306-017-1242-7

24. The surgical intelligent knife distinguishes normal, borderline and malignant gynaecological tissues using rapid evaporative ionisation mass spectrometry (REIMS)

David L Phelps, Júlia Balog, Louise F Gildea, Zsolt Bodai, Adele Savage, Mona A El-Bahrawy, Abigail VM Speller, Francesca Rosini, Hiromi Kudo, James S McKenzie, ... Sadaf Ghaem-Maghami *British Journal of Cancer* (2018-04-19) https://doi.org/gdrdxv

DOI: https://doi.org/10.1038/s41416-018-0048-3

25. Identification of the Species of Origin for Meat Products by Rapid Evaporative Ionization Mass Spectrometry

Julia Balog, Dora Perenyi, Cristina Guallar-Hoyas, Attila Egri, Steven D Pringle, Sara Stead, Olivier P Chevallier, Chris T Elliott, Zoltan Takats

Journal of Agricultural and Food Chemistry (2016-06-02) https://doi.org/f8sfpp

DOI: https://doi.org/10.1021/acs.jafc.6b01041

Rapid Evaporative Ionization Mass Spectrometry Imaging Platform for Direct Mapping 26. from Bulk Tissue and Bacterial Growth Media

Ottmar Golf, Nicole Strittmatter, Tamas Karancsi, Steven D Pringle, Abigail VM Speller, Anna Mroz, James M Kinross, Nima Abbassi-Ghadi, Emrys A Jones, Zoltan Takats Analytical Chemistry (2015-02-17) https://doi.org/f65gwz

DOI: https://doi.org/10.1021/ac5046752

27. Rapid evaporative ionisation mass spectrometry of electrosurgical vapours for the identification of breast pathology: towards an intelligent knife for breast cancer surgery

Edward R St John, Julia Balog, James S McKenzie, Merja Rossi, April Covington, Laura Muirhead, Zsolt Bodai, Francesca Rosini, Abigail VM Speller, Sami Shousha, ... Daniel R Leff Breast Cancer Research (2017-05-23) https://doi.org/gbhgzz

DOI: https://doi.org/10.1186/s13058-017-0845-2

28. Wide area coverage Raman spectroscopy for reliable quantitative analysis and its applications

Kayeong Shin, Hoeil Chung The Analyst (2013) https://doi.org/gm4tmh

DOI: https://doi.org/10.1039/c3an36843b

Bulk Raman Analysis of Pharmaceutical Tablets 29.

P Matousek, AW Parker

Applied Spectroscopy (2016-08-31) https://doi.org/d52pcb

DOI: https://doi.org/10.1366/000370206779321463

30. Recent advances in the application of transmission Raman spectroscopy to pharmaceutical analysis

Kevin Buckley, Pavel Matousek

Journal of Pharmaceutical and Biomedical Analysis (2011-06) https://doi.org/cj83xx

DOI: https://doi.org/10.1016/j.jpba.2010.10.029

Quantitative Transmission Raman Spectroscopy of Pharmaceutical Tablets and Capsules 31.

Jonas Johansson, Anders Sparén, Olof Svensson, Staffan Folestad, Mike Claybourn Applied Spectroscopy (2016-08-31) https://doi.org/fdx3fw DOI: https://doi.org/10.1366/000370207782597085

32. Dependence of Signal on Depth in Transmission Raman Spectroscopy

Pavel Matousek, Neil Everall, David Littlejohn, Alison Nordon, Matthew Bloomfield Applied Spectroscopy (2011-07-01) https://doi.org/bmkq3n

DOI: https://doi.org/10.1366/11-06259

33. A guide to good practice in chemometric methods for vibrational spectroscopy, electrochemistry, and hyphenated mass spectrometry

Manuel David Peris-Díaz, Artur Kreżel

TrAC Trends in Analytical Chemistry (2021-02) https://doi.org/gm4tks

DOI: 10.1016/j.trac.2020.116157

34. Near-infrared Raman spectroscopy for the classification of epithelial pre-cancers and cancers

Nicholas Stone, Catherine Kendall, Neil Shepherd, Paul Crow, Hugh Barr Journal of Raman Spectroscopy (2002-07) https://doi.org/cn23j4 DOI: https://doi.org/10.1002/jrs.882

35. Disease recognition by infrared and Raman spectroscopy

Christoph Krafft, Gerald Steiner, Claudia Beleites, Reiner Salzer

Journal of Biophotonics (2009-02) https://doi.org/cnd3q2

DOI: https://doi.org/10.1002/jbio.200810024

36. Raman Spectroscopy in Colorectal Cancer Diagnostics: Comparison of PCA-LDA and PLS-DA Models

Wenjing Liu, Zhaotian Sun, Jinyu Chen, Chuanbo Jing *Journal of Spectroscopy* (2016) https://doi.org/f9gnw2

DOI: https://doi.org/10.1155/2016/1603609

37. Classification of lactic acid bacteria with UV-resonance Raman spectroscopy

K Gaus, P Rösch, R Petry, K-D Peschke, O Ronneberger, H Burkhardt, K Baumann, J Popp *Biopolymers* (2006-07) https://doi.org/fq8gm7

DOI: https://doi.org/10.1002/bip.20448

38. Model transfer for Raman-spectroscopy-based bacterial classification

Shuxia Guo, Ralf Heinke, Stephan Stöckel, Petra Rösch, Jürgen Popp, Thomas Bocklitz *Journal of Raman Spectroscopy* (2018-04) https://doi.org/gdf76b

DOI: https://doi.org/10.1002/jrs.5343

39. Classification of Narcotics in Solid Mixtures Using Principal Component Analysis and Raman Spectroscopy

Alan G Ryder

Journal of Forensic Sciences (2002-03-01) https://doi.org/gm4tmk

DOI: https://doi.org/10.1520/jfs15244j

40. Identification of pharmaceutical tablets by Raman spectroscopy and chemometrics

Yves Roggo, Klara Degardin, Pierre Margot

Talanta (2010-05-15) https://doi.org/bjz4j6

DOI: https://doi.org/10.1016/j.talanta.2010.01.046

41. Fast detection and identification of counterfeit antimalarial tablets by Raman spectroscopy

Marleen de Veij, Peter Vandenabeele, Krystyn Alter Hall, Facundo M Fernandez, Michael D Green, Nicholas J White, Arjen M Dondorp, Paul N Newton, Luc Moens *Journal of Raman Spectroscopy* (2007-02) https://doi.org/bh7s3g

Tournal of Naman Specifoscopy (2007-02) inteps

DOI: https://doi.org/10.1002/jrs.1621

42. Rapid and non-invasive screening of high renin hypertension using Raman spectroscopy and different classification algorithms

Xiangxiang Zheng, Guodong Lv, Ying Zhang, Xiaoyi Lv, Zhixian Gao, Jun Tang, Jiaqing Mo *Spectrochimica Acta Part A: Molecular and Biomolecular Spectroscopy* (2019-05)

https://doi.org/gm4tmc

DOI: https://doi.org/10.1016/j.saa.2019.02.063

43. Partial least squares discriminant analysis: taking the magic away

Richard G Brereton, Gavin R Lloyd

Journal of Chemometrics (2014-04) https://doi.org/f5wr2f

DOI: https://doi.org/10.1002/cem.2609

44. Comparison of several chemometric methods of libraries and classifiers for the analysis of expired drugs based on Raman spectra

Qun Gao, Yan Liu, Hao Li, Hui Chen, Yifeng Chai, Feng Lu

Journal of Pharmaceutical and Biomedical Analysis (2014-06) https://doi.org/f549x9

DOI: http://dx.doi.org/10.1016/j.jpba.2014.01.027

45. Comparison of multivariate methods for quantitative determination with transmission Raman spectroscopy in pharmaceutical formulations

Magnus Fransson, Jonas Johansson, Anders Sparén, Olof Svensson *Journal of Chemometrics* (2010-11) https://doi.org/fs46d5

DOI: http://dx.doi.org/10.1002/cem.1330

46. Welcome to the Tidyverse

Hadley Wickham, Mara Averick, Jennifer Bryan, Winston Chang, Lucy McGowan, Romain François, Garrett Grolemund, Alex Hayes, Lionel Henry, Jim Hester, ... Hiroaki Yutani *Journal of Open Source Software* (2019-11-21) https://doi.org/ggddkj

DOI: https://doi.org/10.21105/joss.01686

- 47. Tidymodels https://www.tidymodels.org/
- 48. https://cran.r-project.org/package

49. Testing for Multivariate Normality in Mass Spectrometry Imaging Data: A Robust Statistical Approach for Clustering Evaluation and the Generation of Synthetic Mass Spectrometry Imaging Data Sets

Alex Dexter, Alan M Race, Iain B Styles, Josephine Bunch Analytical Chemistry (2016-11-04) https://doi.org/f9cs8j DOI: 10.1021/acs.analchem.6b02139 · PMID: 27641083

50. Sci-Hub provides access to nearly all scholarly literature

Daniel S Himmelstein, Ariel Rodriguez Romero, Jacob G Levernier, Thomas Anthony Munro, Stephen Reid McLaughlin, Bastian Greshake Tzovaras, Casey S Greene *eLife* (2018-03-01) https://doi.org/ckcj

DOI: <u>10.7554/elife.32822</u> · PMID: <u>29424689</u> · PMCID: <u>PMC5832410</u>

51. Reproducibility of computational workflows is automated using continuous analysis

Brett K Beaulieu-Jones, Casey S Greene

Nature biotechnology (2017-04) https://www.ncbi.nlm.nih.gov/pmc/articles/PMC6103790/
DOI: 10.1038/nbt.3780 · PMID: 28288103 · PMCID: PMC6103790

52. **Bitcoin for the biological literature.**

Douglas Heaven

Nature (2019-02) https://www.ncbi.nlm.nih.gov/pubmed/30718888

DOI: 10.1038/d41586-019-00447-9 · PMID: 30718888

53. Plan S: Accelerating the transition to full and immediate Open Access to scientific publications

cOAlition S

(2018-09-04) https://www.wikidata.org/wiki/Q56458321

54. Open access

Peter Suber *MIT Press* (2012) ISBN: 9780262517638

55. Open collaborative writing with Manubot

Daniel S Himmelstein, Vincent Rubinetti, David R Slochower, Dongbo Hu, Venkat S Malladi, Casey S Greene, Anthony Gitter

Manubot (2020-05-25) https://greenelab.github.io/meta-review/

56. Opportunities and obstacles for deep learning in biology and medicine

Travers Ching, Daniel S Himmelstein, Brett K Beaulieu-Jones, Alexandr A Kalinin, Brian T Do, Gregory P Way, Enrico Ferrero, Paul-Michael Agapow, Michael Zietz, Michael M Hoffman, ... Casey S Greene

Journal of The Royal Society Interface (2018-04-04) https://doi.org/gddkhn
DOI: 10.1098/rsif.2017.0387 · PMID: 29618526 · PMCID: PMC5938574

57. **Open collaborative writing with Manubot**

Daniel S Himmelstein, Vincent Rubinetti, David R Slochower, Dongbo Hu, Venkat S Malladi, Casey S Greene, Anthony Gitter

PLOS Computational Biology (2019-06-24) https://doi.org/c7np

DOI: <u>10.1371/journal.pcbi.1007128</u> · PMID: <u>31233491</u> · PMCID: <u>PMC6611653</u>