

Prosthetic Heart Valves

ED15B001 Adarsh
ED15B019 Sanjeev



Natural Heart Valves

- Allow unidirectional flow of blood in the heart
- 4 Valves in the human heart- Tricuspid, Mitral, Pulmonary & Aortic
- Incorporate leaflets or cusps which are pushed open to allow blood flow & which then close together to seal & prevent backflow




Fig: Illustration of Natural Heart Valves[†]

[†] <http://www.yourheartvalve.com/heartbasics/pages/heartvalves.aspx>



Loading Cycle of Natural HVs

- HVs are primarily passive structures that are driven by forces exerted by the surrounding blood & heart
- HVs must replicate their cyclic function over an entire lifetime, with an estimated functional demand of at least 3×10^9 cycles

[†]M.S. Sacks, A.P. Yoganathan, “Heart valve function: a biomechanical perspective”, Philosophical Transactions of the Royal Society, 2007 Aug 29; 362(1484): 1369–1391




Fig: Instantaneous friction streamline & shear stress magnitude plots on the aortic (a,c) & ventricular (b,d) sides of the leaflets during the fully open (a,b) & early closing (c,d) phases of the cardiac cycle[†]



Properties of Natural HVs

- High tensile strength to resist high Trans Valvular Pressures
- Very low flexural rigidity to allow for passive interactions with the surrounding blood
- Undergo large, rapid, directionally dependent strain when in the process of closing
- Allow for rapid cessation of strain when closed



Failure Modes of Natural HVs

Regurgitation

- Leaflets of the valve fail to coapt correctly
- Allows blood to flow in the reverse direction


Stenosis

- Narrowing of the valves
- Prevents adequate outflow of blood



Artificial Heart Valves


Prosthetic heart valves are required to replicate the functioning of a normal heart valve & should last throughout the lifetime of the patient




Mechanical HVs



Starr-Edwards Caged Ball Valve(1960s)[†]



Medtronic Hall Tilting Disc Valve(1970s)[†]



St. Jude's Medical Bileaflet Valve(1979)[†]

A titanium cage, a silastic ball, & a suture ring covered with teflon

A tilting pyrolytic carbon disc with a central perforation for a thin metal strut


Two semicircular leaflets that rotate about struts attached to the valve housing

[†]W. Vongpatanasin, L.D. Hillis, R.A. Lange, “ *Prosthetic Heart Valves*”, The New England Journal of Medicine 1996; 335:407-416



Bioprosthetic HVs

- **Homografts:** Preserved human aortic valves
- **Heterografts:** Composed of bovine or porcine tissue(pericardial or valvular), mounted on a metal support



**Carpentier-Edwards
Porcine Bioprosthesis[†]**

[†]W. Vongpatanasin, L.D. Hillis, R.A. Lange, “ *Prosthetic Heart Valves*”, The New England Journal of Medicine 1996; 335:407-416



Prosthetic HVs Materials

- **Mechanical Heart Valves**
 - Metals - Stainless Steel, Titanium
 - Ceramics - Titanium Dioxide, Pyrolytic Carbon[†]
 - Synthetic Materials- Silicone
- **Bioprosthetic Heart Valves**
 - Natural Materials - Bovine & Porcine tissue
 - Synthetic Materials - PTFE, Dacron

[†]C.R. Gentle, “*The use of ceramics in prosthetic heart valves*”, Journal of Engineering in Medicine, Volume 16, Issue 2, 1987, p.no: 115-117



Market Evaluation

- Increase in valvular disease due to:
 - Socioeconomic factors
 - Growing life expectancy
- Currently a 4.84 billion USD market, the prosthetic HV market is projected to reach 8.86 billion USD by 2022*
- Abbott Healthcare, Edwards LifeSciences Corp., St. Jude's Medical, Medtronic PLC, & Boston Scientific Corp. are the major players in this segment

*<https://www.marketsandmarkets.com/Market-Reports/prosthetic-heart-valve-market-245407958.html>

†J.S. Gammie et al, "Isolated Mitral Valve Surgery: The Society of Thoracic Surgeons Adult Cardiac Surgery Database Analysis", The Annals of Thoracic Surgery, September 2018, Volume 106, Issue 3




Fig: Total number of isolated mitral valve operations performed every year in 1143 major US hospitals (2011-2016)[†]



Issues with Prosthetic HVs

- **Durability**
 - Mechanical HVs: Life of 20 to 30 years
 - Bioprosthetic HVs: Require replacement within 10-15 years[†]
- **Thrombogenicity**
 - Reduced leaflet movement, thickening of valve leaflets, changes in effective prosthesis area (stenosis or regurgitation)
 - Mechanical HVs: Highly thrombogenic; Long term anticoagulant therapy (warfarin) required
 - Bioprosthetic HVs: Less thrombogenic

[†]W. Vongpatanasin, L.D. Hillis, R.A. Lange, “ *Prosthetic Heart Valves*”, The New England Journal of Medicine 1996; 335:407-416



Issues with Prosthetic HVs

- **Structural Failure**

- Mechanical HVs: Rare
- Bioprosthetic HVs: Calcification leads to cusps becoming rigid and rupturing

- **Embolization**

- An unattached mass clogs arterial beds at a point distal to its origin. Eg: Thromboembolism

- **Paravalvular Regurgitation**

- Improper implantation, sutures give way over time
- Causes trauma to blood cells; leads to hemolysis



Solutions- Polymeric Materials

- Combine durability & strength of mechanical HVs and biocompatibility of bioprosthetic HVs
- Polyurethane (PU) is attractive option:
 - Hard (crystalline) and soft (elastomeric) segments
 - Proportion determines features like stiffness
 - Favourable mechanical and hemodynamic properties, resistance to thrombus formation
- Drawbacks: Susceptibility to degradation & risk of calcification;



Solutions- Polymeric Materials

- Soft segments are modified - three variant polymers of PU
- Polyester urethane
 - Soft segments were easily hydrolysed; hence unsuitable for long term implantation
- Polyether urethane
 - Resistant to hydrolysis but susceptible to oxidative degradation and cracking under env. stress *in vivo*
- Polycarbonate urethane
 - Resistant to oxidative degradation
 - Biodegradation is limited to peripheral layer



Solutions- Polymeric Materials

- Modification of chemical structure
 - Linking biodegradation resistant molecules
 - Eg: polydimethylsiloxane (thermal and oxidative stability)
 - Attempts to link anticalcification agents into the polymer structure
 - Eg: covalent binding of bisphosphonate (2-hydroxyethane bisphosphonic acid [HEBP])[†]

[†]Hossein Ghanbari, Helene Viatge, “Polymeric heart valves: new materials, emerging hopes”, Colloids and Surfaces B: Biointerfaces



Solutions- Endothelialization

- Provide a coating of endothelial cells on prosthesis surface so as to reduce risks of thrombosis and autoimmune reactions
- Two techniques - Tissue Engineering & Surface Modification
- **Tissue Engineering:** Seeding of cells on the polymer surface and proliferating them *In vitro*
 - Need for providing the cell culture
 - Risk of infection



Solutions- Endothelialization

- **Surface Modification:** Modify surface of polymer so as to promote adhesion of endothelial progenitor cells (EPCs) *in situ*[†]
 - Application of short peptide sequences recognized by cell receptors
 - Low concentration of EPCs; Need for a **selective** surface
 - The tripeptide Arg-Gly-Asp (RGD) has been in common use
 - Also recognized by platelet integrin receptors
 - Tetrapeptide Arg-Glu-Asp-Val (REDV) is shown to have high selectivity towards human endothelial cells

[†]Beata A.Butruk-Raszeja, Magdalena S.Dresler, Aleksandra Kuzminska, Tomasz Ciach, “*Endothelialization of polyurethanes: surface silanization and immobilization of REDV peptide, Colloids and Surfaces*”



Conclusions

- Proposed design involves the use of Polycarbonate urethane, structurally modified by linking biodegradation resistant molecules and anti calcification molecules
- Further, surface modification techniques used to enhance affinity of human endothelial progenitor cells (EPCs) towards the prosthesis
- Expected to have longer life as a result of reduced risks of biodegradation and calcification
- Also expected to be less susceptible to thrombosis and autoimmune reactions due to endothelialization



THANK YOU