Cabinet de dermatologie Es-Senia


Triage support report — skin lesion

dermIA v2.0.0 · model ham10000-bcn20000-m5-v3 v3.0.0 (full profile)


Patient	P-2026-0422	Analysis date	29/10/2026, 09:22
Age	60 years	Requested by	Dr Karim Haddad
Sex	female	Body site	back
Tracked lesion	#1 — dos, sous l'omoplate gauche	Decision thresholds	original

PROBABLY BENIGN


Melanocytic nevus likely
Probably benign lesion. Follow-up according to clinical judgement; any change in the lesion warrants a new examination.



Analysed image



Most influential regions (Grad-CAM, “Melanocytic nevus”)



Outline and axes used for ABCD

Estimated probabilities

Diagnosis	Probability	In the guaranteed set*
Melanocytic nevus	100.0%	yes
Benign keratosis	0.0%	
Melanoma	0.0%	
Vascular lesion	0.0%	
Basal cell carcinoma	0.0%	
Dermatofibroma	0.0%	
Actinic keratosis / Bowen's disease	0.0%	

* Conformal set (Mondrian, alpha = 0.10): contains the true diagnosis with a probability of at least 90% for each class, assuming the cases resemble the calibration data. Melanoma decision if p(mel) ≥ 0.03 (cost ratio R = 10).
Original model thresholds, calibrated on its calibration set.

Automatic ABCD criteria (indicative)

Criterion	Measure	Detail
A — Asymmetry	0/2	asymmetric principal axes (shape or colours)
B — Border	0/8	sectors with an abrupt edge
C — Colours	3/6	dark brown, blue-grey, black
D — Diameter	9.2 mm	more than 6 mm

Indicative automatic measurements: on the HAM10000 test set, each separates melanoma from nevus only modestly (AUC 0.66 to 0.73). They do not replace dermoscopic examination.

Change since the analysis of 09/03/2026 (233 days)

• Diameter increased: from 6.4 mm to 9.2 mm.

Also compare the images visually: framing and magnification may differ between examinations.

Medical review

Not yet reviewed by a physician.

Physician's signature: _____

Disclaimer. This document is produced by a decision-support tool. It is not a diagnosis and replaces neither clinical examination nor histology. Model status: RESEARCH — non-commercial use — not a medical device. Training data: HAM10000 (Tschandl et al., Sci Data 2018, doi:10.1038/sdata.2018.161) + BCN20000 (Hernández-Pérez et al., Sci Data 2024, doi:10.1038/s41597-024-03387-w) — licence CC BY-NC 4.0 (non-commercial).