Cabinet de dermatologie Es-Senia


Triage support report — skin lesion

DermIA Prudence v2.0.0 · model ham10000-m5-v2 v2.0.0 (full profile)


Patient	P-2026-0422	Analysis date	12/10/2026, 19:00
Age	60 years	Requested by	Dr Karim Haddad
Sex	female	Body site	upper limb
Tracked lesion	#1 — avant-bras droit	Decision thresholds	original

URGENT


Suspected melanoma
The probability of melanoma exceeds the safety threshold. Prompt dermatological assessment recommended.



Analysed image



Most influential regions (Grad-CAM, “Melanoma”)



Outline and axes used for ABCD

Estimated probabilities

Diagnosis	Probability	In the guaranteed set*
Melanoma	98.9%	yes
Melanocytic nevus	1.1%	
Benign keratosis	0.0%	
Actinic keratosis / Bowen's disease	0.0%	
Dermatofibroma	0.0%	
Basal cell carcinoma	0.0%	
Vascular lesion	0.0%	

* Conformal set (Mondrian, alpha = 0.10): contains the true diagnosis with a probability of at least 90% for each class, assuming the cases resemble the calibration data. Melanoma decision if p(mel) ≥ 0.10 (cost ratio R = 10).
Original model thresholds, calibrated on the HAM10000 calibration set.

Automatic ABCD criteria (indicative)

Criterion	Measure	Detail
A — Asymmetry	1/2	asymmetric principal axes (shape or colours)
B — Border	1/8	sectors with an abrupt edge
C — Colours	2/6	light brown, dark brown
D — Diameter	9.3 mm	more than 6 mm

Indicative automatic measurements: on the HAM10000 test set, each separates melanoma from nevus only modestly (AUC 0.66 to 0.73). They do not replace dermoscopic examination.

Change since the analysis of 20/02/2026 (233 days)

• New colour(s): light brown.
• Diameter increased: from 4.7 mm to 9.3 mm.

Also compare the images visually: framing and magnification may differ between examinations.

Medical review

Not yet reviewed by a physician.

Physician's signature: _____

Disclaimer. This document is produced by a decision-support tool. It is not a diagnosis and replaces neither clinical examination nor histology. Model status: DEMONSTRATION — non-commercial use — not a medical device. Training data: HAM10000 (Tschandl et al., Sci Data 2018, doi:10.1038/sdata.2018.161) — licence CC BY-NC 4.0 (non-commercial).