Cabinet de dermatologie Es-Senia


# Triage support report — skin lesion

DermIA Prudence v2.0.0 · model ham10000-m5-v2 v2.0.0 (full profile)


| | | | |
|---|---|---|---|
| Patient | P-2026-0423 | Analysis date | 21/05/2026, 22:38 |
| Age | 70 years | Requested by | Dr Karim Haddad |
| Sex | male | Body site | back |
| Tracked lesion | #1 — dos, sous l'omoplate gauche | Decision thresholds | original |

**PROBABLY BENIGN**


**Benign keratosis likely**
Probably benign lesion. Follow-up according to clinical judgement; any change in the lesion warrants a new examination.



Analysed image



Most influential regions (Grad-CAM, “Benign keratosis”)



Outline and axes used for ABCD

## Estimated probabilities

| Diagnosis | Probability | In the guaranteed set* |
|---|---|---|
| Benign keratosis | 94.1% | yes |
| Melanoma | 4.1% | |
| Melanocytic nevus | 1.8% | |
| Actinic keratosis / Bowen's disease | 0.0% | |
| Dermatofibroma | 0.0% | |
| Basal cell carcinoma | 0.0% | |
| Vascular lesion | 0.0% | |

* Conformal set (Mondrian, alpha = 0.10): contains the true diagnosis with a probability of at least 90% for each class, assuming the cases resemble the calibration data. Melanoma decision if p(mel) ≥ 0.10 (cost ratio R = 10).
Original model thresholds, calibrated on the HAM10000 calibration set.

## Automatic ABCD criteria (indicative)

| Criterion | Measure | Detail |
|---|---|---|
| A — Asymmetry | 2/2 | asymmetric principal axes (shape or colours) |
| B — Border | 1/8 | sectors with an abrupt edge |
| C — Colours | 2/6 | dark brown, blue-grey |
| D — Diameter | 10.7 mm | more than 6 mm |

Indicative automatic measurements: on the HAM10000 test set, each separates melanoma from nevus only modestly (AUC 0.66 to 0.73). They do not replace dermoscopic examination.

## Medical review

Confirmed by Dr Karim Haddad on 22/05/2026.
**Final diagnosis: Benign keratosis**

Physician's signature: _____

Disclaimer. This document is produced by a decision-support tool. It is not a diagnosis and replaces neither clinical examination nor histology. Model status: DEMONSTRATION — non-commercial use — not a medical device. Training data: HAM10000 (Tschandl et al., Sci Data 2018, doi:10.1038/sdata.2018.161) — licence CC BY-NC 4.0 (non-commercial).