Cabinet de dermatologie Es-Senia


Triage support report — skin lesion

DermIA Prudence v2.0.0 · model ham10000-m5-v2 v2.0.0 (full profile)


Patient	P-2026-0422	Analysis date	08/10/2026, 16:09
Age	70 years	Requested by	Dr Karim Haddad
Sex	male	Body site	foot
Tracked lesion	#2	Decision thresholds	original

REFER TO A PHYSICIAN


Uncertain result: the AI abstains
Several diagnoses remain plausible at the guaranteed confidence level. The case must be examined by a physician.



Analysed image



Most influential regions (Grad-CAM, “Melanocytic nevus”)



Outline and axes used for ABCD

Estimated probabilities

Diagnosis	Probability	In the guaranteed set*
Melanocytic nevus	77.2%	yes
Actinic keratosis / Bowen's disease	7.8%	yes
Melanoma	7.2%	yes
Benign keratosis	4.3%	
Dermatofibroma	2.0%	
Basal cell carcinoma	1.5%	
Vascular lesion	0.0%	

* Conformal set (Mondrian, alpha = 0.10): contains the true diagnosis with a probability of at least 90% for each class, assuming the cases resemble the calibration data. Melanoma decision if p(mel) ≥ 0.10 (cost ratio R = 10).
Original model thresholds, calibrated on the HAM10000 calibration set.

Automatic ABCD criteria (indicative)

Criterion	Measure	Detail
A — Asymmetry	1/2	asymmetric principal axes (shape or colours)
B — Border	0/8	sectors with an abrupt edge
C — Colours	2/6	dark brown, blue-grey
D — Diameter	10.4 mm	more than 6 mm

Indicative automatic measurements: on the HAM10000 test set, each separates melanoma from nevus only modestly (AUC 0.66 to 0.73). They do not replace dermoscopic examination.

Medical review

Not yet reviewed by a physician.

Physician's signature: _____

Disclaimer. This document is produced by a decision-support tool. It is not a diagnosis and replaces neither clinical examination nor histology. Model status: DEMONSTRATION — non-commercial use — not a medical device. Training data: HAM10000 (Tschandl et al., Sci Data 2018, doi:10.1038/sdata.2018.161) — licence CC BY-NC 4.0 (non-commercial).