Cabinet de dermatologie Es-Senia


# Triage support report — skin lesion

DermIA Prudence v2.0.0 · model ham10000-m5-v2 v2.0.0 (full profile)


| | | | |
|---|---|---|---|
| Patient | P-2026-0431 | Analysis date | 18/10/2026, 19:08 |
| Age | 46 years | Requested by | Dr Karim Haddad |
| Sex | female | Body site | hand |
| Tracked lesion | #1 | Decision thresholds | original |

**REFER TO A PHYSICIAN**


**Uncertain result: the AI abstains**
Several diagnoses remain plausible at the guaranteed confidence level. The case must be examined by a physician.



Analysed image



Most influential regions (Grad-CAM, “Melanocytic nevus”)



Outline and axes used for ABCD

## Estimated probabilities

| Diagnosis | Probability | In the guaranteed set* |
|---|---|---|
| Melanocytic nevus | 74.1% | yes |
| Benign keratosis | 8.7% | |
| Vascular lesion | 8.6% | yes |
| Melanoma | 3.5% | |
| Dermatofibroma | 2.9% | |
| Basal cell carcinoma | 1.5% | |
| Actinic keratosis / Bowen's disease | 0.7% | |

* Conformal set (Mondrian, alpha = 0.10): contains the true diagnosis with a probability of at least 90% for each class, assuming the cases resemble the calibration data. Melanoma decision if p(mel) ≥ 0.10 (cost ratio R = 10).
Original model thresholds, calibrated on the HAM10000 calibration set.
Caution: image atypical of dermoscopy, analysed at the user's request: interpret the result with great caution.

## Automatic ABCD criteria (indicative)

| Criterion | Measure | Detail |
|---|---|---|
| A — Asymmetry | 0/2 | asymmetric principal axes (shape or colours) |
| B — Border | 4/8 | sectors with an abrupt edge |
| C — Colours | 1/6 | blue-grey |
| D — Diameter | 15.0 mm | more than 6 mm |

Caution: The lesion touches the image border: A, B and D are less reliable.
Indicative automatic measurements: on the HAM10000 test set, each separates melanoma from nevus only modestly (AUC 0.66 to 0.73). They do not replace dermoscopic examination.

## Medical review

Not yet reviewed by a physician.

Physician's signature: _____

Disclaimer. This document is produced by a decision-support tool. It is not a diagnosis and replaces neither clinical examination nor histology. Model status: DEMONSTRATION — non-commercial use — not a medical device. Training data: HAM10000 (Tschandl et al., Sci Data 2018, doi:10.1038/sdata.2018.161) — licence CC BY-NC 4.0 (non-commercial).